

Enclosure 3

**Summer 2011 Compliance Survey for Watts Bar Nuclear Plant Outfall
Passive Mixing Zone**

TENNESSEE VALLEY AUTHORITY
River Operations

SUMMER 2011 COMPLIANCE SURVEY FOR WATTS BAR NUCLEAR PLANT OUTFALL 113 PASSIVE MIXING ZONE

Prepared by

Daniel P. Saint
and
Paul N. Hopping

Knoxville, Tennessee
March 2012



EXECUTIVE SUMMARY

The National Pollutant Discharge Elimination System (NPDES) Permit No. TN0020168 for Watts Bar Nuclear Plant (WBN) identifies the discharge of water to the Tennessee River from the Supplemental Condenser Cooling Water (SCCW) System as Outfall 113. Furthermore, the permit identifies that when there is no flow released from Watts Bar Dam (WBH), the effluent from Outfall 113 shall be regulated based on a passive mixing zone extending in the river from bank-to-bank and 1,000 feet downstream from the outfall. Compliance with the requirements for the passive mixing zone is to be achieved by two annual instream temperature surveys—one for winter conditions and one for summer conditions. Summarized in this report are the measurements, analyses, and results for the passive mixing zone survey performed for 2011 summer conditions. The survey was conducted between 21:00 CDT on August 30 and 05:00 CDT on August 31 (eight hours) and included the collection of temperature data at twelve temporary monitoring stations deployed across the downstream end of the passive mixing zone during a period of no flow in the river. The data were analyzed to determine the three instream compliance parameters specified in the NPDES permit for the outfall: the 1-hour average temperature at the downstream end of mixing zone, T_d ; the 1-hour average temperature rise from upstream to the downstream end of the mixing zone, ΔT ; and the 1-hour average temperature rate-of-change at the downstream end of the mixing zone, TROC. The measured parameters were compared to predicted values from the thermal plume model used by TVA to help determine the safe operation of Outfall 113. The results of the comparisons, in terms of maximum values observed during the no flow event, are as follows:

| Compliance Parameter | Model | Measured | NPDES Limit |
|----------------------|-------------|-------------|-------------|
| Maximum T_d | 80.8°F | 80.6°F | 86.9°F |
| Maximum ΔT | 1.5 F° | 1.6 F° | 5.4 F° |
| Maximum TROC | 0.6 F°/hour | 0.2 F°/hour | 3.6 F°/hr |

As shown, both the model and measured values were well below the NPDES limits for all the compliance parameters. Except for the maximum ΔT , values predicted by the model were larger than those measured in the survey. The maximum value of ΔT from the model underpredicted the measured value by 0.1 F°. This difference was caused by unnatural cooling of the upstream ambient temperature from leakage of cold water through Watts Bar Dam. Based on this, as well as the fact that differences of magnitude 0.1 F° easily fall within the factor of safety currently used in performing hydrothermal forecasts, the thermal plume model is yet considered fully adequate for determining the safe operation of the SCCW system. That is, in combination with TVA procedures for predicting the impact of the Outfall 113 effluent, the model continues to provide a high level of confidence that the SCCW system is being operated in a manner that does not exceed the instream limits for T_d , ΔT , and TROC as specified in the WBN NPDES permit.

TABLE OF CONTENTS

| | <u>Page No.</u> |
|---|------------------------|
| EXECUTIVE SUMMARY | i |
| INTRODUCTION | 1 |
| INSTREAM SURVEY | 2 |
| RESULTS | 3 |
| River Conditions | 3 |
| SCCW Conditions..... | 4 |
| Downstream End of Passive Mixing Zone | 4 |
| NPDES Compliance Parameters..... | 5 |
| CONCLUSIONS..... | 7 |
| REFERENCES | 8 |
| APPENDIX A..... | 19 |
| APPENDIX B | 27 |

LIST OF FIGURES

| | |
|---|----|
| Figure 1. Watts Bar Nuclear Plant Outfall 113 (SCCW) Mixing Zones | 9 |
| Figure 2. Location of HOBO Monitoring Stations | 10 |
| Figure 3. Schematic of HOBO Water Temperature Monitoring Stations | 10 |
| Figure 4. River Conditions..... | 11 |
| Figure 5. SCCW Conditions | 12 |
| Figure 6. HOBO Water Temperature Measurements | 13 |
| Figure 7. Instantaneous Temperature Rise for HOBO Measurements | 15 |
| Figure 8. Measured and Computed Compliance Parameters for Passive Mixing Zone | 18 |

LIST OF TABLES

| | |
|--|---|
| Table 1. NPDES Temperature Limits for Outfall 113 Mixing Zones | 1 |
| Table 2. Sources of Data for Passive Mixing Zone Survey..... | 2 |

WINTER 2011 COMPLIANCE SURVEY FOR WATTS BAR NUCLEAR PLANT OUTFALL 113 PASSIVE MIXING ZONE

INTRODUCTION

Outfall 113 for the Watts Bar Nuclear Plant (WBN) includes the discharge of water to the Tennessee River from the Supplemental Condenser Cooling Water (SCCW) system. Due to the dynamic behavior of the thermal effluent in the river, the National Pollutant Discharge Elimination System (NPDES) Permit No. TN0020168 for the plant specifies two mixing zones for Outfall 113—one for active operation of the river and one for passive operation of the river (TDEC, 2010). The passive mixing zone corresponds to periods when the operation of Watts Bar Dam (WBH) produces no flow in the river (i.e., hydropower and/or spillway releases). The dimensions of the passive mixing zone extend from bank-to-bank and downstream 1,000 feet from the outfall. The active mixing zone applies to all other river flow conditions. The dimensions of the active mixing zone include the right-half of the river (facing downstream) and extend downstream 2,000 feet from the outfall. The passive and the active mixing zones are shown in Figure 1.

Table 1 summarizes the NPDES instream temperature limits for Outfall 113. The limits apply to both the active and passive mixing zones. Compliance for the active mixing zone is monitored by permanent instream water temperature stations situated in the right-half of the river. Due to issues associated with placing permanent stations in the left-half of the river, which contains the navigation channel, a thermal plume model is used to determine the safe operation of Outfall 113 for the passive mixing zone. To verify the thermal plume model, the NPDES permit specifies that two instream temperature surveys shall be conducted each year—one for winter conditions and one for summer conditions. The purpose of this report is to present the results for the passive mixing zone temperature survey performed for summer 2011 conditions. The survey was conducted between 21:00 CDT on August 30 and 05:00 CDT on August 31 (total eight hours). Provided is a brief summary of the survey method, presentations of the measurements and analyses, and discussions of the results and conclusions.

Table 1. NPDES Temperature Limits for Outfall 113 Mixing Zones

| Compliance Parameter | Sampling Period | NPDES Limit |
|--|-----------------|-------------|
| Maximum Temperature, Downstream End of Mixing Zone, T_d | Running 1-hr | 86.9°F |
| Maximum Temperature Rise, Upstream to Downstream, ΔT | Running 1-hr | 5.4 F° |
| Maximum Temperature Rate-of-Change, TROC | Running 1-hr | ±3.6 F°/hr |

INSTREAM SURVEY

The instream survey included the deployment of temporary water temperature stations at twelve locations across the downstream end of the passive mixing zone. Data from these and other monitoring stations were analyzed to obtain measured values for the compliance parameters listed in Table 1. These were then compared with the corresponding values estimated from the SCCW thermal plume model.

The method of conducting the instream survey is the same as that used for the first such survey, performed for winter conditions on May 6, 2005 (McCall and Hopping, 2005). Table 2 provides a summary of the sources of data for the survey. WaterView, a monitoring system for tracking hydroplant operation and performance, was used to obtain measurements for the river discharge from Watts Bar Dam. The WBN Environmental Data Station (EDS) provided measurements from existing permanent monitoring stations for the nuclear plant. These included:

- The river upstream (ambient) water temperature, measured at the EDS Station 30, which is located at the exit of the powerhouse of Watts Bar Dam.
- The river water surface elevation (WSEL) at the EDS Station 30, also known as the tailwater elevation (TWEL) at Watts Bar Dam.
- The SCCW effluent temperature, measured at the EDS Station 32, which is located at the SCCW outfall.
- The SCCW effluent discharge, measured at the EDS Station 32.
- The local air temperature, measured at the EDS meteorological tower.

Table 2. Sources of Data for Passive Mixing Zone Survey

| Data | Source | Frequency |
|---------------------------------------|---------------------------------------|-----------|
| River Discharge from Watts Bar Dam | WaterView | 1 min |
| River ambient water temperature | WBN EDS Station 30 (Tailwater at WBH) | 15 min |
| River water surface elevation | WBN EDS Station 30 (Tailwater at WBH) | 15 min |
| SCCW effluent temperature | WBN EDS Station 32 (SCCW Outfall 113) | 15 min |
| SCCW effluent discharge | WBN EDS Station 32 (SCCW Outfall 113) | 15 min |
| Air temperature | WBN EDS Met Tower | 15 min |
| Passive mixing zone water temperature | Temporary HOBO Monitors | 1 min |

The water temperature at the downstream end of the Outfall 113 passive mixing zone was measured by the aforementioned temporary water temperature stations. Using a global positioning system (GPS) device, the stations were positioned at roughly equal intervals across the river, as shown in Figure 2. The temporary stations recorded water temperatures by using HOBO temperature monitors positioned at depths of 0.5, 3, 5, and 7 feet below the water surface. Shown in Figure 3 is a schematic of the temporary stations. The stations included a string of

HOBO monitors suspended from a tire float, with weights to anchor the station and to keep the sensor string vertical in the water column. The water temperature sensors imbedded in the HOBO monitors have an accuracy of about ± 0.4 F° and resolution of about 0.04 F°, which is comparable to the accuracy and resolution of temperature sensors used elsewhere by TVA for NPDES thermal compliance. The HOBO monitors include an internal data acquisition unit that was programmed to collect measurements once per minute. All the temperature probes used in the survey, including both those contained in the HOBO monitors and the thermistors at the permanent EDS monitoring stations, were calibrated by a quality program with equipment accuracies traceable to the National Institute of Standards and Technology (NIST). The calibration procedure is summarized in APPENDIX A. The temporary monitoring stations were deployed several hours before the beginning of the survey, and retrieved several hours after the end of the survey.

RESULTS

River Conditions

Figure 4 shows the measured ambient conditions of the river during the survey. Included are the river discharge, the river tailwater elevation, and river temperature at the exit of Watts Bar Dam. The river temperature at the exit of Watts Bar Dam serves as the upstream ambient river temperature for WBN Outfall 113. To provide a period of no flow in the river, releases from Watts Bar Dam were suspended between about 21:00 CDT on August 30 and 05:00 CDT on August 31, a total of eight hours (nighttime). Leading up to the survey, as the river flow was stepping down, the WSEL below Watts Bar Dam dropped approximately 0.8 feet, from about 681.4 feet msl to about 680.6 feet msl. During the survey, the elevation slowly increased, due to backflow from the surrounding tailwater and leakage through the hydroturbines, returning to about 681.4 feet msl after four hours of no flow in the river. Afterwards, the elevation slowly receded, reaching about 680.9 feet msl at the end of the survey.

The ambient river temperature was 79.3°F at the beginning of the period of no flow, and in a manner similar to the WSEL, increased in the first half of the survey, reaching a maximum of 79.9°F (increase of 0.6 F°). Afterwards, the temperature first receded slowly, only 0.2 F° in the next 2½ hours. However, in the final 1½ hours of the survey, the temperature dropped more rapidly, an additional 0.8 F°, reaching 78.9°F at the end of the period of no flow. A rapid drop in ambient river temperature in this manner is common in the summer when strong thermal stratification exists behind Watts Bar Dam. During periods of no flow, leakage occurs through the hydroturbines at the dam. Previous studies have suggested the amount of leakage to be roughly 50 cfs for each hydro unit, or a total of 250 cfs for the entire powerhouse (Harper et. al, 1998). The leakage flow is from the very bottom of Watts Bar Reservoir, the coldest part of the water column in front of the dam. As the leakage occurs, it slowly fills the bottom layers of the

tailrace below the powerhouse, eventually reaching the elevation of the sensors that are suspended in the water (from the surface) to measure the upstream ambient river temperature for WBN. Cooling of the ambient river temperature monitor in this manner falsely increases the measured temperature rise for the SCCW system. That is, the temperature rise is elevated not by warming from the SCCW effluent, but by “unnatural” cooling of the upstream monitor via a process that is beyond the operational control of the SCCW system. In forecasting values for the WBN upstream ambient river temperature, the thermal plume model for the SCCW system does not include cooling that occurs as a result of leakage through the hydroturbines at Watts Bar Dam.

SCCW Conditions

During the survey, the SCCW system at WBN was thermally loaded and operating in “summer” mode. That is, the system was operating in a manner producing the largest possible release of heat to the river. Shown in Figure 5 are the measured conditions of the SCCW system during the survey. Included are the discharge and temperature of the SCCW effluent. During the survey, the average discharge of the SCCW system to the river was about 270 cfs. The root-mean-square variation in the SCCW discharge was only about 2 percent of the average—thus, from the standpoint of mixing processes in the river, the discharge was essentially constant. The SCCW effluent temperature decreased throughout the survey from about 86.3°F at the beginning of the survey to about 83.5°F at the end of the survey. This trend coincides with the falling nighttime air temperature, also shown in Figure 5 (note: the discharge temperature of water from the Unit 1 cooling tower, which provides the source of heat for Outfall 113, varies directly with the temperature of the ambient air that is drawn through the tower). Relative to the upstream ambient river temperature, the temperature rise of the Outfall 113 effluent released from the SCCW system, also shown in Figure 5, decreased from about 7.0 F° at the beginning of the survey to about 4.6 F° at the end of the survey.

Downstream End of Passive Mixing Zone

Shown in Figure 6 are the measurements from the HOBO temperature stations at the downstream end of the passive mixing zone. The stations are labeled consecutively from WB1 to WB12, with WB1 situated near the left-hand shoreline of the river and WB12 situated near the right-hand shoreline of the river (i.e., facing downstream—see Figure 2). In Figure 7, the HOBO data has been analyzed to produce contour plots of the local “instantaneous” water temperature rise (ΔT) relative to the SCCW ambient river temperature (i.e., given in Figure 4). The horizontal (x) axis of each contour plot is the span of the river from WB1 to WB12, and the vertical (y) axis is the water depth from 0.5 feet to 7 feet. In this manner, the plots in Figure 7 represent images of the upper 7 feet of the water column in the river, looking downstream. Note that the depth scale in the plots is very distorted so that the data can be viewed in a meaningful manner—that is, whereas the span of the x-axis is about 1000 feet, the span of the y-axis is only about 7 feet

(0.007 times smaller). Plots are provided at the top of each hour from the beginning of the survey at 21:00 CDT on August 30 to the end of the survey at 05:00 CDT on August 31. The following behaviors are emphasized from Figure 6Figure 7:

- At the beginning of the survey, 21:00 CDT on August 30, heat from the SCCW resides primarily on the right-hand-side of the river. Some heat is found in the left-hand-side of the river, perhaps from river sloshing that occurs as a result of deceleration and cessation of the flow at Watts Bar Dam. The maximum local instantaneous temperature rise is about 1.6 F° and occurs in the upper 3 feet of the water column in the right-hand-side of the river.
- Over the next four hours, the temperature rise at the downstream end of the passive mixing zone decreases, and by 01:00 CDT on August 31, the temperature of water in the upper 7 feet of the water column is at most only about 0.4 F° warmer than the ambient water temperature. There is very little temperature variation across the river.
- By 02:00 CDT on August 31, five hours into the survey, heat from the SCCW effluent has arrived in the left-hand-side of the river at the downstream end of the passive mixing zone. That is, in this survey, it took between four and five hours for the leading edge of the SCCW effluent to spread across the river and reach the downstream end of the passive mixing zone.
- In the remaining three hours of the survey, heat from the SCCW effluent slowly backfills from the left-hand-side of the river to the right-hand-side of the river. The maximum local instantaneous temperature rise is about 1.8 F° and occurs in the upper 3 feet of the water column in the left-hand-side of the river. Overall, however, at the end of the survey, 05:00 CDT on August 31, there again is very little temperature variation across the river—at most about 0.4 F°.

NPDES Compliance Parameters

Since heat from the SCCW effluent is distributed across the full width of the river, data from all of the HOBO stations were used to compute the NPDES compliance parameters, which is consistent with the dimensions of the passive mixing zone (i.e., the passive mixing zone spans the full width of the river). The compliance parameters examined include all those given in Table 1—the temperature at the downstream end of mixing zone, T_d ; the temperature rise from upstream to the downstream end of the mixing zone, ΔT ; and the temperature rate-of-change at the downstream end of the mixing zone, TROC. The fundamental equations used to compute the compliance parameters are provided in APPENDIX B, based on the criteria specified in the NPDES permit. The temperature at the downstream end of the mixing zone was determined from the HOBO measurements by averaging the readings from the sensors at depths 3, 5, and 7 feet for all twelve HOBO stations. The temperature rise was computed as the difference between the measured temperature at the downstream end of the mixing zone and the upstream

temperature measured at Watts Bar Dam (i.e., Station 30). The temperature rate-of-change was determined by the change in the measured temperature at the downstream end of the mixing zone from one hour to the next. The data were averaged over a period of one hour using 15-minute readings, as specified in the NPDES permit, and compared with the WBN thermal plume model. The measurements are presented in Figure 8, along with the results obtained by the thermal plume model. The following behaviors are emphasized:

- Temperature at the downstream end of the passive mixing zone, T_d : The maximum 1-hour average T_d estimated by the thermal plume model was 80.8°F, whereas the maximum measured value was about 80.6°F. Thus, the model overpredicted the maximum measured T_d by 0.2°F. Compared to the measurements, the increase in river temperature due to the no flow event was predicted to occur much more rapidly by the model. This is because the model assumes impacts due to changes in the river and/or Outfall 113 conditions are fully realized as a steady-state episode within one hour (i.e., the model time-step); whereas in reality, the actual time for the thermal plume to evolve is much longer. Both the predictions from the model and measurements from the survey were well below the NPDES limit of 86.9°F.
- Temperature rise, ΔT : The maximum 1-hour average ΔT predicted by the plume model was 1.5 F°, whereas the maximum measured value was about 1.6 F°. Thus, the model underpredicted the maximum measured temperature rise by 0.1 F°. For the reason cited above (i.e., computational time-step of one hour), the model predicted the maximum temperature rise to occur one hour into the no flow event. A close examination of the data reveals that the maximum measured value of the temperature rise occurred at end of the survey, when the impact of leakage at Watts Bar Dam reduced the upstream ambient river temperature relative to the model value (see previous discussion in section entitled “River Conditions”). The model value for the upstream ambient river temperature was 79.3°F, whereas due to leakage of cold water at Watts Bar Dam, the measured ambient temperature was unnaturally lowered to 78.9°F (i.e., 0.4 F° lower than the model value, see Figure 4). Both the predictions from the model and measurements from the survey were well below the NPDES limit of 5.4 F°.
- Temperature rate-of-change, TROC: The maximum 1-hour average TROC predicted by the plume model was 0.6 F°/hour, whereas the maximum measured value was about 0.2 F°/hour (absolute values). Thus, the model overpredicted the temperature rate-of-change by 0.4 F°/hour. Both the predictions from the model and measurements from the survey were well below the NPDES limit of ± 3.6 F°/hour.

CONCLUSIONS

The compliance survey for 2011 summer conditions was successful in measuring the NPDES instream water temperature parameters for the Outfall 113. These included the temperature, T_d , temperature rise, ΔT , and temperature rate-of-change, TROC, all at the downstream end of the passive mixing zone. The measurements were compared with values predicted by the thermal plume model that TVA currently uses to determine the safe operation of the SCCW system.

Since 2005, when the first compliance survey was performed for the Outfall 113 passive mixing zone, the model value for the maximum downstream temperature T_d , including that for the survey summarized herein, has always bounded the measured value for the maximum T_d . That is, the model value has always been greater than or equal to the measured value. Such is not the case, however, for ΔT and TROC. In this survey, and for the first time, the model value for the maximum ΔT underpredicted the measured value for the maximum ΔT by 0.1 F°. In the summer survey for 2005, the model value for the maximum TROC underpredicted the measured value for the maximum TROC by 0.3 F°/hour (McCall and Hopping, 2006). These differences are not surprising in light of the fact that the model, like any mathematical representation of an actual complex physical process, contains inherent accuracy limitations. The TVA model for predicting the Outfall 113 thermal plume uses CORMIX, which has a stated accuracy of about 50% of the standard deviation of field measurements (Jirka, et al., 1996). In the survey summarized herein, the difference of 0.1 F° between the model and measured values of the maximum ΔT was not caused by any inadequacy in CORMIX, but by unnatural cooling of the upstream ambient river temperature from leakage of cold water through the hydroturbines at Watts Bar Dam. Based on this, as well as the fact that differences as small as 0.1 F° for ΔT and 0.3 F°/hour for TROC fall within the factor of safety currently used in performing hydrothermal forecasts, the thermal plume model is yet considered fully adequate for determining the safe operation of the SCCW system. That is, in combination with TVA procedures for predicting the impact of the Outfall 113 effluent, the model continues to provide a high level of confidence that the SCCW system is being operated in a manner that does not exceed the instream limits for T_d , ΔT , and TROC as specified in the WBN NPDES permit for the passive mixing zone.

REFERENCES

Harper, Walter L., and Bo Hadjerioua, Mark Reeves, Gary Hickman, and John Jenkinson, "Hydrodynamics and Water Temperature Modeling at Watts Bar SCCW Discharge Structure," TVA Resource Group, Water Management, Report No. WR98-1-85-142, November 1998.

Jirka, Gerhard H., Robert L. Doneker, and Steven W. Hinton, "User's Manual for CORMIX: A Hydrodynamic Mixing Zone Model and Decision Support System for Pollutant Discharges into Surface Waters," Office of Science and Technology, U.S. Environmental Protection Agency, Washington, DC, September 1996.

McCall, Michael J., and P.N. Hopping, "Summer 2005 Compliance Survey for Watts Bar Nuclear Plant Outfall 113 Passive Mixing Zone," TVA River Operations, Report No. WR2006-2-85-152, February 2006.

McCall, Michael J., and P.N. Hopping, "Winter 2005 Compliance Survey for Watts Bar Nuclear Plant Outfall 113 Passive Mixing Zone," TVA River Operations, Report No. WR2005-2-85-151, October 2005.

TDEC, *State of Tennessee NPDES Permit No. TN0020168*, Tennessee Department of Environment and Conservation, Issued June 2010.




Figure 1. Watts Bar Nuclear Plant Outfall 113 (SCCW) Mixing Zones




Figure 2. Location of HOBO Monitoring Stations




Figure 3. Schematic of HOBO Water Temperature Monitoring Stations




Figure 4. River Conditions




Figure 5. SCCW Conditions




Figure 6. HOBO Water Temperature Measurements




Figure 6 (Continued). HOBO Water Temperature Measurements




Figure 7. Instantaneous Temperature Rise for HOB0 Measurements




Figure 7 (Continued). Instantaneous Temperature Rise for HOB0 Measurements




Figure 7 (Continued). Instantaneous Temperature Rise for HOB0 Measurements




Figure 8. Measured and Computed Compliance Parameters for Passive Mixing Zone

APPENDIX A

Calibration of NPDES Water Temperature Sensors



All sensors used by TVA for monitoring compliance of NPDES water temperature requirements are certified and maintained to meet the following industry and regulatory standards:

- ISO/IEC 17025—Quality assurance requirements for the competence to carry out sampling, testing, and calibrations using standard, non-standard, and laboratory-developed methods (ISO=International Organization for Standardization, IEC=International Electrotechnical Commission).
- 10CFR50 Appendix B—Quality assurance criteria for design, fabrication, construction, and testing of the structures, systems, and components of nuclear power plants (CFR=Code of Federal Regulations).
- 40CFR136—Guidelines establishing test procedures for the analysis of pollutants under the Clean Water Act.
- ANSI N45.2. 1971—Quality assurance requirements for Nuclear Power Plants (ANSI=American National Standards Institute).
- ANSI/NCSL Z540-1-1994—General requirements for calibration laboratories and equipment used for measurements and testing (NCSL=National Conference of Standards Laboratories).

The standard used to certify the thermistors for the permanent EDS stations and the temporary HOB0 stations is traceable to the National Institute of Standards and Technology (NIST). The standard includes two pieces of equipment—a platinum resistance temperature detector (RTD) manufactured by Burns Engineering, Inc. and an ohmmeter manufactured by Azonix Inc. The latter is used to measure the resistance of the RTD (i.e., the resistance of platinum varies with temperature). The NIST traceable calibration certificates for the Burns RTD and the Azonix ohmmeter used to calibrate the HOB0 monitors in the field survey summarized herein are available upon request. The overall accuracy of the system for the temperature standard is about $\pm 0.05^{\circ}\text{F}$. The tolerance of the thermistors used for the WBN field survey is about $\pm 0.4^{\circ}\text{F}$, thus providing a calibration test accuracy ratio (TAR) of about 1:8. That is, the accuracy of temperature standard used for the sensor calibrations is about 8 times greater than the minimum acceptable field accuracy of temperature sensors. This is twice the recommended maximum TAR of 1:4 for sensor calibrations.

The TVA procedure to calibrate the HOB0 water temperature monitors, Instruction No. 450.01-020, is provided below. Briefly, the HOB0 monitors are immersed in a stirred temperature-

controlled water bath along with the standard (i.e., along with the Burns RTD probe). After the bath stabilizes, temperature readings from the HOBO monitors are compared to the temperature readings from the standard. Experience has shown that in nearly all cases, the readings from both the HOBO monitors and the standard are essentially constant, so that the 95 percent confidence interval of the readings is diminutive. Under these conditions, the accuracy of each HOBO monitor is recorded simply as the difference between the HOBO reading and that of the standard (negative difference = HOBO reading low/below standard, positive difference = HOBO reading high/above standard). The HOBO monitors are tested at three temperatures between 30°F and 100°F, covering the range of expected water temperature for natural river conditions. The three temperatures are at about the 10 percent, 50 percent, and 90 percent intervals, or 37°F, 65°F and 93°F, respectively. Any HOBO monitor with measured accuracy in excess of the maximum allowable tolerance of $\pm 0.4^\circ\text{F}$ for any one of the three temperatures fails the calibration test and is removed from the field survey inventory. The calibration certificates for HOBO monitors used in this field survey summarized herein are available upon request. All the HOBO monitors passed both the pre-survey and post-survey calibration tests. The mean square error of the HOBO monitors was 0.14°F for both the pre-survey and post-survey calibrations.

| | | |
|---|---|--|
|  CENTRAL LABORATORIES SERVICES QUALITY PROGRAM INSTRUCTION | TITLE Certification of HOBO Water Temp Pro Data Acquisition Systems H ₂ O-001 | Instruction No. 450.01-020 Rev. No. 0 Page No. 1 of 7 |
| | | Effective Date <u>5/19/03</u> |
| LEVEL OF USE <input type="checkbox"/> Continuous <input checked="" type="checkbox"/> Reference <input type="checkbox"/> Information | | |
| QA RECORD | | |
| <u>Dennis T. Darby</u> Preparer | | <u>5/19/03</u> Date |
| <u>Paul B. Loiseau, Jr.</u> Technical Reviewer | | <u>5/19/03</u> Date |
|  Administrative Review | | <u>6/5/03</u> Date |
| APPROVAL | | |
| <u>Jerry D. Hubble</u> Department Manager | | <u>5/19/03</u> Date |

| | |
|--|--|
| TITLE: Certification of HOBO Water Temp Pro Data Acquisition Systems H₂O-001 | Instruction No. 450.01-020 Rev. 0 Eff. Date 5/19/03 Page 3 of 7 |
|--|--|

1.0 PURPOSE

To provide uniform and effective certifications of Hobo Water Temp Pro data acquisition systems meeting the accuracy and performance requirements of TVA's water temperature-monitoring programs. This technical instruction uses the method of comparison with a laboratory standard thermometer.

2.0 SCOPE

This instruction applies to the certification of Hobo Water Temp Pro data loggers manufactured by Onset Computer Corporation of Bourne, Massachusetts. The Hobo Water Temp Pro is a data acquisition system containing a temperature sensor, data logger and battery sealed in a single submersible case. The Hobo Water Temp Pro is programmed and data retrieved by use of an infrared interface located in one end of the case. Hobo Water Temp Pros are certified upon receipt from the manufacturer at no greater than 12 month intervals during use or when requested.

3.0 SUMMARY

In this three-point certification systems are tested as actually used over the historical water temperature range of 30° to 100°F and submerged in water. The three test points are 37°, 65° and 93°F. The systems are required to perform within Onset Computer Corporation tolerances. System conformity at each temperature point is determined by comparing system temperature, logged by the Hobo Water Temp Pro and a laboratory standard thermometer.

Systems are programmed and submerged with a standard thermometer in a stirred, temperature-controlled temperature bath. The systems are read after the test by an infrared interface adapter connected to a computer running Onset Computer Corporation's Boxcar Pro software. Traceability of the certification is through the thermometer.

"As-found" certifications are performed on new systems as an acceptance test and on sensors returned from field service. "As-left" certifications are performed before delivery for field service if more than 12 months has elapsed since the last certification. "As-found" and "as-left" certifications may be combined on the same record if there is clear indication which type each system is undergoing.

Multiple HOBOS may be certified at the same time in the temperature bath.

| | |
|--|--|
| TITLE: Certification of HOBO Water Temp Pro Data Acquisition Systems H₂O-001 | Instruction No. 450.01-020 Rev. 0 Eff. Date 5/19/03 Page 4 of 7 |
|--|--|

- Accuracy of $\pm 0.2^{\circ}\text{C}$ at 25°C (0.33°F at 70°F)
- Waterproof case, submersible to 100 feet
- Capacity to store up to 21,580 temperature measurements
- Selectable sampling interval from 1 second to 9 hours
- Programmable start time/date
- Two data recording modes: Stop when full or wrap around when full.
- Two data offload modes: Halt then offload or offload while logging.
- Nonvolatile EEPROM memory that retains data even if batteries fail
- Light-emitting diode (LED) operation, indicator, which can be disabled during logging by selecting "Stealth" mode
- High-speed IR communications for offloading data; can readout full logger in less than 30 seconds while logging continues
- Battery life of 6 years with typical usage

4.0 PRACTICES/EXCEPTIONS

N/A

5.0 SAFETY

- 5.1 Standard electrical equipment safety.

6.0 STANDARDS USED

- 6.1 Laboratory reference thermometer, range 30° to 100°F or greater, 0.01°F resolution, 0.1°F accuracy or better, with current calibration sticker.

7.0 EQUIPMENT/APPARATUS

- 7.1 Temperature bath, stirred, temperature-controlled.
7.2 Computer with Onset Boxcar Pro software installed (version 4.3 or later)
7.3 IR Base station, Onset Part # BST -IR

8.0 PREREQUISITE ACTIONS

- 8.1 Turn on temperature bath and set for 37°F .
8.2 Check the IR interface to verify that it is plugged into the correct serial port on the PC. Set the correct time on the PC.
8.3 Align the IR port on the Base station with the HOBO Water Temp Pro communications window. Place the logger no further than 4 to 5 inches away from the Base station (see Figure 2) and make sure the IR windows in both devices point at each other. There is a 30° acceptance angle for the IR beam, so some misalignment is acceptable.

| | |
|---|--|
| TITLE: Certification of HOBO Water Temp Pro Data Acquisition Systems H20-001 | Instruction No. 450.01-020 Rev. 0 Eff. Date 5/19/03 Page 5 of 7 |
|---|--|

- 8.4 Start the Onset Box Car Software and select **Logger** then **Hobo Water Temp Pro** and **Launch**.
- 8.5 The computer will respond with a list of loggers found. The serial number in this list should match the serial number printed on the side of the logger. If these numbers do not match, click the **Refresh** button. Record this serial number on the certification form. Either wait or click the **Stop Searching** button. Using the mouse select the logger and click the **Launch** button.
- 8.6 After a few seconds the screen will display the status of the HOBO Water Temp Pro. Record the battery percentage on the certification form.
- 8.7 Verify that the Hobo is set to Fahrenheit and program it to a recording interval of 0:1:0 for a reading once a minute. Verify that the start logging immediately box is checked and that the set data logger clock with host launch is also checked.
- 8.8 Using the mouse click the **Launch Immediately** button.
- 8.9 If last HOBO is programmed click the **DONE** button, else select the **Launch Another** and repeat steps 8.5 through 8.9.
- 9.0 **TEST PROCEDURE/METHOD**
- 9.1 On the certification form record the serial number of the laboratory reference thermometer.
- 9.2 Place the HOBO Water Temp Pro in the temperature bath, making sure the end opposite the IR windows is submerged, and allow the bath to stabilize at 37°F ±0.5°F on the thermometer. Adjust the bath set point if needed. After the bath reaches the desired temperature allow 20 minutes 'soak time' for the HOBO to reach its final temperature.
- 9.3 Record the thermometer reading on the certification form and the time. (The time will be needed to get the correct reading from the HOBO.)
- 9.4 Repeat steps 9.2 and 9.3 for bath settings of 65.0°F ± 0.5°F and 93°F ± 0.5°F.
- 9.5 Remove the HOBO from the temperature bath and align the IR port on the Base station with the HOBO Water Temp Pro communications window.
- 9.6 Restart Onset BoxCar Pro if it is not running and select **Logger** then **Hobo Water Temp Pro** and **Readout**.
- 9.7 The computer will respond with a list of loggers found. Using the mouse select the logger and click the **Readout** button. The computer will ask to download data and continue logging or the stop logging and offload data. Select the **Stop Logging and Offload data**. After a few seconds the computer will respond with a suggested file name. Select **Save** and allow the HOBO to transfer the data.
- 9.8 After a successful download click the **OK** button. The computer will then ask if the data should be displayed in Centigrade or Fahrenheit. Deselect °C and select °F and click **OK**. The computer should display a graph of the collected data. Click the view details button (this is the button just left of the question mark button.)

| | |
|--|--|
| TITLE: Certification of HOBO Water Temp Pro Data Acquisition Systems H₂O-001 | Instruction No. 450.01-020 Rev. 0 Eff. Date 5/19/03 Page 6 of 7 |
|--|--|

9.9 Scroll down the displayed list until the time recorded for the 37°F point is found. Record the corresponding temperature on the certification form. Repeat this step for 65° and 93°.

9.10 Close the view details windows and repeat steps 9.6 through 9.9 for additional HOBOs.

9.11 Fill out the rest of the certification form.

10.0 ACCEPTANCE CRITERIA

10.1 Based upon the manufacturer specifications the HOBO Water Temp Pro should be within $\pm 0.4^{\circ}\text{F}$ over the range of 32°F to 100°F. Any HOBO with an error of greater than $\pm 0.5^{\circ}\text{F}$ at any of the three measured points shall fail certification.

11.0 POST PROCEDURE ACTIVITY

11.1 Close the BoxCar Software.

12.0 RECORDS

12.1 Completed HOBO Water Temperature Pro Certification form and associated Report of Certification cover sheet is a QA record.

13.0 REFERENCE

13.1 HOBO Water Temp Pro User's Manual, version 1.0 or later

13.2 Onset BoxCar Pro4 Manual Version 1.0 or later

APPENDIX B
WBN Outfall 113 NPDES Compliance Parameters

- Current Instantaneous Upstream Temperature:

Tu_i (measured at EDS Station 30 by the first sensor below a depth of 5 feet).

- Current 1-Hour Average Upstream Temperature:

$$Tu1_i = \frac{Tu_i + Tu_{i-1} + Tu_{i-2} + Tu_{i-3} + Tu_{i-4}}{5},$$

where the subscripts i, i-1, i-2, i-3, and i-4 denote the current and previous four 15-minute (0.25 hour) values of Tu .

- Current Instantaneous Downstream Temperature:

$$Td_i = \frac{Td3_i + Td5_i + Td7_i}{3},$$

where $Td3_i$, $Td5_i$, and $Td7_i$ denote the current measurements of river temperature at the downstream end of the mixing zone at water depths 3 feet, 5 feet, and 7 feet, respectively.

- Current 1-Hour Average Downstream Temperature:

$$Td1_i = \frac{Td_i + Td_{i-1} + Td_{i-2} + Td_{i-3} + Td_{i-4}}{5},$$

where the subscripts i, i-1, i-2, i-3, and i-4 denote the current and previous four 15-minute (0.25 hour) values of Td .

- Current Instantaneous Temperature Rise:

$$\Delta T_i = Td_i - Tu_i.$$

- Current 1-Hour Average Temperature Rise:

$$\Delta T1_i = \frac{\Delta T_i + \Delta T_{i-1} + \Delta T_{i-2} + \Delta T_{i-3} + \Delta T_{i-4}}{5},$$

where the subscripts i , $i-1$, $i-2$, $i-3$, and $i-4$ denote the current and previous four 15-minute (0.25 hour) values of ΔT .

- Current Temperature Rate-of-Change:

$$TROC_i = \frac{Td_i - Td_{i-4}}{1 \text{ hour}}.$$

- Current 1-Hour Average Temperature Rate-of-Change:

$$TROC1_i = \frac{TROC_i + TROC_{i-1} + TROC_{i-2} + TROC_{i-3} + TROC_{i-4}}{5},$$

where the subscripts i , $i-1$, $i-2$, $i-3$, and $i-4$ denote the current and previous four 15-minute (0.25 hour) values of TROC.

Enclosure 4

**Winter 2011 Compliance Survey for Watts Bar Nuclear Plant Outfall Passive
Mixing Zone**

TENNESSEE VALLEY AUTHORITY
River Operations

**WINTER 2011 COMPLIANCE SURVEY FOR WATTS BAR
NUCLEAR PLANT OUTFALL 113 PASSIVE MIXING ZONE**

Prepared by

Brandin L. Ruth
and
Paul N. Hopping

Knoxville, Tennessee
December 2011



EXECUTIVE SUMMARY

The National Pollutant Discharge Elimination System (NPDES) Permit No. TN0020168 for Watts Bar Nuclear Plant (WBN) identifies the discharge of water to the Tennessee River from the Supplemental Condenser Cooling Water (SCCW) System as Outfall 113. Furthermore, the permit identifies that when there is no flow released from Watts Bar Dam (WBH), the effluent from Outfall 113 shall be regulated based on a passive mixing zone extending in the river from bank-to-bank and 1,000 feet downstream from the outfall. Compliance with the requirements for the passive mixing zone is to be achieved by two annual instream temperature surveys—one for winter conditions and one for summer conditions. Summarized in this report are the measurements, analyses, and results for the passive mixing zone survey conducted for 2011 winter conditions. The survey was conducted between 23:00 CDT on June 2 and 06:00 CDT on June 3 (seven hours) and included the collection of temperature data at twelve temporary monitoring stations deployed across the downstream edge of the passive mixing zone during a period of no flow in the river. The data were analyzed to compute three compliance parameters: the 1-hour average temperature at the downstream edge of mixing zone, T_d ; the 1-hour average temperature rise from upstream to the downstream edge of the mixing zone, ΔT ; and the 1-hour average temperature rate-of-change at the downstream edge of the mixing zone, TROC. The measured parameters were compared to predicted values from the thermal plume model used by TVA to help determine the safe operation of Outfall 113. The results of the comparisons, in terms of maximum values observed during the no flow event, are as follows:

| Compliance Parameter | Model | Measured | NPDES Limit |
|----------------------|-------------|-------------|-------------|
| Maximum T_d | 72.5°F | 70.8°F | 86.9°F |
| Maximum ΔT | 4.0 F° | 1.7 F° | 5.4°F |
| Maximum TROC | 1.0 F°/hour | 0.6 F°/hour | 3.6 F°/hr |

As shown, values predicted by the model were larger than those measured in the survey. Thus, for the conditions of the survey, the plume model was found to be good for enforcing the operation of Outfall 113 at levels of T_d , ΔT , and TROC below the NPDES limits. For T_d and ΔT , these results are consistent with those of all the previous surveys for the passive mixing zone. For TROC, however, previous surveys have revealed that the model is capable of underpredicting measured values for TROC by as much as 0.3 F°/hour (e.g., see McCall and Hopping, 2006). Under these conditions, a factor of safety of 0.3 F°/hour currently is used for tracking TROC in the operation of the SCCW system. That is, for the passive mixing zone, the safe operation of Outfall 113 is evaluated based on a maximum allowable value of TROC from the thermal plume model of ± 3.3 F°/hour rather than ± 3.6 F°/hour. This practice will continue until further notice.

TABLE OF CONTENTS

| | <u>Page No.</u> |
|-------------------------|------------------------|
| EXECUTIVE SUMMARY | i |
| INTRODUCTION | 1 |
| INSTREAM SURVEY | 2 |
| RESULTS | 3 |
| River Conditions | 3 |
| SCCW Conditions..... | 3 |
| Effluent Behavior..... | 4 |
| CONCLUSIONS..... | 7 |
| REFERENCES | 8 |
| APPENDIX A..... | 17 |
| APPENDIX B | 62 |

LIST OF FIGURES

| | |
|---|----|
| Figure 1. Watts Bar Nuclear Plant Outfall 113 (SCCW) Mixing Zones | 9 |
| Figure 2. Location of HOBO Monitoring Stations | 10 |
| Figure 3. Schematic of HOBO Water Temperature Monitoring Stations | 10 |
| Figure 4. River Conditions..... | 11 |
| Figure 5. SCCW Conditions | 12 |
| Figure 6. HOBO Water Temperature Measurements During Survey..... | 13 |
| Figure 7. Profiles of Instantaneous Compliance Temperature across Downstream End of Passive Mixing Zone..... | 15 |
| Figure 8. Measured and Computed Compliance Parameters for Passive Mixing Zone | 16 |

LIST OF TABLES

| | |
|--|---|
| Table 1. NPDES Temperature Limits for Outfall 113 Mixing Zones | 1 |
| Table 2. Sources of Data for Passive Mixing Zone Survey | 2 |

WINTER 2011 COMPLIANCE SURVEY FOR WATTS BAR NUCLEAR PLANT OUTFALL 113 PASSIVE MIXING ZONE

INTRODUCTION

Outfall 113 for the Watts Bar Nuclear Plant (WBN) includes the discharge of water to the Tennessee River from the Supplemental Condenser Cooling Water (SCCW) system. Due to the dynamic behavior of the thermal effluent in the river, the National Pollutant Discharge Elimination System (NPDES) Permit No. TN0020168 for the plant specifies two mixing zones for Outfall 113—one for active operation of the river and one for passive operation of the river (TDEC, 2010). The passive mixing zone corresponds to periods when the operation of Watts Bar Dam (WBH) produces no flow in the river (i.e., hydropower and/or spillway releases). The dimensions of the passive mixing zone extend from bank-to-bank and downstream 1,000 feet from the outfall. The active mixing zone applies to all other river flow conditions. The dimensions of the active mixing zone include the right-half of the river (facing downstream) and extend downstream 2,000 feet from the outfall. The passive and the active mixing zones are illustrated in Figure 1.

Table 1 summarizes the NPDES temperature limits for Outfall 113. The limits apply to both the active and passive mixing zones. Compliance for the active mixing zone is monitored by permanent instream water temperature stations situated in the right-half of the river. Due to navigation issues associated with placing permanent stations in the left-half of the river, a thermal plume model is used to determine the safe operation of Outfall 113 for the passive mixing zone. To verify the thermal plume model, the NPDES permit specifies that two instream temperature surveys shall be conducted each year—one for winter conditions and one for summer conditions. The purpose of this report is to present the results for the passive mixing zone temperature survey conducted for winter 2011 conditions. Provided is a brief summary of the survey method, presentations of the measurements and analyses, and discussions for the results and conclusions.

Table 1. NPDES Temperature Limits for Outfall 113 Mixing Zones

| Compliance Parameter | Sampling Period | NPDES Limit |
|--|------------------------|--------------------|
| Maximum Temperature, Downstream Edge of Mixing Zone, T_d | Running 1-hr | 86.9°F |
| Maximum Temperature Rise, Upstream to Downstream, ΔT | Running 1-hr | 5.4 F° |
| Maximum Temperature Rate-of-Change, TROC | Running 1-hr | ±3.6 F°/hr |

The survey was conducted between 23:00 CDT on June 2 and 06:00 CDT on June 3 (seven hours). The winter survey usually is conducted in March or April when the ambient river temperature is cool, but when daytime air temperatures can be high. These conditions produce

above normal effluent temperatures from Outfall 113. That is, TVA prefers to evaluate the outfall at a time when the effluent from the SCCW system “challenges” the method used by TVA to monitor compliance for the outfall. In 2011, due to high rainfall, TVA was in a flood control operation at Watts Bar Dam during most of March. Under these conditions, river flow could not be discontinued for the purpose of a survey. Then in early April, WBN was removed from service for a routine refueling and maintenance outage. During the outage, Outfall 113 was not thermally loaded. For these reasons, the winter survey was not conducted until early June, when the flood operation had expired and the plant had returned to service with a sustained level of generation.

INSTREAM SURVEY

The instream survey included the deployment of temporary water temperature stations at twelve locations across the downstream edge of the passive mixing zone. Data from these and other monitoring stations were analyzed to obtain measured values for the compliance parameters listed in Table 1. These were then compared with the corresponding values estimated from the SCCW thermal plume model.

The method of conducting the instream survey is the same as that used for the first such survey, performed for winter conditions on May 6, 2005 (McCall and Hopping, 2005). Table 2 provides a summary of the sources of data for the survey. WaterView, a monitoring system for tracking hydroplant operation and performance, was used to obtain measurements for the river discharge from Watts Bar Dam. The WBN Environmental Data Station (EDS) provided measurements from existing permanent monitoring stations for the nuclear plant. These included the upstream (ambient) river temperature, river water surface elevation, SCCW effluent temperature, SCCW effluent discharge, and air temperature.

Table 2. Sources of Data for Passive Mixing Zone Survey

| Data | Source | Frequency |
|---------------------------------------|---------------------------------------|------------------|
| River ambient water temperature | WBN EDS Station 30 (Tailwater at WBH) | 15 min |
| River water surface elevation | WBN EDS Station 30 (Tailwater at WBH) | 15 min |
| SCCW effluent temperature | WBN EDS Station 32 (Outfall 113) | 15 min |
| SCCW effluent discharge | WBN EDS Station 32 (Outfall 113) | 15 min |
| Air temperature | WBN EDS Met Tower | 15 min |
| Passive mixing zone water temperature | Temporary HOBO Monitors | 1 min |

The water temperature at the downstream edge of the Outfall 113 passive mixing zone was measured by the temporary water temperature stations. The stations were positioned at roughly equal intervals across the river, as shown in Figure 2, using a Global Positioning System (GPS) device. The temporary stations recorded water temperatures by using HOBO temperature

monitors positioned at depths of 0.5, 3, 5, and 7 feet below the water surface. Shown in Figure 3 is a schematic of the temporary stations. The stations included a string of HOBO monitors suspended from a tire float, with weights to anchor the station and to keep the sensor string vertical in the water column. The water temperature sensors used in the HOBO monitors had an accuracy of about ± 0.4 F° and resolution of about 0.04 F°, which is consistent with other temperature sensors used by TVA for tracking hydrothermal compliance. The HOBO monitors include an internal data acquisition unit that was programmed to collect measurements once per minute. All the temperature probes used in the survey, including those contained in the HOBO monitors and the thermistors at the permanent EDS monitoring stations, were calibrated by a quality program with equipment accuracies traceable to the National Institute of Standards and Technology (NIST). The calibration procedure is summarized in Appendix A. The temporary monitoring stations were deployed several hours before the beginning of the survey, and retrieved several hours after the end of the survey.

RESULTS

River Conditions

Figure 4 shows the measured ambient conditions of the river during the survey. Included are the river discharge at Watts Bar Dam, the river water surface elevation (WSEL) at the exit of Watts Bar Dam, and river temperature at the exit of Watts Bar Dam. The river temperature at the exit of Watts Bar Dam serves as the upstream ambient river temperature for WBN Outfall 113. To provide a period of no flow in the river, releases from Watts Bar Dam were suspended between about 23:00 CDT on June 2 and 06:00 CDT on June 3, a total of seven hours (nighttime). Leading up to the survey, as the river flow was stepping down, the water surface elevation below Watts Bar Dam dropped approximately 0.8 feet. During the survey, the elevation slowly increased due to filling (i.e., backflow) from the surrounding tailwater, reaching a value of about 682 feet msl at the end of the survey. The ambient river temperature was 68.5°F at the beginning of the survey and increased to 69.1°F by the end of the survey. In June, the ambient river temperature often increases in this manner because the temperature of the bottom water released through the hydroturbines (before the onset of the no flow event) usually is cooler than that of the surrounding tailwater, which is warmed by daytime solar heating.

SCCW Conditions

During the survey, the SCCW system at WBN was thermally loaded and operating in “summer” mode. That is, the system was operating in a manner producing the largest possible heat load to the river. Shown in Figure 5 are the measured conditions of the SCCW system during the survey. Included are the discharge and temperature of the SCCW effluent. Due to an unexpected outage of data acquisition equipment, the measurement for the SCCW discharge was

unavailable between 22:30 CDT on June 2 and 04:00 CDT on June 3. However, since WBN was operating in a near steady manner throughout the survey, it is known that the SCCW discharge, in like manner, was near steady. Based on data collected in the hours immediately before and after the equipment outage, the average SCCW discharge during the survey was estimated to be about 294 cfs. The SCCW effluent temperature decreased throughout the survey from about 87.1°F at the beginning of the survey to about 83.7°F at the end of the survey. This trend coincides with the falling nighttime air temperature, also shown in Figure 5 (note: the discharge temperature of water from the Unit 1 cooling tower, which provides the source of heat for Outfall 113, varies directly with the temperature of the ambient air that is drawn into the tower). Relative to the upstream ambient river temperature, the temperature rise of the Outfall 113 effluent released from the SCCW system, also shown in Figure 5, decreased from about 18.6 F° at the beginning of the survey to about 14.6 F° at the end of the survey.

Effluent Behavior

Individual Temperature Stations

Shown in Figure 6 are the measurements from the HOBO temperature stations at the downstream end of the passive mixing zone. The stations are labeled consecutively from WB1 to WB12, with WB1 situated near the left-hand shoreline of the river and WB12 situated near the right-hand shoreline of the river (i.e., facing downstream—see Figure 2). The following behaviors are noted:

- At the beginning of the survey, between 23:00 CDT and 23:30 CDT on June 2, stations WB2 through WB4 had to be removed from the navigation channel to allow passage of a tow. No data is available from these stations during this time.
- In the first three hours of the survey, temperature undulations at the 0.5 foot depth were more intense than in previous surveys. This perhaps was due to large-scale “swirls” created in the surface layer of the river by the passing tow, as well as water released upstream as part of the operation of the navigation lock. The undulations are attenuated at larger depths.
- It took about three hours for the leading edge of the SCCW effluent to spread across the river and reach the downstream edge of the passive mixing zone. This is observed by the increase in temperature that begins for all stations at about 02:00 CDT on June 3. The increase is more sudden in the left-hand-side of the river than in the right-hand-side of the river (i.e., WB1 through WB6 verses WB7 through WB12). This is because in a no flow situation, the effluent traverses across the river as it transported downstream more rapidly along the left-hand shoreline.

- In the remaining hours of the survey, the temperature at all stations slowly increased—as much as 3 F° at the 0.5-foot depth, and as much as 1.5 F° at the 7-foot depth. The smaller increase at the 7-foot depth suggests that for the prevailing conditions of the river and WBN, most of the thermal effluent from Outfall 113 resided in the surface layer of the water column (i.e., the bottom layer of the river is protected).

Distribution Across The Mixing Zone

At each HOBO station, the instantaneous compliance temperature was determined by averaging the measurements for the sensors at the 3-foot, 5-foot, and 7-foot depths. Plotted in Figure 7 are the resulting temperatures across the downstream end of the passive mixing zone, measured at the top of each hour from 23:00 CDT on June 2 to 06:00 CDT on June 3. The following behaviors are noted:

- As previously stated, between 23:00 CDT and 23:30 CDT on June 2, stations WB2 through WB4 had to be removed from the navigation channel to allow passage of a tow. As such, no data is shown for these stations for 23:00 CDT.
- During the first hour of the survey, the temperatures at WB1 through WB6 decreased about 0.5°F. Then, between 00:00 CDT and 01:00 CDT, the temperature at all the stations remained fairly constant with only small variations, typically between 0.2°F to 0.3°F.
- Between 01:00 CDT and 02:00 CDT, the temperature at WB1 through WB3 increased by about 1.0°F, indicating the arrival of the leading edge of the SCCW effluent at the downstream, left-hand-side of the passive mixing zone.
- By 03:00 CDT, the effluent had spread across the entire width of the river (at the downstream end of the passive mixing zone). Over the remainder of the survey, from 04:00 CDT to 06:00 CDT, temperatures continued to increase, on the average climbing about an additional 0.3°F. The temperature for stations WB3 through WB5 were somewhat higher, suggesting the center of the effluent plume resided in the left-hand-side of the river.

Compliance Parameters

Since heat from the SCCW effluent is distributed across the full width of the river, data from all of the HOBO stations were used to compute the NPDES compliance parameters, which is consistent with the dimensions of the passive mixing zone (e.g., as shown in Figure 1). The compliance parameters examined include those given in Table 1—the temperature at the downstream edge of mixing zone, T_d ; the temperature rise from upstream to the downstream edge of the mixing zone, ΔT ; and the temperature rate-of-change at the downstream edge of the mixing zone, TROC. The fundamental equations used to compute the compliance parameters

are provided in Appendix B, based on the criteria specified in the NPDES permit. The temperature at the downstream end of the mixing zone was determined from the HOBO measurements (i.e., average of sensors at depths 3, 5, and 7 feet for all twelve HOBO stations). The temperature rise was computed as the difference between the temperature at the downstream end of the mixing zone and the upstream temperature measured at Station 30. The temperature rate-of-change was determined by the change in the temperature at the downstream end of the mixing zone from one hour to the next. The data were averaged over a period of one hour using 15-minute readings, as specified in the NPDES permit, and compared with the WBN thermal plume model. The results are presented in Figure 8, along with the results obtained by the thermal plume model. The following comments are provided.

- Temperature at the downstream edge of the passive mixing zone, T_d : The maximum 1-hour average T_d estimated by the thermal plume model was 72.5°F, whereas the maximum measured value was about 70.8°F. Thus, the model overpredicted the maximum measured T_d by 1.7°F. Compared to the measurements, the increase in river temperature due to the no flow event was predicted to occur much more rapidly by the model. This is because the model assumes impacts due to changes in the river and/or Outfall 113 are fully realized within one hour (i.e., the model time-step); whereas in reality, the actual time for the development of these impacts is much longer, at least for events with little or no river flow. Both the predictions from the model and measurements from the survey were well below the NPDES limit of 86.9°F.
- Temperature rise, ΔT : The maximum 1-hour average ΔT predicted by the plume model was 4.0 F°, whereas the maximum measured value was about 1.7 F°. Thus, the model overpredicted the maximum measured temperature rise by 2.3 F°. For the reason cited above (i.e., computational time-step of one hour), the model predicted the temperature rise to occur sooner than that found by the measurements. Both the predictions from the model and measurements from the survey were well below the NPDES limit of 5.4 F°.
- Temperature rate-of-change, TROC: The maximum 1-hour average TROC predicted by the plume model was 1.0 F°/hour, whereas the maximum measured value was about 0.6 F°/hour (absolute values). Thus, the model overpredicted the temperature rate-of-change by 0.4 F°/hour. Both the predictions from the model and measurements from the survey were well below the NPDES limit of ± 3.6 F°/hour.

CONCLUSIONS

The survey for 2011 winter conditions was successful in measuring the NPDES water temperature parameters for the Outfall 113 passive mixing zone. The measurements were compared with values predicted by the thermal plume model that TVA currently uses to judge the safe operation of the SCCW system. Overall, for the conditions of the 2011 winter survey, the model was found to be good for estimating the potential impact of Outfall 113 on the temperature, T_d , temperature rise, ΔT , and temperature rate-of-change, TROC, at the downstream end of the passive mixing zone. This is because the model overpredicted, or bounded, the maximum values measured for T_d , ΔT , and TROC. In this manner, for the conditions of the 2011 winter survey, the thermal plume model assured the operation of Outfall 113 at levels of T_d , ΔT , and TROC below the NPDES limits. For T_d and ΔT , these results are consistent with those for all of the previous surveys for the passive mixing zone. The same is not true, however, for TROC. Previous surveys have revealed that the model is capable of underpredicting measured values for TROC by as much as 0.3 F°/hour (e.g., see McCall and Hopping, 2006). Under these conditions, and despite the results summarized herein, a factor of safety of 0.3 F°/hour currently is used for tracking TROC in the operation of the SCCW system. This is accomplished by limiting the operation of Outfall 113 for the passive mixing zone based on a maximum allowable value of TROC from the thermal plume model of ± 3.3 F°/hour rather than ± 3.6 F°/hour.

REFERENCES

McCall, Michael J., and P.N. Hopping, "Summer 2005 Compliance Survey for Watts Bar Nuclear Plant Outfall 113 Passive Mixing Zone," TVA River Operations, Report No. WR2006-2-85-152, February 2006.

McCall, Michael J., and P.N. Hopping, "Winter 2005 Compliance Survey for Watts Bar Nuclear Plant Outfall 113 Passive Mixing Zone," TVA River Operations, Report No. WR2005-2-85-151, October 2005.

TDEC, *State of Tennessee NPDES Permit No. TN0020168*, Tennessee Department of Environment and Conservation, Issued June 2010.




Figure 1. Watts Bar Nuclear Plant Outfall 113 (SCCW) Mixing Zones




Figure 2. Location of HOBO Monitoring Stations




Figure 3. Schematic of HOBO Water Temperature Monitoring Stations




Figure 4. River Conditions




Figure 5. SCCW Conditions




Figure 6. HOBO Water Temperature Measurements During Survey




Figure 6 (Continued). HOBO Water Temperature Measurements During Survey




Figure 7. Profiles of Instantaneous Compliance Temperature across Downstream End of Passive Mixing Zone
(Average of Readings at 3-Foot, 5-Foot, and 7-Foot Depths)




Figure 8. Measured and Computed Compliance Parameters for Passive Mixing Zone

APPENDIX A

(The following information is provided per request of Mike Kelly of TDEC on August 26, 2008)

All sensors used by TVA for monitoring compliance of NPDES water temperature requirements are certified and maintained to meet the following industry and regulatory standards:

- ISO/IEC 17025—Quality assurance requirements for the competence to carry out sampling, testing, and calibrations using standard, non-standard, and laboratory-developed methods (ISO=International Organization for Standardization, IEC=International Electrotechnical Commission).
- 10CFR50 Appendix B—Quality assurance criteria for design, fabrication, construction, and testing of the structures, systems, and components of nuclear power plants (CFR=Code of Federal Regulations).
- 40CFR136—Guidelines establishing test procedures for the analysis of pollutants under the Clean Water Act.
- ANSI N45.2. 1971—Quality assurance requirements for Nuclear Power Plants (ANSI=American National Standards Institute).
- ANSI/NCSL Z540-1-1994—General requirements for calibration laboratories and equipment used for measurements and testing (NCSL=National Conference of Standards Laboratories).

The standard used to certify the thermistors for the permanent EDS stations and the temporary HOBO stations is traceable to the National Institute of Standards and Technology (NIST). The standard includes two pieces of equipment—a platinum resistance temperature detector (RTD) manufactured by Burns Engineering, Inc. and an ohmmeter manufactured by Azonix Inc. The latter is used to measure the resistance of the RTD (i.e., the resistance of platinum varies with temperature). The NIST traceable calibration certificates for the Burns RTDs and the Azonix ohmmeter that were used to calibrate the HOBO probes are provided below. The end result of the RTD calibration is a set of International Temperature Scale 1990 (ITS 90) coefficients that are used to compute water temperature from the measured RTD resistance. Based on the calibration certificates, the accuracy of the system for the temperature standard is about $\pm 0.05^{\circ}\text{F}$. The tolerance of the thermistors used for the WBN field survey is about $\pm 0.4^{\circ}\text{F}$, thus providing a calibration test accuracy ratio (TAR) of about 1:8. That is, the accuracy of temperature standard used for the sensor calibrations is 8 times greater than the minimum acceptable field accuracy of temperature sensors. This is twice the recommended maximum TAR of 1:4 for sensor calibrations.

The TVA procedure to calibrate the HOBO water temperature probes, Instruction No. 450.01-020, is provided below. Briefly, the HOBO probes are immersed in a stirred temperature-controlled water bath along with the standard (i.e., along with the Burns RTD probe). After the bath stabilizes, temperature readings from the HOBO probes are compared to the temperature readings from the standard. Experience has shown that in nearly all cases, the readings from both the HOBO probes and the standard are essentially constant, so that the 95 percent confidence interval of the readings is diminutive. Under these conditions, the accuracy of each HOBO probe is recorded simply as the difference between the HOBO reading and that of the standard (negative difference = HOBO reading low/below standard, positive difference = HOBO reading high/above standard). The HOBO probes are tested at three temperatures between 30°F and 100°F, covering the range of expected water temperature for natural river conditions. Specifically, the three temperatures are at about the 10 percent, 50 percent, and 90 percent intervals, or 37°F, 65°F and 93°F, respectively. Any HOBO probe with measured accuracy (i.e., difference) in excess of the maximum allowable tolerance of $\pm 0.4^{\circ}\text{F}$ for any one temperature fails the calibration test and is removed from the field survey inventory. In general, based on TVA experience, most HOBO probes that pass the calibration test usually have measured accuracies better than about $\pm 0.25^{\circ}\text{F}$ for all three temperatures examined in the bath tests. The calibration certificates for HOBO probes used in field survey summarized herein are provided below. Included are certificates for both the pre- and post-survey calibration tests. A close examination of the certificates shows that all the HOBO probes passed the calibration test both before and after the field survey.

Calibration Certificates for Burns Platinum Resistance Thermometer (RTD)

RTD ID No. 906535 was used for both pre-survey and post-survey calibrations.



LAB STANDARD REPORT of CALIBRATION

Tennessee Valley Authority
Central Laboratories Services

Mailing Address: 1101 Market Street, PSC-1B-C, Chattanooga, TN 37402

Shipping Address: 4601 North Access Road, Bldg. A, Chattanooga, TN 37415

Phone: (423) 876-4318 Fax: (423) 876-4137

Customer:

CLS KNOXVILLE
400 W. SUMMIT HILL DR.
KNOXVILLE, TN 37902

Asset ID: 906535
Certificate No: 34481
Page 1 of 6



QA RECORD

Instrument Information:

Description: RTD
Manufacturer: BURNS
Model: 3925
Serial Number:

Calibration Information:

Cal Date: 12/16/2010
Due Date: 12/16/2011
Interval: 12 Months
Cal Instruction: 307.04-004
As Found: In Tolerance
As Left: In Tolerance - Adjusted

Ambient Temperature: 72°F +/- 0°F

Ambient Humidity: <=80% RH

This is to certify that all instrumentation, testing methods and personnel used comply with the requirements of the Central Laboratories Services (CLS) Quality Assurance Program which is designed to meet the requirements of ISO/IEC 17025, 10CFR50 Appendix B, ANSI N45.2-1971, and ANSI/NCSL Z540-1-1994. Standards used are traceable to the National Institute of Standards and Technology (NIST), officially recognized agencies, commercially accepted practices or natural physical constants. This report shall not be reproduced, except in full, without the written approval of CLS.

Technical Remarks:

Recalculated coefficients to improve As Left data.

Standards Utilized

| TVA I.D. | Mfg. | Model No. | Description | Cal. Date | Due Date |
|----------|-----------------|-------------------|------------------------------|------------|------------|
| 906643 | ISOTECH | MERCURY CELL | FIXED POINT CELL | 12/08/2009 | 12/08/2014 |
| 906644 | HART SCIENTIFIC | WATER TRIPLE CELL | TRIPLE POINT BATH & CELL | 12/08/2009 | 12/08/2014 |
| 906645 | ISOTECH | GALLIUM CELL | FIXED POINT CELL | 12/08/2009 | 12/08/2014 |
| 906646 | ISOTECH | TIN CELL | FIXED POINT CELL | 12/08/2009 | 12/08/2014 |
| 906647 | ISOTECH | ZINC CELL | FIXED POINT CELL | 12/08/2009 | 12/08/2014 |
| 906722 | GUILDLINE | 6622T | TEMPERATURE MEASURING SYSTEM | 08/23/2010 | 08/23/2011 |
| 906737 | GUILDLINE | 9334A | STANDARD RESISTOR | 08/20/2010 | 08/20/2011 |

| | | |
|---|---|--------------------|
| Calibrated by: David R. Bird Sr Metrology Tech | Approved By: Sam Bertram Calibration Supv. | 12/20/2010 Date |
|---|---|--------------------|

This report was electronically approved using Edison Mulcats Metrology Suite Ver. 2.2.1.

| | |
|---|---------------------------------|
| CENTRAL LABORATORIES SERVICES CHATTANOOGA, TENNESSEE CALIBRATION REPORT | Cust. I.D. No.: 906535 |
| | Page No.: 2 of 6 |
| | Date of Report: 12/16/10 |

Remarks: Accuracy = \pm 0.02 deg C
 Recalculated coefficients prior to As Left test to improve accuracy.
 For As Left data and coefficients refer to page 3 of 6.

AS FOUND TEST

| UUT (deg C) | STD (deg C) | Error (deg C) |
|-------------|-------------|---------------|
| -38.84 | -38.834 | -0.0051 |
| 0.01 | 0.010 | 0.0000 |
| 29.75 | 29.765 | -0.0121 |
| 231.91 | 231.928 | -0.0199 |
| 419.54 | 419.527 | 0.0159 |

As Found ITS 90 Coefficients

Rtpw 100.000221

a5 -4.15854650E-04

b5 -1.55621388E-04

a8 -2.72593907E-04

b8 -2.28004426E-04

Test current 1mA

All meas. ratios between the stds referenced in this instruction and the M&TE calibrated are greater than or equal to 4:1 except as noted.
 This instrument was tested and calibrated to prescribed test procedures and the condition of the instrument is indicated.

Report for ITS-90 Coefficients

Model: 3925

Serial: TVA 906535

Date: December 17, 2010

TPW:

| Reference (°C) | UUT (Ohms) | Residual (°C) |
|----------------|------------|---------------|
| 0.0100 | 100.0002 | N/A |

Low Range:

| Reference (°C) | UUT (Ohms) | Residual (°C) |
|----------------|------------|---------------|
| -38.8344 | 84.4207 | 0.0002 |
| 29.7646 | 111.8088 | 0.0000 |

High Range:

| Reference (°C) | UUT (Ohms) | Residual (°C) |
|----------------|------------|---------------|
| 231.9280 | 189.2353 | 0.0001 |
| 419.5270 | 256.8059 | 0.0000 |

Coefficients:

RTPW = 100.000221

Low Range:

a5 = -4.33797355 E-04

b5 = -1.87516921 E-04

High Range:

a8 = -4.39650984 E-04

b8 = -7.09976322 E-05

Model: 3925 Serial: TVA 906535

ITS-90 Temperature vs. Resistance Table

| °C | Resistance | dr/dT | °C | Resistance | dr/dT | °C | Resistance | dr/dT |
|--------|------------|-----------|-------|------------|-----------|--------|------------|-----------|
| -39.00 | 84.353834 | 0.4034888 | 20.00 | 107.94545 | 0.3961843 | 79.00 | 131.11375 | 0.3890764 |
| -38.00 | 84.757322 | 0.4033577 | 21.00 | 108.34164 | 0.3960630 | 80.00 | 131.50283 | 0.3889566 |
| -37.00 | 85.160680 | 0.4032272 | 22.00 | 108.73770 | 0.3959417 | 81.00 | 131.89179 | 0.3888369 |
| -36.00 | 85.563907 | 0.4030971 | 23.00 | 109.13364 | 0.3958205 | 82.00 | 132.28062 | 0.3887171 |
| -35.00 | 85.967004 | 0.4029674 | 24.00 | 109.52946 | 0.3956993 | 83.00 | 132.66934 | 0.3885974 |
| -34.00 | 86.369972 | 0.4028383 | 25.00 | 109.92516 | 0.3955781 | 84.00 | 133.05794 | 0.3884778 |
| -33.00 | 86.772810 | 0.4027095 | 26.00 | 110.32074 | 0.3954570 | 85.00 | 133.44641 | 0.3883581 |
| -32.00 | 87.175520 | 0.4025812 | 27.00 | 110.71619 | 0.3953359 | 86.00 | 133.83477 | 0.3882385 |
| -31.00 | 87.578101 | 0.4024533 | 28.00 | 111.11153 | 0.3952148 | 87.00 | 134.22301 | 0.3881189 |
| -30.00 | 87.980554 | 0.4023258 | 29.00 | 111.50675 | 0.3950938 | 88.00 | 134.61113 | 0.3879993 |
| -29.00 | 88.382880 | 0.4021987 | 30.00 | 111.90184 | 0.3949728 | 89.00 | 134.99913 | 0.3878797 |
| -28.00 | 88.785079 | 0.4020720 | 31.00 | 112.29681 | 0.3948518 | 90.00 | 135.38701 | 0.3877601 |
| -27.00 | 89.187151 | 0.4019457 | 32.00 | 112.69166 | 0.3947309 | 91.00 | 135.77477 | 0.3876406 |
| -26.00 | 89.589096 | 0.4018198 | 33.00 | 113.08640 | 0.3946100 | 92.00 | 136.16241 | 0.3875211 |
| -25.00 | 89.990916 | 0.4016942 | 34.00 | 113.48101 | 0.3944891 | 93.00 | 136.54993 | 0.3874016 |
| -24.00 | 90.392610 | 0.4015689 | 35.00 | 113.87549 | 0.3943683 | 94.00 | 136.93733 | 0.3872821 |
| -23.00 | 90.794179 | 0.4014440 | 36.00 | 114.26986 | 0.3942475 | 95.00 | 137.32461 | 0.3871627 |
| -22.00 | 91.195623 | 0.4013193 | 37.00 | 114.66411 | 0.3941267 | 96.00 | 137.71178 | 0.3870432 |
| -21.00 | 91.596943 | 0.4011950 | 38.00 | 115.05824 | 0.3940059 | 97.00 | 138.09882 | 0.3869238 |
| -20.00 | 91.998138 | 0.4010709 | 39.00 | 115.45224 | 0.3938852 | 98.00 | 138.48574 | 0.3868044 |
| -19.00 | 92.399208 | 0.4009472 | 40.00 | 115.84613 | 0.3937645 | 99.00 | 138.87255 | 0.3866850 |
| -18.00 | 92.800156 | 0.4008236 | 41.00 | 116.23989 | 0.3936438 | 100.00 | 139.25923 | 0.3865657 |
| -17.00 | 93.200979 | 0.4007003 | 42.00 | 116.63354 | 0.3935232 | 101.00 | 139.64580 | 0.3864464 |
| -16.00 | 93.601680 | 0.4005773 | 43.00 | 117.02706 | 0.3934025 | 102.00 | 140.03225 | 0.3863271 |
| -15.00 | 94.002257 | 0.4004544 | 44.00 | 117.42046 | 0.3932819 | 103.00 | 140.41857 | 0.3862078 |
| -14.00 | 94.402711 | 0.4003317 | 45.00 | 117.81374 | 0.3931614 | 104.00 | 140.80478 | 0.3860885 |
| -13.00 | 94.803043 | 0.4002091 | 46.00 | 118.20690 | 0.3930408 | 105.00 | 141.19087 | 0.3859692 |
| -12.00 | 95.203252 | 0.4000867 | 47.00 | 118.59995 | 0.3929203 | 106.00 | 141.57684 | 0.3858500 |
| -11.00 | 95.603339 | 0.3999645 | 48.00 | 118.99287 | 0.3927998 | 107.00 | 141.96269 | 0.3857308 |
| -10.00 | 96.003303 | 0.3998423 | 49.00 | 119.38567 | 0.3926794 | 108.00 | 142.34842 | 0.3856116 |
| -9.00 | 96.403146 | 0.3997202 | 50.00 | 119.77835 | 0.3925589 | 109.00 | 142.73403 | 0.3854924 |
| -8.00 | 96.802866 | 0.3995981 | 51.00 | 120.17090 | 0.3924385 | 110.00 | 143.11952 | 0.3853733 |
| -7.00 | 97.202464 | 0.3994761 | 52.00 | 120.56334 | 0.3923181 | 111.00 | 143.50490 | 0.3852542 |
| -6.00 | 97.601940 | 0.3993541 | 53.00 | 120.95566 | 0.3921978 | 112.00 | 143.89015 | 0.3851351 |
| -5.00 | 98.001294 | 0.3992321 | 54.00 | 121.34786 | 0.3920774 | 113.00 | 144.27529 | 0.3850160 |
| -4.00 | 98.400526 | 0.3991100 | 55.00 | 121.73994 | 0.3919571 | 114.00 | 144.66030 | 0.3848969 |
| -3.00 | 98.799636 | 0.3989878 | 56.00 | 122.13189 | 0.3918368 | 115.00 | 145.04520 | 0.3847779 |
| -2.00 | 99.198624 | 0.3988656 | 57.00 | 122.52373 | 0.3917166 | 116.00 | 145.42998 | 0.3846588 |
| -1.00 | 99.597490 | 0.3987432 | 58.00 | 122.91545 | 0.3915963 | 117.00 | 145.81464 | 0.3845398 |
| 0.00 | 99.996233 | 0.3986186 | 59.00 | 123.30704 | 0.3914761 | 118.00 | 146.19918 | 0.3844209 |
| 1.00 | 100.39485 | 0.3984965 | 60.00 | 123.69852 | 0.3913559 | 119.00 | 146.58360 | 0.3843019 |
| 2.00 | 100.79335 | 0.3983744 | 61.00 | 124.08987 | 0.3912357 | 120.00 | 146.96790 | 0.3841829 |
| 3.00 | 101.19172 | 0.3982523 | 62.00 | 124.48111 | 0.3911156 | 121.00 | 147.35208 | 0.3840640 |
| 4.00 | 101.58998 | 0.3981303 | 63.00 | 124.87223 | 0.3909954 | 122.00 | 147.73615 | 0.3839451 |
| 5.00 | 101.98811 | 0.3980084 | 64.00 | 125.26322 | 0.3908753 | 123.00 | 148.12009 | 0.3838262 |
| 6.00 | 102.38611 | 0.3978865 | 65.00 | 125.65410 | 0.3907552 | 124.00 | 148.50392 | 0.3837074 |
| 7.00 | 102.78400 | 0.3977646 | 66.00 | 126.04485 | 0.3906352 | 125.00 | 148.88762 | 0.3835885 |
| 8.00 | 103.18177 | 0.3976428 | 67.00 | 126.43549 | 0.3905151 | 126.00 | 149.27121 | 0.3834697 |
| 9.00 | 103.57941 | 0.3975211 | 68.00 | 126.82600 | 0.3903951 | 127.00 | 149.65468 | 0.3833509 |
| 10.00 | 103.97693 | 0.3973993 | 69.00 | 127.21640 | 0.3902751 | 128.00 | 150.03803 | 0.3832321 |
| 11.00 | 104.37433 | 0.3972776 | 70.00 | 127.60667 | 0.3901552 | 129.00 | 150.42127 | 0.3831134 |
| 12.00 | 104.77161 | 0.3971560 | 71.00 | 127.99683 | 0.3900352 | 130.00 | 150.80438 | 0.3829946 |
| 13.00 | 105.16876 | 0.3970344 | 72.00 | 128.38686 | 0.3899153 | 131.00 | 151.18737 | 0.3828759 |
| 14.00 | 105.56580 | 0.3969128 | 73.00 | 128.77678 | 0.3897954 | 132.00 | 151.57025 | 0.3827572 |
| 15.00 | 105.96271 | 0.3967913 | 74.00 | 129.16657 | 0.3896755 | 133.00 | 151.95301 | 0.3826386 |
| 16.00 | 106.35950 | 0.3966698 | 75.00 | 129.55625 | 0.3895556 | 134.00 | 152.33565 | 0.3825199 |
| 17.00 | 106.75617 | 0.3965484 | 76.00 | 129.94580 | 0.3894358 | 135.00 | 152.71816 | 0.3824013 |
| 18.00 | 107.15272 | 0.3964270 | 77.00 | 130.33524 | 0.3893159 | 136.00 | 153.10057 | 0.3822827 |
| 19.00 | 107.54915 | 0.3963056 | 78.00 | 130.72456 | 0.3891961 | 137.00 | 153.48285 | 0.3821641 |

Model: 3925 Serial: TVA 906535

ITS-90 Temperature vs. Resistance Table

| °C | Resistance | dr/dT | °C | Resistance | dr/dT | °C | Resistance | dr/dT |
|--------|------------|-----------|--------|------------|-----------|--------|------------|-----------|
| 138.00 | 153.86501 | 0.3820455 | 197.00 | 176.20357 | 0.3750883 | 256.00 | 198.13379 | 0.3682033 |
| 139.00 | 154.24706 | 0.3819270 | 198.00 | 176.57865 | 0.3749710 | 257.00 | 198.50199 | 0.3680872 |
| 140.00 | 154.62899 | 0.3818084 | 199.00 | 176.95363 | 0.3748537 | 258.00 | 198.87008 | 0.3679710 |
| 141.00 | 155.01079 | 0.3816899 | 200.00 | 177.32848 | 0.3747365 | 259.00 | 199.23805 | 0.3678549 |
| 142.00 | 155.39248 | 0.3815715 | 201.00 | 177.70322 | 0.3746193 | 260.00 | 199.60590 | 0.3677388 |
| 143.00 | 155.77406 | 0.3814530 | 202.00 | 178.07784 | 0.3745021 | 261.00 | 199.97364 | 0.3676227 |
| 144.00 | 156.15551 | 0.3813346 | 203.00 | 178.45234 | 0.3743850 | 262.00 | 200.34126 | 0.3675066 |
| 145.00 | 156.53684 | 0.3812161 | 204.00 | 178.82672 | 0.3742678 | 263.00 | 200.70877 | 0.3673905 |
| 146.00 | 156.91806 | 0.3810977 | 205.00 | 179.20099 | 0.3741507 | 264.00 | 201.07616 | 0.3672745 |
| 147.00 | 157.29916 | 0.3809794 | 206.00 | 179.57514 | 0.3740336 | 265.00 | 201.44344 | 0.3671584 |
| 148.00 | 157.68014 | 0.3808610 | 207.00 | 179.94917 | 0.3739165 | 266.00 | 201.81059 | 0.3670424 |
| 149.00 | 158.06100 | 0.3807427 | 208.00 | 180.32309 | 0.3737995 | 267.00 | 202.17764 | 0.3669264 |
| 150.00 | 158.44174 | 0.3806244 | 209.00 | 180.69689 | 0.3736824 | 268.00 | 202.54465 | 0.3668104 |
| 151.00 | 158.82236 | 0.3805061 | 210.00 | 181.07057 | 0.3735654 | 269.00 | 202.91137 | 0.3666944 |
| 152.00 | 159.20287 | 0.3803878 | 211.00 | 181.44414 | 0.3734484 | 270.00 | 203.27807 | 0.3665784 |
| 153.00 | 159.58326 | 0.3802696 | 212.00 | 181.81759 | 0.3733314 | 271.00 | 203.64465 | 0.3664624 |
| 154.00 | 159.96353 | 0.3801513 | 213.00 | 182.19092 | 0.3732145 | 272.00 | 204.01111 | 0.3663465 |
| 155.00 | 160.34368 | 0.3800331 | 214.00 | 182.56413 | 0.3730975 | 273.00 | 204.37745 | 0.3662305 |
| 156.00 | 160.72371 | 0.3799150 | 215.00 | 182.93723 | 0.3729806 | 274.00 | 204.74369 | 0.3661146 |
| 157.00 | 161.10363 | 0.3797968 | 216.00 | 183.31021 | 0.3728637 | 275.00 | 205.10980 | 0.3659987 |
| 158.00 | 161.48342 | 0.3796787 | 217.00 | 183.68307 | 0.3727469 | 276.00 | 205.47580 | 0.3658828 |
| 159.00 | 161.86310 | 0.3795605 | 218.00 | 184.05582 | 0.3726300 | 277.00 | 205.84168 | 0.3657669 |
| 160.00 | 162.24266 | 0.3794424 | 219.00 | 184.42845 | 0.3725132 | 278.00 | 206.20745 | 0.3656510 |
| 161.00 | 162.62211 | 0.3793244 | 220.00 | 184.80096 | 0.3723963 | 279.00 | 206.57310 | 0.3655351 |
| 162.00 | 163.00143 | 0.3792063 | 221.00 | 185.17336 | 0.3722796 | 280.00 | 206.93863 | 0.3654193 |
| 163.00 | 163.38064 | 0.3790883 | 222.00 | 185.54564 | 0.3721628 | 281.00 | 207.30405 | 0.3653034 |
| 164.00 | 163.75972 | 0.3789703 | 223.00 | 185.91780 | 0.3720460 | 282.00 | 207.66936 | 0.3651876 |
| 165.00 | 164.13869 | 0.3788523 | 224.00 | 186.28985 | 0.3719293 | 283.00 | 208.03454 | 0.3650718 |
| 166.00 | 164.51755 | 0.3787343 | 225.00 | 186.66178 | 0.3718126 | 284.00 | 208.39962 | 0.3649559 |
| 167.00 | 164.89628 | 0.3786164 | 226.00 | 187.03359 | 0.3716959 | 285.00 | 208.76457 | 0.3648401 |
| 168.00 | 165.27490 | 0.3784985 | 227.00 | 187.40529 | 0.3715792 | 286.00 | 209.12941 | 0.3647243 |
| 169.00 | 165.65340 | 0.3783806 | 228.00 | 187.77687 | 0.3714625 | 287.00 | 209.49414 | 0.3646085 |
| 170.00 | 166.03178 | 0.3782627 | 229.00 | 188.14833 | 0.3713459 | 288.00 | 209.85875 | 0.3644928 |
| 171.00 | 166.41004 | 0.3781448 | 230.00 | 188.51967 | 0.3712293 | 289.00 | 210.22324 | 0.3643770 |
| 172.00 | 166.78818 | 0.3780270 | 231.00 | 188.89090 | 0.3711127 | 290.00 | 210.58762 | 0.3642612 |
| 173.00 | 167.16621 | 0.3779092 | 232.00 | 189.26202 | 0.3709961 | 291.00 | 210.95188 | 0.3641455 |
| 174.00 | 167.54412 | 0.3777914 | 233.00 | 189.63301 | 0.3708795 | 292.00 | 211.31602 | 0.3640297 |
| 175.00 | 167.92191 | 0.3776736 | 234.00 | 190.00389 | 0.3707630 | 293.00 | 211.68005 | 0.3639140 |
| 176.00 | 168.29959 | 0.3775559 | 235.00 | 190.37465 | 0.3706464 | 294.00 | 212.04397 | 0.3637982 |
| 177.00 | 168.67714 | 0.3774382 | 236.00 | 190.74530 | 0.3705299 | 295.00 | 212.40776 | 0.3636825 |
| 178.00 | 169.05458 | 0.3773205 | 237.00 | 191.11583 | 0.3704134 | 296.00 | 212.77145 | 0.3635668 |
| 179.00 | 169.43190 | 0.3772028 | 238.00 | 191.48624 | 0.3702970 | 297.00 | 213.13501 | 0.3634511 |
| 180.00 | 169.80910 | 0.3770851 | 239.00 | 191.85654 | 0.3701805 | 298.00 | 213.49846 | 0.3633354 |
| 181.00 | 170.18619 | 0.3769675 | 240.00 | 192.22672 | 0.3700641 | 299.00 | 213.86180 | 0.3632197 |
| 182.00 | 170.56316 | 0.3768499 | 241.00 | 192.59679 | 0.3699477 | 300.00 | 214.22502 | 0.3631040 |
| 183.00 | 170.94001 | 0.3767323 | 242.00 | 192.96673 | 0.3698312 | 301.00 | 214.58812 | 0.3629883 |
| 184.00 | 171.31674 | 0.3766147 | 243.00 | 193.33657 | 0.3697149 | 302.00 | 214.95111 | 0.3628727 |
| 185.00 | 171.69335 | 0.3764972 | 244.00 | 193.70628 | 0.3695985 | 303.00 | 215.31398 | 0.3627570 |
| 186.00 | 172.06985 | 0.3763796 | 245.00 | 194.07588 | 0.3694821 | 304.00 | 215.67674 | 0.3626413 |
| 187.00 | 172.44623 | 0.3762621 | 246.00 | 194.44536 | 0.3693658 | 305.00 | 216.03938 | 0.3625257 |
| 188.00 | 172.82249 | 0.3761446 | 247.00 | 194.81473 | 0.3692495 | 306.00 | 216.40191 | 0.3624100 |
| 189.00 | 173.19864 | 0.3760272 | 248.00 | 195.18398 | 0.3691332 | 307.00 | 216.76432 | 0.3622944 |
| 190.00 | 173.57466 | 0.3759097 | 249.00 | 195.55311 | 0.3690169 | 308.00 | 217.12661 | 0.3621787 |
| 191.00 | 173.95057 | 0.3757923 | 250.00 | 195.92213 | 0.3689006 | 309.00 | 217.48879 | 0.3620631 |
| 192.00 | 174.32637 | 0.3756749 | 251.00 | 196.29103 | 0.3687844 | 310.00 | 217.85085 | 0.3619474 |
| 193.00 | 174.70204 | 0.3755575 | 252.00 | 196.65981 | 0.3686681 | 311.00 | 218.21280 | 0.3618318 |
| 194.00 | 175.07760 | 0.3754402 | 253.00 | 197.02848 | 0.3685519 | 312.00 | 218.57463 | 0.3617162 |
| 195.00 | 175.45304 | 0.3753229 | 254.00 | 197.39703 | 0.3684357 | 313.00 | 218.93635 | 0.3616005 |
| 196.00 | 175.82836 | 0.3752055 | 255.00 | 197.76547 | 0.3683195 | 314.00 | 219.29795 | 0.3614849 |

Model: 3925 Serial: TVA 906535

ITS-90 Temperature vs. Resistance Table

| °C | Resistance | dr/dT | °C | Resistance | dr/dT |
|--------|------------|-----------|--------|------------|-----------|
| 315.00 | 219.65944 | 0.3613693 | 374.00 | 240.78236 | 0.3545432 |
| 316.00 | 220.02080 | 0.3612537 | 375.00 | 241.13690 | 0.3544273 |
| 317.00 | 220.38206 | 0.3611381 | 376.00 | 241.49133 | 0.3543114 |
| 318.00 | 220.74320 | 0.3610224 | 377.00 | 241.84564 | 0.3541954 |
| 319.00 | 221.10422 | 0.3609068 | 378.00 | 242.19983 | 0.3540794 |
| 320.00 | 221.46513 | 0.3607912 | 379.00 | 242.55391 | 0.3539634 |
| 321.00 | 221.82592 | 0.3606756 | 380.00 | 242.90788 | 0.3538474 |
| 322.00 | 222.18659 | 0.3605600 | 381.00 | 243.26173 | 0.3537314 |
| 323.00 | 222.54715 | 0.3604444 | 382.00 | 243.61546 | 0.3536154 |
| 324.00 | 222.90760 | 0.3603288 | 383.00 | 243.96907 | 0.3534993 |
| 325.00 | 223.26793 | 0.3602132 | 384.00 | 244.32257 | 0.3533833 |
| 326.00 | 223.62814 | 0.3600976 | 385.00 | 244.67595 | 0.3532672 |
| 327.00 | 223.98824 | 0.3599820 | 386.00 | 245.02922 | 0.3531511 |
| 328.00 | 224.34822 | 0.3598664 | 387.00 | 245.38237 | 0.3530350 |
| 329.00 | 224.70808 | 0.3597508 | 388.00 | 245.73541 | 0.3529189 |
| 330.00 | 225.06784 | 0.3596351 | 389.00 | 246.08833 | 0.3528027 |
| 331.00 | 225.42747 | 0.3595195 | 390.00 | 246.44113 | 0.3526865 |
| 332.00 | 225.78699 | 0.3594039 | 391.00 | 246.79382 | 0.3525704 |
| 333.00 | 226.14639 | 0.3592883 | 392.00 | 247.14639 | 0.3524542 |
| 334.00 | 226.50568 | 0.3591727 | 393.00 | 247.49884 | 0.3523379 |
| 335.00 | 226.86486 | 0.3590571 | 394.00 | 247.85118 | 0.3522217 |
| 336.00 | 227.22391 | 0.3589415 | 395.00 | 248.20340 | 0.3521055 |
| 337.00 | 227.58285 | 0.3588258 | 396.00 | 248.55551 | 0.3519892 |
| 338.00 | 227.94168 | 0.3587102 | 397.00 | 248.90749 | 0.3518729 |
| 339.00 | 228.30039 | 0.3585946 | 398.00 | 249.25937 | 0.3517566 |
| 340.00 | 228.65898 | 0.3584789 | 399.00 | 249.61112 | 0.3516403 |
| 341.00 | 229.01746 | 0.3583633 | 400.00 | 249.96276 | 0.3515239 |
| 342.00 | 229.37583 | 0.3582477 | 401.00 | 250.31429 | 0.3514075 |
| 343.00 | 229.73407 | 0.3581320 | 402.00 | 250.66570 | 0.3512912 |
| 344.00 | 230.09221 | 0.3580164 | 403.00 | 251.01699 | 0.3511748 |
| 345.00 | 230.45022 | 0.3579007 | 404.00 | 251.36816 | 0.3510583 |
| 346.00 | 230.80812 | 0.3577850 | 405.00 | 251.71922 | 0.3509419 |
| 347.00 | 231.16591 | 0.3576694 | 406.00 | 252.07016 | 0.3508254 |
| 348.00 | 231.52358 | 0.3575537 | 407.00 | 252.42099 | 0.3507089 |
| 349.00 | 231.88113 | 0.3574380 | 408.00 | 252.77170 | 0.3505924 |
| 350.00 | 232.23857 | 0.3573223 | 409.00 | 253.12229 | 0.3504759 |
| 351.00 | 232.59589 | 0.3572066 | 410.00 | 253.47276 | 0.3503593 |
| 352.00 | 232.95310 | 0.3570909 | 411.00 | 253.82312 | 0.3502427 |
| 353.00 | 233.31019 | 0.3569752 | 412.00 | 254.17337 | 0.3501261 |
| 354.00 | 233.66716 | 0.3568595 | 413.00 | 254.52349 | 0.3500095 |
| 355.00 | 234.02402 | 0.3567438 | 414.00 | 254.87350 | 0.3498928 |
| 356.00 | 234.38077 | 0.3566281 | 415.00 | 255.22340 | 0.3497762 |
| 357.00 | 234.73740 | 0.3565123 | 416.00 | 255.57317 | 0.3496595 |
| 358.00 | 235.09391 | 0.3563966 | 417.00 | 255.92283 | 0.3495428 |
| 359.00 | 235.45030 | 0.3562808 | 418.00 | 256.27237 | 0.3494260 |
| 360.00 | 235.80659 | 0.3561651 | 419.00 | 256.62180 | 0.3493092 |
| 361.00 | 236.16275 | 0.3560493 | 420.00 | 256.97111 | 0.3491925 |
| 362.00 | 236.51880 | 0.3559335 | | | |
| 363.00 | 236.87473 | 0.3558177 | | | |
| 364.00 | 237.23055 | 0.3557019 | | | |
| 365.00 | 237.58625 | 0.3555861 | | | |
| 366.00 | 237.94184 | 0.3554703 | | | |
| 367.00 | 238.29731 | 0.3553544 | | | |
| 368.00 | 238.65266 | 0.3552386 | | | |
| 369.00 | 239.00790 | 0.3551227 | | | |
| 370.00 | 239.36303 | 0.3550069 | | | |
| 371.00 | 239.71803 | 0.3548910 | | | |
| 372.00 | 240.07292 | 0.3547751 | | | |
| 373.00 | 240.42770 | 0.3546592 | | | |

Calibration Certificate for Azonix Ohmmeter

Instrument used to read resistance of Burns RTD thermometers.

Azonix Ohmmeter ID No. 906527 was used for both pre-survey and post-survey calibrations.



LAB STANDARD REPORT of CALIBRATION

Tennessee Valley Authority
Central Laboratories Services

Mailing Address: 1101 Market Street, PSC-1B-C, Chattanooga, TN 37402
Shipping Address: 4601 North Access Road, Bldg. A, Chattanooga, TN 37415
Phone: (423) 876-4318 Fax: (423) 876-4137

Customer:

CLS KNOXVILLE
400 W. SUMMIT HILL DR.
KNOXVILLE, TN 37902

Asset ID: 906527
Certificate No: 34480
Page 1 of 2



QA RECORD

Instrument Information:

Description: DIGITAL THERMOMETER
Manufacturer: AZONIX
Model: A1011-RS-AD-RT41
Serial Number:

Calibration Information:

Cal Date: 01/07/2011
Due Date: 01/07/2012
Interval: 12 Months
Cal Instruction: 308.02-003
As Found: In Tolerance
As Left: In Tolerance

Ambient Temperature: 72°F +/- 2°F

Ambient Humidity: <=50% RH

This is to certify that all instrumentation, testing methods and personnel used comply with the requirements of the Central Laboratories Services (CLS) Quality Assurance Program which is designed to meet the requirements of ISO/IEC 17025, 10CFR50 Appendix B, ANSI N45.2-1971, and ANSI/NCSL Z540-1-1994. Standards used are traceable to the National Institute of Standards and Technology (NIST), officially recognized agencies, commercially accepted practices or natural physical constants. This report shall not be reproduced, except in full, without the written approval of CLS.

Technical Remarks:

Left as found. Certification is limited to channels 1 and 2. Channels 3 and 4 are not certified. Limited certification label is attached.

Standards Utilized

| TVA I.D. | Mfg. | Model No. | Description | Cal. Date | Due Date |
|----------|-----------|-----------|----------------------------|------------|------------|
| 259303 | HONEYWELL | 1190 | RESISTANCE STANDARD, 1 OHM | 07/21/2010 | 07/21/2015 |
| 906523 | OMEGA | HH 42 | DIGITAL THERMOMETER | 12/17/2010 | 12/17/2011 |
| E29099 | GUILDLINE | 8875A | DC RESISTANCE BRIDGE | 11/17/2010 | 02/10/2011 |

| | | |
|---|---|--------------------|
| Calibrated by: Keith Roberts Sr Metrology Tech | Approved By: Sam Bertram Calibration Supv. | 01/11/2011 Date |
|---|---|--------------------|

This report was electronically approved using Edison Mulcats Metrology Suite Ver. 2.2.1.

CENTRAL LABORATORIES SERVICES

CHATTANOOGA, TENNESSEE

Cust. I. D. No.: **906527**Page No.: **2 of 2****CALIBRATION REPORT**Date of Report: **1/7/11****Remarks:** Accuracy = **0.004** Ohms

Certification is limited to channels 1 and 2; channels 3 and 4 are not certified.
 Limited certification label is attached.



Left as found.

*Denotes out of tolerance.

AS FOUND

| Probe | Standard Resistance (Ohms) | UUT Reading (Ohms) | Error (Ohms) |
|-------|----------------------------|--------------------|--------------|
| 1 | 89.9995 | 90.001 | 0.002 |
| | 99.9995 | 100.002 | 0.003 |
| | 119.9993 | 120.002 | 0.003 |
| 2 | 89.9995 | 90.002 | 0.003 |
| | 99.9995 | 100.002 | 0.003 |
| | 119.9993 | 120.002 | 0.003 |

TVA Procedure for Calibration of HOBO Water Temperature Probes

| | | |
|---|--|---|
|  CENTRAL LABORATORIES SERVICES QUALITY PROGRAM INSTRUCTION | TITLE Certification of HOBO Water Temp Pro Data Acquisition SystemsH ₂ O-001 | Instruction No. 450.01-020 Rev. No. 0 Page No. 1 of 7 Effective Date 5/19/03 |
| | | |
| LEVEL OF USE <input type="checkbox"/> Continuous <input checked="" type="checkbox"/> Reference <input type="checkbox"/> Information | | |
| <div style="text-align: right; font-weight: bold; font-size: 1.2em;">QA RECORD</div> <div style="display: flex; justify-content: space-between; margin-top: 20px;"> <div style="width: 45%;"> <p>_____</p> <p style="text-align: center;">Dennis T. Darby Preparer</p> </div> <div style="width: 45%;"> <p>_____</p> <p style="text-align: center;">5/19/03 Date</p> </div> </div> <div style="display: flex; justify-content: space-between; margin-top: 40px;"> <div style="width: 45%;"> <p>_____</p> <p style="text-align: center;">Paul B. Loiseau, Jr. Technical Reviewer</p> </div> <div style="width: 45%;"> <p>_____</p> <p style="text-align: center;">5/19/03 Date</p> </div> </div> <div style="display: flex; justify-content: space-between; margin-top: 40px;"> <div style="width: 45%;"> <p>_____</p> <p style="text-align: center;"> Administrative Review</p> </div> <div style="width: 45%;"> <p>_____</p> <p style="text-align: center;">6/5/03 Date</p> </div> </div> | | |
| APPROVAL | | |
| <div style="display: flex; justify-content: space-between; margin-top: 20px;"> <div style="width: 45%;"> <p>_____</p> <p style="text-align: center;">Jerry D. Hubble Department Manager</p> </div> <div style="width: 45%;"> <p>_____</p> <p style="text-align: center;">5/19/03 Date</p> </div> </div> | | |

| | |
|--|-----------------------------------|
| TITLE: Certification of HOBO Water Temp Pro Data Acquisition Systems H₂O-001 | Instruction No. 450.01-020 |
| | Rev. 0 |
| | Eff. Date 5/19/03 |
| | Page 3 of 7 |

1.0 PURPOSE

To provide uniform and effective certifications of Hobo Water Temp Pro data acquisition systems meeting the accuracy and performance requirements of TVA's water temperature-monitoring programs. This technical instruction uses the method of comparison with a laboratory standard thermometer.

2.0 SCOPE

This instruction applies to the certification of Hobo Water Temp Pro data loggers manufactured by Onset Computer Corporation of Bourne, Massachusetts. The Hobo Water Temp Pro is a data acquisition system containing a temperature sensor, data logger and battery sealed in a single submersible case. The Hobo Water Temp Pro is programmed and data retrieved by use of an infrared interface located in one end of the case. Hobo Water Temp Pros are certified upon receipt from the manufacturer at no greater than 12 month intervals during use or when requested.

3.0 SUMMARY

In this three-point certification systems are tested as actually used over the historical water temperature range of 30° to 100°F and submerged in water. The three test points are 37°, 65° and 93°F. The systems are required to perform within Onset Computer Corporation tolerances. System conformity at each temperature point is determined by comparing system temperature, logged by the Hobo Water Temp Pro and a laboratory standard thermometer.

Systems are programmed and submerged with a standard thermometer in a stirred, temperature-controlled temperature bath. The systems are read after the test by an infrared interface adapter connected to a computer running Onset Computer Corporation's Boxcar Pro software. Traceability of the certification is through the thermometer.

"As-found" certifications are performed on new systems as an acceptance test and on sensors returned from field service. "As-left" certifications are performed before delivery for field service if more than 12 months has elapsed since the last certification. "As-found" and "as-left" certifications may be combined on the same record if there is clear indication which type each system is undergoing.

Multiple HOBOs may be certified at the same time in the temperature bath.

| | |
|--|-----------------------------------|
| TITLE: Certification of HOBO Water Temp Pro Data Acquisition Systems H₂O-001 | Instruction No. 450.01-020 |
| | Rev. 0 |
| | Eff. Date 5/19/03 |
| | Page 4 of 7 |

- Accuracy of $\pm 0.2^{\circ}\text{C}$ at 25°C (0.33°F at 70°F)
- Waterproof case, submersible to 100 feet
- Capacity to store up to 21,580 temperature measurements
- Selectable sampling interval from 1 second to 9 hours
- Programmable start time/date
- Two data recording modes: Stop when full or wrap around when full.
- Two data offload modes: Halt then offload or offload while logging.
- Nonvolatile EEPROM memory that retains data even if batteries fail
- Light-emitting diode (LED) operation, indicator, which can be disabled during logging by selecting "Stealth" mode
- High-speed IR communications for offloading data; can readout full logger in less than 30 seconds while logging continues
- Battery life of 6 years with typical usage

4.0 PRACTICES/EXCEPTIONS

N/A

5.0 SAFETY

- 5.1 Standard electrical equipment safety.

6.0 STANDARDS USED

- 6.1 Laboratory reference thermometer, range 30° to 100°F or greater, 0.01°F resolution, 0.1°F accuracy or better, with current calibration sticker.

7.0 EQUIPMENT/APPARATUS

- 7.1 Temperature bath, stirred, temperature-controlled.
7.2 Computer with Onset Boxcar Pro software installed (version 4.3 or later)
7.3 IR Base station, Onset Part # BST -IR

8.0 PREREQUISITE ACTIONS

- 8.1 Turn on temperature bath and set for 37°F .
8.2 Check the IR interface to verify that it is plugged into the correct serial port on the PC. Set the correct time on the PC.
8.3 Align the IR port on the Base station with the HOBO Water Temp Pro communications window. Place the logger no further than 4 to 5 inches away from the Base station (see Figure 2) and make sure the IR windows in both devices point at each other. There is a 30° acceptance angle for the IR beam, so some misalignment is acceptable.

| | |
|--|--|
| TITLE: Certification of HOBO Water Temp Pro Data Acquisition Systems H₂O-001 | Instruction No. 450.01-020 Rev. 0 Eff. Date 5/19/03 Page 5 of 7 |
|--|--|

- 8.4 Start the Onset Box Car Software and select **Logger** then **Hobo Water Temp Pro** and **Launch**.
- 8.5 The computer will respond with a list of **loggers** found. The serial number in this list should match the serial number printed on the side of the logger. If these numbers do not match, click the **Refresh** button. Record this serial number on the certification form. Either wait or click the **Stop Searching** button. Using the mouse select the logger and click the **Launch** button.
- 8.6 After a few seconds the screen will display the status of the HOBO Water Temp Pro. Record the battery percentage on the certification form.
- 8.7 Verify that the Hobo is set to Fahrenheit and program it to a recording interval of 0:1:0 for a reading once a minute. Verify that the start logging immediately box is checked and that the set data logger clock with host launch is also checked.
- 8.8 Using the mouse click the **Launch Immediately** button.
- 8.9 If last HOBO is programmed click the **DONE** button, else select the **Launch Another** and repeat steps 8.5 through 8.9.
- 9.0 **TEST PROCEDURE/METHOD**
- 9.1 On the certification form record the serial number of the laboratory reference thermometer.
- 9.2 Place the HOBO Water Temp Pro in the temperature bath, making sure the end opposite the IR windows is submerged, and allow the bath to stabilize at 37°F ±0.5°F on the thermometer. Adjust the bath set point if needed. After the bath reaches the desired temperature allow 20 minutes 'soak time' for the HOBO to reach its final temperature.
- 9.3 Record the thermometer reading on the certification form and the time. (The time will be needed to get the correct reading from the HOBO.)
- 9.4 Repeat steps 9.2 and 9.3 for bath settings of 65.0°F ± 0.5°F and 93°F ± 0.5°F.
- 9.5 Remove the HOBO from the temperature bath and align the IR port on the Base station with the HOBO Water Temp Pro communications window.
- 9.6 Restart Onset BoxCar Pro if it is not running and select **Logger** then **Hobo Water Temp Pro** and **Readout**.
- 9.7 The computer will respond with a list of **loggers** found. Using the mouse select the logger and click the **Readout** button. The computer will ask to download data and continue logging or the stop logging and offload data. Select the **Stop Logging and Offload data**. After a few seconds the computer will respond with a suggested file name. Select **Save** and allow the HOBO to transfer the data.
- 9.8 After a successful download click the **OK** button. The computer will then ask if the data should be displayed in Centigrade or Fahrenheit. Deselect °C and select °F and click **OK**. The computer should display a graph of the collected data. Click the view details button (this is the button just left of the question mark button.)

| | |
|--|-----------------------------------|
| TITLE: Certification of HOBO Water Temp Pro Data Acquisition Systems H₂O-001 | Instruction No. 450.01-020 |
| | Rev. 0 |
| | Eff. Date 5/19/03 |
| | Page 6 of 7 |

9.9 Scroll down the displayed list until the time recorded for the 37°F point is found. Record the corresponding temperature on the certification form. Repeat this step for 65° and 93°.

9.10 Close the view details windows and repeat steps 9.6 through 9.9 for additional HOBOs.

9.11 Fill out the rest of the certification form.

10.0 ACCEPTANCE CRITERIA

10.1 Based upon the manufacturer specifications the HOBO Water Temp Pro should be within $\pm 0.4^{\circ}\text{F}$ over the range of 32°F to 100°F. Any HOBO with an error of greater than $\pm 0.5^{\circ}\text{F}$ at any of the three measured points shall fail certification.

11.0 POST PROCEDURE ACTIVITY

11.1 Close the BoxCar Software.

12.0 RECORDS

12.1 Completed HOBO Water Temperature Pro Certification form and associated Report of Certification cover sheet is a QA record.

13.0 REFERENCE

13.1 HOBO Water Temp Pro User's Manual, version 1.0 or later

13.2 Onset BoxCar Pro4 Manual Version 1.0 or later

| | |
|--|----------------------------|
| TITLE: Certification of HOBO Water Temp Pro Data Acquisition Systems H20-001 | Instruction No. 450.01-020 |
| | Rev. 0 |
| | Eff. Date 5/19/03 |
| | Page 7 of 7 |

| | |
|--|--------------------|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | SN Page Date |
|--|--------------------|

WATER TEMPERATURE HOBO WATER TEMP PRO CALIBRATION RECORD

Date of Certification: April 25, 2001

Type of Certification: As-found As-Left X

| SENSOR INFO For As-Found List Plant S/N & PLNT | 37 deg F | | 65 degF | | 93 degF | | P A S S | F A I L | Battery L I F E |
|---|-----------------|------------------|-----------------|------------------|-----------------|------------------|------------------|------------------|-----------------------------|
| | BATH TEMP | | BATH TEMP | | BATH TEMP | | | | |
| | Limits | | Limits | | Limits | | | | |
| | 0.40 degF OBSVD | -0.40 degF ERROR | 0.40 degF OBSVD | -0.40 degF ERROR | 0.40 degF OBSVD | -0.40 degF ERROR | | | |
| 1 | | 0.00 | | 0.00 | | 0.00 | ✓ | | |
| 2 | | 0.00 | | 0.00 | | 0.00 | ✓ | | |
| 3 | | 0.00 | | 0.00 | | 0.00 | ✓ | | |
| 4 | | 0.00 | | 0.00 | | 0.00 | ✓ | | |
| 5 | | 0.00 | | 0.00 | | 0.00 | ✓ | | |
| 6 | | 0.00 | | 0.00 | | 0.00 | ✓ | | |
| 7 | | 0.00 | | 0.00 | | 0.00 | ✓ | | |
| 8 | | 0.00 | | 0.00 | | 0.00 | ✓ | | |
| 9 | | 0.00 | | 0.00 | | 0.00 | ✓ | | |
| 10 | | 0.00 | | 0.00 | | 0.00 | ✓ | | |

SENSOR TYPE: HOBO Water Temp Pro H20-001

Remarks: _____

Calibration Certificates for HOBO Water Temperature Probes

Table of HOBO Probes Used for the WBN Survey Summarized Herein

| Station (Figure 3) | Depth (feet) | HOBO Logger (Serial Number) | Station (Figure 3) | Depth (feet) | HOBO Logger (Serial Number) |
|-----------------------|-----------------|--------------------------------|-----------------------|-----------------|--------------------------------|
| WB1 | 0.5 | 1304864 | WB7 | 0.5 | 1305136 |
| | 3 | 1304872 | | 3 | 1305160 |
| | 5 | 1305177 | | 5 | 1304855 |
| | 7 | 1304860 | | 7 | 1304890 |
| WB2 | 0.5 | 1305152 | WB8 | 0.5 | 1305139 |
| | 3 | 1304888 | | 3 | 1304886 |
| | 5 | 1304891 | | 5 | 1305174 |
| | 7 | 1304874 | | 7 | 1305143 |
| WB3 | 0.5 | 1305159 | WB9 | 0.5 | 1304866 |
| | 3 | 1305144 | | 3 | 1305140 |
| | 5 | 1305184 | | 5 | 1305150 |
| | 7 | 1304867 | | 7 | 1304870 |
| WB4 | 0.5 | 1305192 | WB10 | 0.5 | 1304861 |
| | 3 | 1304854 | | 3 | 1305156 |
| | 5 | 1304865 | | 5 | 1304877 |
| | 7 | 1304889 | | 7 | 1305179 |
| WB5 | 0.5 | 1304882 | WB11 | 0.5 | 1134040 |
| | 3 | 1305164 | | 3 | 1305176 |
| | 5 | 1304853 | | 5 | 1304878 |
| | 7 | 1305182 | | 7 | 1305153 |
| WB6 | 0.5 | 1304883 | WB12 | 0.5 | 1305141 |
| | 3 | 1304868 | | 3 | 1304851 |
| | 5 | 1305161 | | 5 | 1304857 |
| | 7 | 1304863 | | 7 | 1305155 |

Pre-Survey Calibrations

| | |
|--|---|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | ID E44909 Page 1 of 2 Date 02/07/2011 |
|--|---|

METEOROLOGICAL MONITORING INSTRUMENTATION REPORT OF CALIBRATION

Calibrated For: Hydrothermal Compliance Date of Report: 02/07/2011

Item Description: HOBO WATER PRO TVA I.D. No.: E44909

Manufacturer: Onset Computer Corporation

Model: U22-001 CLS Instruction No.: 450.01-020

S/N No.: See Attached Sheet

Dispositioned to: CLS Norris Lab As-Left calibration in tolerance

Standards Used Log:

| I.D. No. | Description | Calibration Date | Calibration Due Date |
|----------|-----------------------------------|------------------|----------------------|
| 906527 | Azonix A1011-RS-XX Therm/Ohmmeter | 01/07/2011 | 01/07/2012 |
| 906535 | Burns Engineering 12001 PRT | 12/16/2010 | 12/16/2011 |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

This is to certify that all instrumentation, testing methods and personnel used comply with the requirements of the Central Laboratories Services (CLS) Quality Assurance Program which is designed to meet the requirements of ISO/IEC 17025, 10 CFR 50 Appendix B and ANSI N45.2-1971, and ANSI/NCSL Z540-1-1994. Standards used are traceable to the National Institute of Standards and Technology (NIST), officially recognized agencies, commercially accepted practices or natural physical constants. This report shall not be reproduced except in full, without the written approval of CLS.

Calibrated By: Elaine Houston

Approved By: Randy Cooper

Date Approved: 2/9/11

| | | |
|---|--|---|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | | ID E44909 Page 2 of 2 Date 02/07/2011 |
|---|--|---|

WATER TEMPERATURE HOBO WATER PRO CALIBRATION RECORD

Range 0 to 100°F

Accuracy ±0.4°F

| Sensor Serial Number | 37 deg F | | | 65 degF | | | 93 degF | | | P A S S | F A I L | Battery L I F E |
|----------------------------|-----------|-------|-------|-----------|-------|-------|-----------|-------|-------|------------------|------------------|-----------------------------|
| | BATH TEMP | | | BATH TEMP | | | BATH TEMP | | | | | |
| | 36.951 | | | 64.996 | | | 92.958 | | | | | |
| | Limits | | | Limits | | | Limits | | | | | |
| | 0.40 | deg F | OBSVD | 0.40 | deg F | OBSVD | 0.40 | deg F | OBSVD | | | |
| | -0.40 | deg F | ERROR | -0.40 | deg F | ERROR | -0.40 | deg F | ERROR | | | |
| WB11 - ½ ft | 1134040 | 37.04 | 0.08 | 65.06 | | 0.06 | 92.90 | | -0.05 | ✓ | | 3.60 |
| WB12 - 3 ft | 1304851 | 36.89 | -0.06 | 65.02 | | 0.02 | 92.95 | | -0.01 | ✓ | | 3.60 |
| WB5 - 5 ft | 1304853 | 36.89 | -0.06 | 65.06 | | 0.06 | 93.05 | | 0.09 | ✓ | | 3.57 |
| WB4 - 3 ft | 1304854 | 36.89 | -0.06 | 65.02 | | 0.02 | 93.05 | | 0.09 | ✓ | | 3.57 |
| WB7 - 5 ft | 1304855 | 37.08 | 0.13 | 65.19 | | 0.19 | 93.19 | | 0.23 | ✓ | | 3.60 |
| WB12 - 5 ft | 1304857 | 37.04 | 0.08 | 65.19 | | 0.19 | 93.19 | | 0.23 | ✓ | | 3.57 |
| WB1 - 7 ft | 1304860 | 36.99 | 0.03 | 65.15 | | 0.15 | 93.19 | | 0.23 | ✓ | | 3.51 |
| WB10 - ½ ft | 1304861 | 36.94 | -0.01 | 65.10 | | 0.11 | 93.09 | | 0.13 | ✓ | | 3.57 |
| WB6 - 7 ft | 1304863 | 36.89 | -0.06 | 65.06 | | 0.06 | 93.05 | | 0.09 | ✓ | | 3.57 |
| WB1 - ½ ft | 1304864 | 36.94 | -0.01 | 65.10 | | 0.11 | 93.05 | | 0.09 | ✓ | | 3.60 |

SENSOR TYPE: HOBO Water Temp Pro U22-001

Remarks These Instruments are submerged in water for a long period of time and no calibration label will be attached.

The current calibration report will be in the Instrument Log.

WBN SCCW Testing Pre Cal 2011

Pre-Survey Calibrations (Continued)

| | |
|--|---|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | ID E44910 Page 1 of 2 Date 02/07/2011 |
|--|---|

METEOROLOGICAL MONITORING INSTRUMENTATION REPORT OF CALIBRATION

Calibrated For: Hydrothermal Compliance Date of Report: 02/07/2011

Item Description: HOBO WATER PRO TVA I.D. No.: E44910

Manufacturer: Onset Computer Corporation

Model: U22-001 CLS Instruction No.: 450.01-020

S/N No.: See Attached Sheet

Dispositioned to: CLS Norris Lab As-Left calibration in tolerance

Standards Used Log:

| I.D. No. | Description | Calibration Date | Calibration Due Date |
|----------|-----------------------------------|------------------|----------------------|
| 906527 | Azonix A1011-RS-XX Therm/Ohmmeter | 01/07/2011 | 01/07/2012 |
| 906535 | Burns Engineering 12001 PRT | 12/16/2010 | 12/16/2011 |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

This is to certify that all instrumentation, testing methods and personnel used comply with the requirements of the Central Laboratories Services (CLS) Quality Assurance Program which is designed to meet the requirements of ISO/IEC 17025, 10 CFR 50 Appendix B and ANSI N45.2-1971, and ANSI/NCSL Z540-1-1994. Standards used are traceable to the National Institute of Standards and Technology (NIST), officially recognized agencies, commercially accepted practices or natural physical constants. This report shall not be reproduced except in full, without the written approval of CLS.

Calibrated By: *Debbie Hanson*

Approved By: *Randy Lingen*

Date Approved: 2/9/11

| | | | |
|---|--|--|---|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | | | ID E44910 Page 2 of 2 Date 02/07/2011 |
|---|--|--|---|

WATER TEMPERATURE HOBO WATER PRO CALIBRATION RECORD

Range 0 to 100°F

Accuracy ±0.4°F

| Sensor Serial Number | 37 deg F | | | 65 degF | | | 93 degF | | | P A S S | F A I L | Battery L I F E |
|----------------------------|-----------|-------|-------|-----------|-------|-------|-----------|-------|-------|------------------|------------------|-----------------------------|
| | BATH TEMP | | | BATH TEMP | | | BATH TEMP | | | | | |
| | 36.951 | | | 64.996 | | | 92.958 | | | | | |
| | Limits | | | Limits | | | Limits | | | | | |
| | 0.40 | deg F | OBSVD | 0.40 | deg F | OBSVD | 0.40 | deg F | OBSVD | | | |
| | -0.40 | deg F | ERROR | -0.40 | deg F | ERROR | -0.40 | deg F | ERROR | | | |
| WB4 - 5 ft | 1304865 | 37.04 | 0.08 | 65.15 | 0.15 | | 93.14 | 0.18 | | ✓ | | 3.57 |
| WB9 - ½ ft | 1304866 | 36.94 | -0.01 | 65.10 | 0.11 | | 93.09 | 0.13 | | ✓ | | 3.57 |
| WB3 - 7 ft | 1304867 | 36.89 | -0.06 | 65.06 | 0.06 | | 93.09 | 0.13 | | ✓ | | 3.60 |
| WB6 - 3 ft | 1304868 | 36.94 | -0.01 | 65.10 | 0.11 | | 93.09 | 0.13 | | ✓ | | 3.57 |
| WB9 - 7 ft | 1304870 | 37.04 | 0.08 | 65.19 | 0.19 | | 93.19 | 0.23 | | ✓ | | 3.57 |
| WB1 - 3 ft | 1304872 | 36.89 | -0.06 | 65.02 | 0.02 | | 93.00 | 0.04 | | ✓ | | 3.60 |
| WB2 - 7 ft | 1304874 | 37.08 | 0.13 | 65.19 | 0.19 | | 93.19 | 0.23 | | ✓ | | 3.57 |
| WB10 - 5 ft | 1304877 | 36.84 | -0.11 | 64.97 | -0.02 | | 93.00 | 0.04 | | ✓ | | 3.57 |
| WB11 - 5 ft | 1304878 | 37.08 | 0.13 | 65.19 | 0.19 | | 93.19 | 0.23 | | ✓ | | 3.57 |

SENSOR TYPE: HOBO Water Temp Pro U22-001

Remarks These Instruments are submerged in water for a long period of time and no calibration label will be attached.

The current calibration report will be in the Instrument Log.

WBN SCCW Testing Pre Cal 2011

Pre-Survey Calibrations (Continued)

| | |
|--|---|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | ID E44911 Page 1 of 2 Date 02/08/2011 |
|--|---|

METEOROLOGICAL MONITORING INSTRUMENTATION REPORT OF CALIBRATION

Calibrated For: Hydrothermal Compliance Date of Report: 02/08/2011

Item Description: HOBO WATER PRO TVA I.D. No.: E44911

Manufacturer: Onset Computer Corporation

Model: U22-001 CLS Instruction No.: 450.01-020

S/N No.: See Attached Sheet

Dispositioned to: CLS Norris Lab As-Left calibration in tolerance

Standards Used Log:

| I.D. No. | Description | Calibration Date | Calibration Due Date |
|----------|-----------------------------------|------------------|----------------------|
| 906527 | Azonix A1011-RS-XX Therm/Ohmmeter | 01/07/2011 | 01/07/2012 |
| 906535 | Burns Engineering 12001 PRT | 12/16/2010 | 12/16/2011 |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

This is to certify that all instrumentation, testing methods and personnel used comply with the requirements of the Central Laboratories Services (CLS) Quality Assurance Program which is designed to meet the requirements of ISO/IEC 17025, 10 CFR 50 Appendix B and ANSI N45.2-1971, and ANSI/NCSS Z540-1-1994. Standards used are traceable to the National Institute of Standards and Technology (NIST), officially recognized agencies, commercially accepted practices or natural physical constants. This report shall not be reproduced except in full, without the written approval of CLS.

Calibrated By: Robin Hansen

Approved By: Randy Long
Date Approved: 2/9/11

| | |
|---|---|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | ID E44911 Page 2 of 2 Date 02/08/2011 |
|---|---|

WATER TEMPERATURE HOBO WATER PRO CALIBRATION RECORD

Range 0 to 100°F

Accuracy ±0.4°F

| Sensor Serial Number | 37 deg F | | | 65 degF | | | 93 degF | | | P A S S | F A I L | Battery L I F E |
|----------------------------|-----------|-------|-------|-----------|-------|-------|-----------|-------|-------|------------------|------------------|-----------------------------|
| | BATH TEMP | | | BATH TEMP | | | BATH TEMP | | | | | |
| | 36.951 | | | 64.997 | | | 92.959 | | | | | |
| | Limits | | | Limits | | | Limits | | | | | |
| | 0.40 | deg F | OBSVD | 0.40 | deg F | OBSVD | 0.40 | deg F | OBSVD | | | |
| | -0.40 | deg F | ERROR | -0.40 | deg F | ERROR | -0.40 | deg F | ERROR | | | |
| WB5 - ½ ft | 1304882 | 36.89 | -0.06 | 65.06 | | 0.06 | 93.05 | | 0.09 | ✓ | | 3.57 |
| WB6 - ½ ft | 1304883 | 37.13 | 0.18 | 65.23 | | 0.23 | 93.19 | | 0.23 | ✓ | | 3.57 |
| WB8 - 3 ft | 1304886 | 36.94 | -0.01 | 65.06 | | 0.06 | 93.05 | | 0.09 | ✓ | | 3.57 |
| WB2 - 3 ft | 1304888 | 36.94 | -0.01 | 65.06 | | 0.06 | 93.05 | | 0.09 | ✓ | | 3.57 |
| WB4 - 7 ft | 1304889 | 37.04 | 0.08 | 65.10 | | 0.11 | 93.09 | | 0.13 | ✓ | | 3.57 |
| WB7 - 7 ft | 1304890 | 36.94 | -0.01 | 65.06 | | 0.06 | 93.05 | | 0.09 | ✓ | | 3.60 |
| WB2 - 5 ft | 1304891 | 36.99 | 0.03 | 65.19 | | 0.19 | 93.19 | | 0.23 | ✓ | | 3.60 |
| WB7 - ½ ft | 1305136 | 36.84 | -0.11 | 64.97 | | -0.02 | 93.00 | | 0.04 | ✓ | | 3.57 |
| WB8 - ½ ft | 1305139 | 37.08 | 0.13 | 65.27 | | 0.28 | 93.23 | | 0.27 | ✓ | | 3.60 |

SENSOR TYPE: HOBO Water Temp Pro U22-001

Remarks These Instruments are submerged in water for a long period of time and no calibration label will be attached.

The current calibration report will be in the Instrument Log.

NBN SCCW Testing Pre Cal 2011

Pre-Survey Calibrations (Continued)

| | |
|--|---|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | ID E44912 Page 1 of 2 Date 02/08/2011 |
|--|---|

METEOROLOGICAL MONITORING INSTRUMENTATION REPORT OF CALIBRATION

Calibrated For: Hydrothermal Compliance Date of Report: 02/08/2011

Item Description: HOBO WATER PRO TVA I.D. No.: E44912

Manufacturer: Onset Computer Corporation

Model: U22-001 CLS Instruction No.: 450.01-020

S/N No.: See Attached Sheet

Dispositioned to: CLS Norris Lab As-Left calibration in tolerance

Standards Used Log:

| I.D. No. | Description | Calibration Date | Calibration Due Date |
|----------|-----------------------------------|------------------|----------------------|
| 906527 | Azonix A1011-RS-XX Therm/Ohmmeter | 01/07/2011 | 01/07/2012 |
| 906535 | Burns Engineering 12001 PRT | 12/16/2010 | 12/16/2011 |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

This is to certify that all instrumentation, testing methods and personnel used comply with the requirements of the Central Laboratories Services (CLS) Quality Assurance Program which is designed to meet the requirements of ISO/IEC 17025, 10 CFR 50 Appendix B and ANSI N45.2-1971, and ANSI/NCSL Z540-1-1994. Standards used are traceable to the National Institute of Standards and Technology (NIST), officially recognized agencies, commercially accepted practices or natural physical constants. This report shall not be reproduced except in full, without the written approval of CLS.

Calibrated By: *Heidi Hanson* Approved By: *Randy Coyer*

Date Approved: 2/9/11

| | | |
|---|--|---|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | | ID E44912 Page 2 of 2 Date 02/08/2011 |
|---|--|---|

WATER TEMPERATURE HOBO WATER PRO CALIBRATION RECORD

Range 0 to 100°F

Accuracy ±0.4°F

| | Sensor Serial Number | 37 deg F | | | 65 degF | | | 93 degF | | | P A S S | F A I L | Battery L I F E |
|-------------|----------------------------|-----------|-------|-------|-----------|-------|-------|-----------|-------|-------|------------------|------------------|-----------------------------|
| | | BATH TEMP | | | BATH TEMP | | | BATH TEMP | | | | | |
| | | 36.951 | | | 64.997 | | | 92.959 | | | | | |
| | | Limits | | | Limits | | | Limits | | | | | |
| | | 0.40 | deg F | OBSVD | 0.40 | deg F | OBSVD | 0.40 | deg F | OBSVD | | | |
| | | -0.40 | deg F | ERROR | -0.40 | deg F | ERROR | -0.40 | deg F | ERROR | | | |
| WB9 - 3 ft | 1305140 | 37.08 | | 0.13 | 65.23 | | 0.23 | 93.23 | | 0.27 | ✓ | | 3.60 |
| WB12 - ½ ft | 1305141 | 36.94 | | -0.01 | 65.15 | | 0.15 | 93.23 | | 0.27 | ✓ | | 3.57 |
| WB8 - 7 ft | 1305143 | 37.23 | | 0.27 | 64.97 | | -0.02 | 92.95 | | -0.01 | ✓ | | 3.60 |
| WB3 - 3 ft | 1305144 | 36.94 | | -0.01 | 65.06 | | 0.06 | 93.05 | | 0.09 | ✓ | | 3.57 |
| WB9 - 5 ft | 1305150 | 36.84 | | -0.11 | 65.02 | | 0.02 | 93.00 | | 0.04 | ✓ | | 3.57 |
| WB2 - ½ ft | 1305152 | 36.89 | | -0.06 | 65.10 | | 0.11 | 93.14 | | 0.18 | ✓ | | 3.57 |
| WB11 - 7 ft | 1305153 | 37.04 | | 0.08 | 65.06 | | 0.06 | 93.05 | | 0.09 | ✓ | | 3.57 |
| WB12 - 7 ft | 1305155 | 36.94 | | -0.01 | 65.06 | | 0.06 | 93.00 | | 0.04 | ✓ | | 3.57 |
| WB10 - 3 ft | 1305156 | 36.94 | | -0.01 | 65.10 | | 0.11 | 93.09 | | 0.13 | ✓ | | 3.57 |
| WB3 - ½ ft | 1305159 | 37.04 | | 0.08 | 65.19 | | 0.19 | 93.14 | | 0.18 | ✓ | | 3.60 |

SENSOR TYPE: HOBO Water Temp Pro U22-001

Remarks These Instruments are submerged in water for a long period of time and no calibration label will be attached.

The current calibration report will be in the Instrument Log.

WBN SCCW Testing Pre Cal 2011

Pre-Survey Calibrations (Continued)

| | |
|---|---|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | ID E44913 Page 1 of 2 Date 02/09/2011 |
|---|---|

METEOROLOGICAL MONITORING INSTRUMENTATION REPORT OF CALIBRATION

Calibrated For: Hydrothermal Compliance Date of Report: 02/09/2011

Item Description: HOBO WATER PRO TVA I.D. No.: E44913

Manufacturer: Onset Computer Corporation

Model: U22-001 CLS Instruction No.: 450.01-020

S/N No.: See Attached Sheet

Dispositioned to: CLS Norris Lab As-Left calibration in tolerance

Standards Used Log:

| I.D. No. | Description | Calibration Date | Calibration Due Date |
|----------|-----------------------------------|------------------|----------------------|
| 906527 | Azonix A1011-RS-XX Therm/Ohmmeter | 01/07/2011 | 01/07/2012 |
| 906535 | Burns Engineering 12001 PRT | 12/16/2010 | 12/16/2011 |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

This is to certify that all instrumentation, testing methods and personnel used comply with the requirements of the Central Laboratories Services (CLS) Quality Assurance Program which is designed to meet the requirements of ISO/IEC 17025, 10 CFR 50 Appendix B and ANSI N45.2-1971, and ANSI/NCSL Z540-1-1994. Standards used are traceable to the National Institute of Standards and Technology (NIST), officially recognized agencies, commercially accepted practices or natural physical constants. This report shall not be reproduced except in full, without the written approval of CLS.

Calibrated By: Debbie Houser

Approved By: Randy Cosper

Date Approved: 2/9/11

| | |
|---|---|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | ID E44913 Page 2 of 2 Date 02/09/2011 |
|---|---|

WATER TEMPERATURE HOBO WATER PRO CALIBRATION RECORD

Range 0 to 100°F

Accuracy ±0.4°F

| Sensor Serial Number | 37 deg F | | | 65 degF | | | 93 degF | | | P A S S | F A I L | Battery L I F E |
|----------------------------|-----------|-------|-------|-----------|-------|-------|-----------|-------|-------|------------------|------------------|-----------------------------|
| | BATH TEMP | | | BATH TEMP | | | BATH TEMP | | | | | |
| | 36.948 | | | 64.998 | | | 92.962 | | | | | |
| | Limits | | | Limits | | | Limits | | | | | |
| | 0.40 | deg F | OBSVD | 0.40 | deg F | OBSVD | 0.40 | deg F | OBSVD | | | |
| | -0.40 | deg F | ERROR | -0.40 | deg F | ERROR | -0.40 | deg F | ERROR | | | |
| WB7 - 3 ft | 1305160 | 36.94 | -0.01 | 65.10 | | 0.10 | 93.05 | | 0.08 | ✓ | | 3.57 |
| WB6 - 5 ft | 1305161 | 36.94 | -0.01 | 65.06 | | 0.06 | 93.05 | | 0.08 | ✓ | | 3.57 |
| WB5 - 3 ft | 1305164 | 36.94 | -0.01 | 65.10 | | 0.10 | 93.09 | | 0.13 | ✓ | | 3.57 |
| WB8 - 5 ft | 1305174 | 37.04 | 0.09 | 65.15 | | 0.15 | 93.14 | | 0.18 | ✓ | | 3.60 |
| WB11 - 3 ft | 1305176 | 37.08 | 0.13 | 65.27 | | 0.27 | 93.28 | | 0.32 | ✓ | | 3.57 |
| WB1 - 5 ft | 1305177 | 36.89 | -0.06 | 65.02 | | 0.02 | 93.00 | | 0.03 | ✓ | | 3.57 |
| WB10 - 7 ft | 1305179 | 37.08 | 0.13 | 65.19 | | 0.19 | 93.19 | | 0.22 | ✓ | | 3.57 |
| WB5 - 7 ft | 1305182 | 36.89 | -0.06 | 65.06 | | 0.06 | 93.05 | | 0.08 | ✓ | | 3.60 |

SENSOR TYPE: HOBO Water Temp Pro U22-001

Remarks These Instruments are submerged in water for a long period of time and no calibration label will be attached.

The current calibration report will be in the Instrument Log.

WBN SCCW Testing Pre Cal 2011

Pre-Survey Calibrations (Continued)

| | |
|--|---|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | ID E44914 Page 1 of 2 Date 02/09/2011 |
|--|---|

METEOROLOGICAL MONITORING INSTRUMENTATION REPORT OF CALIBRATION

Calibrated For: Hydrothermal Compliance Date of Report: 02/09/2011

Item Description: HOBO WATER PRO TVA I.D. No.: E44914

Manufacturer: Onset Computer Corporation

Model: U22-001 CLS Instruction No.: 450.01-020



S/N No.: See Attached Sheet

Dispositioned to: CLS Norris Lab As-Left calibration in tolerance

Standards Used Log:

| I.D. No. | Description | Calibration Date | Calibration Due Date |
|----------|-----------------------------------|------------------|----------------------|
| 906527 | Azonix A1011-RS-XX Therm/Ohmmeter | 01/07/2011 | 01/07/2012 |
| 906535 | Burns Engineering 12001 PRT | 12/16/2010 | 12/16/2011 |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

This is to certify that all instrumentation, testing methods and personnel used comply with the requirements of the Central Laboratories Services (CLS) Quality Assurance Program which is designed to meet the requirements of ISO/IEC 17025, 10 CFR 50 Appendix B and ANSI N45.2-1971, and ANSI/NCSS Z540-1-1994. Standards used are traceable to the National Institute of Standards and Technology (NIST), officially recognized agencies, commercially accepted practices or natural physical constants. This report shall not be reproduced except in full, without the written approval of CLS.

Calibrated By:  Approved By: 

Date Approved: 2/9/11

Post-Survey Calibrations

| | |
|--|---|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | ID E44909 Page 1 of 2 Date 06/21/2011 |
|--|---|

METEOROLOGICAL MONITORING INSTRUMENTATION REPORT OF CALIBRATION

Calibrated For: Hydrothermal Compliance Date of Report: 06/21/2011

Item Description: HOB0 WATER PRO TVA I.D. No.: E44909

Manufacturer: Onset Computer Corporation

Model: U22-001 CLS Instruction No.: 450.01-020

S/N No.: See Attached Sheet

Dispositioned to: CLS Norris Lab As-Left calibration in tolerance

Standards Used Log

| I.D. No. | Description | Calibration Date | Calibration Due Date |
|----------|-----------------------------------|------------------|----------------------|
| 906527 | Azonix A1011-RS-XX Therm/Ohmmeter | 01/07/2011 | 01/07/2012 |
| 906535 | Burns Engineering 12001 PRT | 12/16/2010 | 12/16/2011 |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

This is to certify that all instrumentation, testing methods and personnel used comply with the requirements of the Central Laboratories Services (CLS) Quality Assurance Program which is designed to meet the requirements of ISO/IEC 17025, 10 CFR 50 Appendix B and ANSI N45.2-1971, and ANSI/NCSL Z540-1-1994. Standards used are traceable to the National Institute of Standards and Technology (NIST), officially recognized agencies, commercially accepted practices or natural physical constants. This report shall not be reproduced except in full, without the written approval of CLS.

Calibrated By: Heather Hansen

Approved By: Randy Lora

Date Approved: 6/30/11

TENNESSEE VALLEY AUTHORITY
CENTRAL LABORATORIES SERVICES
 400 W. Summit Hill Drive, Mail Stop SPB BA-K
 Knoxville, Tennessee 37902
 Phone: (865) 632-2304 Fax: (865) 632-4996

ID E44909
 Page 2 of 2
 Date 06/21/2011

WATER TEMPERATURE HOBO WATER PRO CALIBRATION RECORD

Range 0 to 100°F

Accuracy ±0.4°F

| | Sensor Serial Number | 37 deg F | | | 65 degF | | | 93 degF | | | P A S S | F A I L | Battery L I F E |
|-------------|----------------------------|-----------|-------|-------|-----------|-------|-------|-----------|-------|-------|------------------|------------------|-----------------------------|
| | | BATH TEMP | | | BATH TEMP | | | BATH TEMP | | | | | |
| | | 36.955 | | | 65.004 | | | 92.969 | | | | | |
| | | Limits | | | Limits | | | Limits | | | | | |
| | | 0.40 | deg F | OBSVD | 0.40 | deg F | OBSVD | 0.40 | deg F | OBSVD | | | |
| | | -0.40 | deg F | ERROR | -0.40 | deg F | ERROR | -0.40 | deg F | ERROR | | | |
| WB11 - ½ ft | 1134040 | 36.99 | | 0.03 | 65.10 | | 0.10 | 93.00 | | 0.03 | ✓ | | 3.60 |
| WB12 - 3 ft | 1304851 | 36.89 | | -0.06 | 65.02 | | 0.01 | 93.00 | | 0.03 | ✓ | | 3.60 |
| WB5 - 5 ft | 1304853 | 36.94 | | -0.02 | 65.06 | | 0.05 | 93.09 | | 0.12 | ✓ | | 3.57 |
| WB4 - 3 ft | 1304854 | 36.94 | | -0.02 | 65.06 | | 0.05 | 93.05 | | 0.08 | ✓ | | 3.57 |
| WB7 - 5 ft | 1304855 | 37.08 | | 0.13 | 65.19 | | 0.18 | 93.19 | | 0.22 | ✓ | | 3.60 |
| WB12 - 5 ft | 1304857 | 37.08 | | 0.13 | 65.23 | | 0.23 | 93.19 | | 0.22 | ✓ | | 3.57 |
| WB1 - 7 ft | 1304860 | 37.04 | | 0.08 | 65.19 | | 0.18 | 93.19 | | 0.22 | ✓ | | 3.57 |
| WB10 - ½ ft | 1304861 | 36.99 | | 0.03 | 65.10 | | 0.10 | 93.14 | | 0.17 | ✓ | | 3.57 |
| WB6 - 7 ft | 1304863 | 36.89 | | -0.06 | 65.06 | | 0.05 | 93.09 | | 0.12 | ✓ | | 3.60 |
| WB1 - ½ ft | 1304834 | 36.94 | | -0.02 | 65.10 | | 0.10 | 93.09 | | 0.12 | ✓ | | 3.60 |

SENSOR TYPE: HOBO Water Temp Pro U22-001

All measurement ratios between the standards referenced in this instruction and the M & TE calibrated are greater than or equal to 4:1 except as noted.

Remarks These Instruments are submerged in water for a long period of time and no calibration label will be attached.

The current calibration report will be in the Instrument Log.

Initial Pre Calibration.

Post-Survey Calibrations (Continued)

| | |
|--|---|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | ID E44910 Page 1 of 2 Date 06/21/2011 |
|--|---|

METEOROLOGICAL MONITORING INSTRUMENTATION REPORT OF CALIBRATION

Calibrated For: Hydrothermal Compliance Date of Report: 06/21/2011

Item Description: HOBO WATER PRO TVA I.D. No.: E44910

Manufacturer: Onset Computer Corporation

Model: U22-001 CLS Instruction No.: 450.01-020

S/N No.: See Attached Sheet

Dispositioned to: CLS Norris Lab As-Left calibration in tolerance

Standards Used Log.

| I.D. No. | Description | Calibration Date | Calibration Due Date |
|----------|-----------------------------------|------------------|----------------------|
| 906527 | Azonix A1011-RS-XX Therm/Ohmmeter | 01/07/2011 | 01/07/2012 |
| 906535 | Burns Engineering 12001 PRT | 12/16/2010 | 12/16/2011 |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

This is to certify that all instrumentation, testing methods and personnel used comply with the requirements of the Central Laboratories Services (CLS) Quality Assurance Program which is designed to meet the requirements of ISO/IEC 17025, 10 CFR 50 Appendix B and ANSI N45.2-1971, and ANSI/NCCL Z540-1-1994. Standards used are traceable to the National Institute of Standards and Technology (NIST), officially recognized agencies, commercially accepted practices or natural physical constants. This report shall not be reproduced except in full, without the written approval of CLS

Calibrated By: Heidi Houser

Approved By: Randy Loper

Date Approved: 6/30/11

| | |
|---|---|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | ID E44910 Page 2 of 2 Date 06/21/2011 |
|---|---|

WATER TEMPERATURE HOBO WATER PRO CALIBRATION RECORD

Range 0 to 100°F

Accuracy ±0.4°F

| Sensor Serial Number | | 37 deg F | | | 65 degF | | | 93 degF | | | P A S S | F A I L | Battery L I F E |
|----------------------------|-------------|-----------|-------|-------------|-----------|--|-------------|-----------|---|--|------------------|------------------|-----------------------------|
| | | BATH TEMP | | | BATH TEMP | | | BATH TEMP | | | | | |
| | | 36.955 | | | 65.004 | | | 92.969 | | | | | |
| | | Limits | | | Limits | | | Limits | | | | | |
| | 0.40 deg F | OBSVD | | 0.40 deg F | OBSVD | | 0.40 deg F | OBSVD | | | | | |
| | -0.40 deg F | ERROR | | -0.40 deg F | ERROR | | -0.40 deg F | ERROR | | | | | |
| WB4 - 5 ft | 1304865 | 37.04 | 0.08 | 65.15 | 0.14 | | 93.14 | 0.17 | ✓ | | | 3.57 | |
| WB9 - ½ ft | 1304866 | 36.94 | -0.02 | 65.10 | 0.10 | | 93.14 | 0.17 | ✓ | | | 3.57 | |
| WB3 - 7 ft | 1304867 | 36.94 | -0.02 | 65.10 | 0.10 | | 93.09 | 0.12 | ✓ | | | 3.60 | |
| WB6 - 3 ft | 1304868 | 36.94 | -0.02 | 65.10 | 0.10 | | 93.04 | 0.07 | ✓ | | | 3.57 | |
| WB9 - 7 ft | 1304870 | 37.08 | 0.13 | 65.23 | 0.23 | | 93.23 | 0.27 | ✓ | | | 3.57 | |
| WB1 - 3 ft | 1304872 | 36.89 | -0.06 | 65.06 | 0.05 | | 93.00 | 0.03 | ✓ | | | 3.60 | |
| WB2 - 7 ft | 1304874 | 37.08 | 0.13 | 65.23 | 0.23 | | 93.23 | 0.27 | ✓ | | | 3.60 | |
| WB10 - 5 ft | 1304877 | 36.89 | -0.06 | 65.02 | 0.01 | | 93.00 | 0.03 | ✓ | | | 3.57 | |
| WB11 - 5 ft | 1304878 | 37.08 | 0.13 | 65.23 | 0.23 | | 93.23 | 0.27 | ✓ | | | 3.57 | |

SENSOR TYPE: HOBO Water Temp Pro U22-001

All measurement ratios between the standards referenced in this instruction and the M & TE calibrated are greater than or equal to 4:1 except as noted.

Remarks These Instruments are submerged in water for a long period of time and no calibration label will be attached.

The current calibration report will be in the Instrument Log.

Initial Pre Calibration.

Post-Survey Calibrations (Continued)

| | |
|--|--|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | ID E44911 Page 1 of 2 Date 06/21/2011 |
|--|--|

METEOROLOGICAL MONITORING INSTRUMENTATION REPORT OF CALIBRATION

Calibrated For: Hydrothermal Compliance Date of Report: 06/21/2011

Item Description: HOBO WATER PRO TVA I.D. No.: E44911

Manufacturer: Onset Computer Corporation

Model: U22-001 CLS Instruction No : 450.01-020

S/N No : See Attached Sheet

Dispositioned to: CLS Norris Lab As-Left calibration in tolerance

Standards Used Log:

| I.D. No. | Description | Calibration Date | Calibration Due Date |
|----------|-----------------------------------|------------------|----------------------|
| 906527 | Azonix A1011-RS-XX Therm/Ohmmeter | 01/07/2011 | 01/07/2012 |
| 906535 | Burns Engineering 12001 PRT | 12/16/2010 | 12/16/2011 |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

This is to certify that all instrumentation, testing methods and personnel used comply with the requirements of the Central Laboratories Services (CLS) Quality Assurance Program which is designed to meet the requirements of ISO/IEC 17025, 10 CFR 50 Appendix B and ANSI N45.2-1971, and ANSI/NCSS Z540-1-1994. Standards used are traceable to the National Institute of Standards and Technology (NIST), officially recognized agencies, commercially accepted practices or natural physical constants. This report shall not be reproduced except in full, without the written approval of CLS.

Calibrated By: *Debbie Houser* Approved By: *Randy Lopez*

Date Approved: 6/30/11

| | |
|---|---|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | ID E44911 Page 2 of 2 Date 06/21/2011 |
|---|---|

WATER TEMPERATURE HOBO WATER PRO CALIBRATION RECORD

Range 0 to 100°F

Accuracy ±0.4°F

| Sensor Serial Number | 37 deg F | | | 65 degF | | | 93 degF | | | P A S S | F A I L | Battery L I F E |
|----------------------------|-------------|-------|-------|-------------|-------|--|-------------|-------|---|------------------|------------------|-----------------------------|
| | BATH TEMP | | | BATH TEMP | | | BATH TEMP | | | | | |
| | 36.954 | | | 65.004 | | | 92.969 | | | | | |
| | Limits | | | Limits | | | Limits | | | | | |
| | 0.40 deg F | OBSVD | | 0.40 deg F | OBSVD | | 0.40 deg F | OBSVD | | | | |
| | -0.40 deg F | ERROR | | -0.40 deg F | ERROR | | -0.40 deg F | ERROR | | | | |
| WB5 - ½ ft | 1304882 | 36.94 | -0.02 | 65.06 | 0.05 | | 93.09 | 0.12 | ✓ | | 3.57 | |
| WB6 - ½ ft | 1304883 | 37.13 | 0.18 | 65.23 | 0.23 | | 93.19 | 0.22 | ✓ | | 3.57 | |
| WB8 - 3 ft | 1304886 | 36.94 | -0.02 | 65.06 | 0.05 | | 93.05 | 0.08 | ✓ | | 3.57 | |
| WB2 - 3 ft | 1304888 | 36.94 | -0.02 | 65.06 | 0.05 | | 93.05 | 0.08 | ✓ | | 3.57 | |
| WB4 - 7 ft | 1304889 | 37.04 | 0.08 | 65.15 | 0.14 | | 93.09 | 0.12 | ✓ | | 3.57 | |
| WB7 - 7 ft | 1304890 | 36.99 | 0.03 | 65.10 | 0.10 | | 93.05 | 0.08 | ✓ | | 3.60 | |
| WB2 - 5 ft | 1304891 | 37.04 | 0.08 | 65.19 | 0.18 | | 93.19 | 0.22 | ✓ | | 3.60 | |
| WB7 - ½ ft | 1305136 | 36.89 | -0.06 | 65.02 | 0.01 | | 93.00 | 0.03 | ✓ | | 3.57 | |
| WB8 - ½ ft | 1305139 | 37.08 | 0.13 | 65.27 | 0.27 | | 93.28 | 0.31 | ✓ | | 3.57 | |

SENSOR TYPE: HOBO Water Temp Pro U22-001

All measurement ratios between the standards referenced in this instruction and the M & TE calibrated are greater than or equal to 4:1 except as noted.

Remarks These instruments are submerged in water for a long period of time and no calibration label will be attached.

The current calibration report will be in the Instrument Log.

Initial Pre Calibration.

Post-Survey Calibrations (Continued)

| | |
|--|---|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | ID E44912 Page 1 of 2 Date 06/21/2011 |
|--|---|

METEOROLOGICAL MONITORING INSTRUMENTATION REPORT OF CALIBRATION

Calibrated For: Hydrothermal Compliance Date of Report: 06/21/2011

Item Description: HOB0 WATER PRO TVA I.D. No.: E44912

Manufacturer: Onset Computer Corporation

Model: U22-001 CLS Instruction No.: 450.01-020

S/N No.: See Attached Sheet

Dispositioned to: CLS Norris Lab As-Left calibration in tolerance

Standards Used Log:

| I.D. No. | Description | Calibration Date | Calibration Due Date |
|----------|-----------------------------------|------------------|----------------------|
| 906527 | Azonix A1011-RS-XX Therm/Ohmmeter | 01/07/2011 | 01/07/2012 |
| 906535 | Burns Engineering 12001 PRT | 12/16/2010 | 12/16/2011 |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

This is to certify that all instrumentation, testing methods and personnel used comply with the requirements of the Central Laboratories Services (CLS) Quality Assurance Program which is designed to meet the requirements of ISO/IEC 17025, 10 CFR 50 Appendix B and ANSI N45.2-1971, and ANSI/NCSL Z540-1-1994. Standards used are traceable to the National Institute of Standards and Technology (NIST), officially recognized agencies, commercially accepted practices or natural physical constants. This report shall not be reproduced except in full, without the written approval of CLS.

Calibrated By: *Shelbie Hansen*

Approved By: *Randy Lopez*

Date Approved: 6/30/11

| | | | |
|---|--|--|---|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | | | ID E44912 Page 2 of 2 Date 06/21/2011 |
|---|--|--|---|

WATER TEMPERATURE HOBO WATER PRO CALIBRATION RECORD

Range 0 to 100°F

Accuracy ±0.4°F

| | | 37 deg F | | | 65 degF | | | 93 degF | | | P A S S | F A I L | Battery L I E |
|-------------|---------|-----------|-------|-------|-----------|-------|-------|-----------|-------|-------|------------------|------------------|------------------------|
| | | BATH TEMP | | | BATH TEMP | | | BATH TEMP | | | | | |
| | | 36.954 | | | 65.004 | | | 92.969 | | | | | |
| Sensor | Serial | Limits | | | Limits | | | Limits | | | | | |
| Number | | 0.40 | deg F | OBSVD | 0.40 | deg F | OBSVD | 0.40 | deg F | OBSVD | | | |
| | | -0.40 | deg F | ERROR | -0.40 | deg F | ERROR | -0.40 | deg F | ERROR | | | |
| WB9 - 3 ft | 1305140 | 37.13 | | 0.18 | 65.23 | | 0.23 | 93.23 | | 0.27 | ✓ | | 3.60 |
| WB12 - ½ ft | 1305141 | 36.94 | | -0.02 | 65.15 | | 0.14 | 93.23 | | 0.27 | ✓ | | 3.57 |
| WB8 - 7 ft | 1305143 | 36.84 | | -0.11 | 64.97 | | -0.03 | 92.95 | | -0.02 | ✓ | | 3.57 |
| WB3 - 3 ft | 1305144 | 36.99 | | 0.03 | 65.10 | | 0.10 | 93.05 | | 0.08 | ✓ | | 3.57 |
| WB9 - 5 ft | 1305150 | 36.89 | | -0.06 | 65.02 | | 0.01 | 93.00 | | 0.03 | ✓ | | 3.57 |
| WB2 - ½ ft | 1305152 | 36.94 | | -0.02 | 65.10 | | 0.10 | 93.19 | | 0.22 | ✓ | | 3.57 |
| WB11 - 7 ft | 1305153 | 36.89 | | -0.06 | 65.06 | | 0.05 | 93.05 | | 0.08 | ✓ | | 3.57 |
| WB12 - 7 ft | 1305155 | 36.94 | | -0.02 | 65.06 | | 0.05 | 93.05 | | 0.08 | ✓ | | 3.57 |
| WB10 - 3 ft | 1305156 | 36.99 | | 0.03 | 65.10 | | 0.10 | 93.14 | | 0.17 | ✓ | | 3.57 |
| WB3 - ½ ft | 1305159 | 37.08 | | 0.13 | 65.19 | | 0.18 | 93.19 | | 0.22 | ✓ | | 3.60 |

SENSOR TYPE: HOBO Water Temp Pro U22-001

All measurement ratios between the standards referenced in this instruction and the M & TE calibrated are greater than or equal to 4:1 except as noted.

Remarks These Instruments are submerged in water for a long period of time and no calibration label will be attached.

The current calibration report will be in the Instrument Log.

Initial Pre Calibration.

Post-Survey Calibrations (Continued)

| | |
|--|---|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | ID E44913 Page 1 of 2 Date 06/21/2011 |
|--|---|

METEOROLOGICAL MONITORING INSTRUMENTATION REPORT OF CALIBRATION

Calibrated For: Hydrothermal Compliance

Date of Report: 06/21/2011

Item Description: HOBO WATER PRO

TVA I.D. No.: E44913

Manufacturer: Onset Computer Corporation

Model: U22-001

CLS Instruction No.: 450.01-020

S/N No.: See Attached Sheet

Dispositioned to: CLS Norris Lab

As-Left calibration in tolerance

Standards Used Log:

| I.D. No. | Description | Calibration Date | Calibration Due Date |
|----------|-----------------------------------|------------------|----------------------|
| 906527 | Azonix A1011-RS-XX Therm/Ohmmeter | 01/07/2011 | 01/07/2012 |
| 906535 | Burns Engineering 12001 PRT | 12/16/2010 | 12/16/2011 |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

This is to certify that all instrumentation, testing methods and personnel used comply with the requirements of the Central Laboratories Services (CLS) Quality Assurance Program which is designed to meet the requirements of ISO/IEC 17025, 10 CFR 50 Appendix B and ANSI N45.2-1971, and ANSI/NCSL Z540-1-1994. Standards used are traceable to the National Institute of Standards and Technology (NIST), officially recognized agencies, commercially accepted practices or natural physical constants. This report shall not be reproduced except in full, without the written approval of CLS

Calibrated By: *Debbie Houser*

Approved By: *Randy Longen*

Date Approved: 6/30/11

| | | | | | |
|---|--|--|--|------|------------|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | | | | ID | E44913 |
| | | | | Page | 2 of 2 |
| | | | | Date | 06/21/2011 |

WATER TEMPERATURE HOBO WATER PRO CALIBRATION RECORD

Range 0 to 100°F

Accuracy ±0.4°F

| | Sensor Serial Number | 37 deg F | | | 65 degF | | | 93 degF | | | P A S S | F A I L | Battery L I F E |
|-------------|----------------------------|-----------|-------|-------|-----------|-------|-------|-----------|-------|-------|------------------|------------------|-----------------------------|
| | | BATH TEMP | | | BATH TEMP | | | BATH TEMP | | | | | |
| | | 36.955 | | | 65.007 | | | 92.966 | | | | | |
| | | Limits | deg F | OBSVD | Limits | deg F | OBSVD | Limits | deg F | OBSVD | | | |
| | | 0.40 | deg F | ERROR | 0.40 | deg F | ERROR | 0.40 | deg F | ERROR | | | |
| | | -0.40 | deg F | | -0.40 | deg F | | -0.40 | deg F | | | | |
| WB7 - 3 ft | 1305160 | 36.94 | | -0.02 | 65.10 | | 0.09 | 93.09 | | 0.13 | ✓ | | 3.57 |
| WB6 - 5 ft | 1305161 | 36.94 | | -0.02 | 65.10 | | 0.09 | 93.09 | | 0.13 | ✓ | | 3.57 |
| WB5 - 3 ft | 1305164 | 36.94 | | -0.02 | 65.10 | | 0.09 | 93.09 | | 0.13 | ✓ | | 3.57 |
| WB8 - 5 ft | 1305174 | 37.04 | | 0.08 | 65.19 | | 0.18 | 93.16 | | 0.19 | ✓ | | 3.57 |
| WB11 - 3 ft | 1305176 | 37.13 | | 0.18 | 65.27 | | 0.27 | 93.28 | | 0.32 | ✓ | | 3.57 |
| WB1 - 5 ft | 1305177 | 36.94 | | -0.02 | 65.02 | | 0.01 | 93.00 | | 0.03 | ✓ | | 3.57 |
| WB10 - 7 ft | 1305179 | 37.08 | | 0.13 | 65.23 | | 0.22 | 93.19 | | 0.22 | ✓ | | 3.57 |
| WB5 - 7 ft | 1305182 | 36.94 | | -0.02 | 65.06 | | 0.05 | 93.09 | | 0.13 | ✓ | | 3.60 |

SENSOR TYPE: HOBO Water Temp Pro U22-001

All measurement ratios between the standards referenced in this instruction and the M & TE calibrated are greater than or equal to 4:1 except as noted.

Remarks These Instruments are submerged in water for a long period of time and no calibration label will be attached.

The current calibration report will be in the Instrument Log.

Initial Pre Calibration.

Post-Survey Calibrations (Continued)

| | |
|--|---|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | ID E44914 Page 1 of 2 Date 06/21/2011 |
|--|---|

METEOROLOGICAL MONITORING INSTRUMENTATION REPORT OF CALIBRATION

Calibrated For: Hydrothermal Compliance Date of Report: 06/21/2011

Item Description: HOB0 WATER PRO TVA I.D. No.: E44914

Manufacturer: Onset Computer Corporation

Model: U22-001 CLS Instruction No.: 450.01-020

S/N No.: See Attached Sheet

Dispositioned to: CLS Norris Lab As-Left calibration in tolerance

Standards Used Log:

| I.D. No. | Description | Calibration Date | Calibration Due Date |
|----------|-----------------------------------|------------------|----------------------|
| 906527 | Azonix A1011-RS-XX Therm/Ohmmeter | 01/07/2011 | 01/07/2012 |
| 906535 | Burns Engineering 12001 PRT | 12/16/2010 | 12/16/2011 |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

This is to certify that all instrumentation, testing methods and personnel used comply with the requirements of the Central Laboratories Services (CLS) Quality Assurance Program which is designed to meet the requirements of ISO/IEC 17025, 10 CFR 50 Appendix B and ANSI N45.2-1971, and ANSI/NCSL Z540-1-1994. Standards used are traceable to the National Institute of Standards and Technology (NIST), officially recognized agencies, commercially accepted practices or natural physical constants. This report shall not be reproduced except in full, without the written approval of CLS.

Calibrated By:

Heather Howard

Approved By:

Randy Goren

Date Approved:

6/30/11

| | |
|---|---|
| TENNESSEE VALLEY AUTHORITY CENTRAL LABORATORIES SERVICES 400 W. Summit Hill Drive, Mail Stop SPB BA-K Knoxville, Tennessee 37902 Phone: (865) 632-2304 Fax: (865) 632-4996 | ID E44914 Page 2 of 2 Date 06/21/2011 |
|---|---|

WATER TEMPERATURE HOBO WATER PRO CALIBRATION RECORD

Range 0 to 100°F

Accuracy ±0.4°F

| Sensor Serial Number | 37 deg F | | | 65 degF | | | 93 degF | | | P A S S | F A I L | Battery L I F E |
|----------------------------|-----------|-------|-------|-----------|-------|-------|-----------|-------|-------|------------------|------------------|-----------------------------|
| | BATH TEMP | | | BATH TEMP | | | BATH TEMP | | | | | |
| | 36.955 | | | 65.007 | | | 92.966 | | | | | |
| | Limits | | | Limits | | | Limits | | | | | |
| | 0.40 | deg F | OBSVD | 0.40 | deg F | OBSVD | 0.40 | deg F | OBSVD | | | |
| | -0.40 | deg F | ERROR | -0.40 | deg F | ERROR | -0.40 | deg F | ERROR | | | |
| WB3 - 5 ft | 1305184 | 37.08 | 0.13 | 65.23 | 0.22 | | 93.19 | 0.22 | | ✓ | | 3.57 |
| WB4 - ½ ft | 1305192 | 37.04 | 0.08 | 65.19 | 0.18 | | 93.23 | 0.27 | | ✓ | | 3.60 |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |

SENSOR TYPE: HOBO Water Temp Pro U22-001

All measurement ratios between the standards referenced in this instruction and the M & TE calibrated are greater than or equal to 4:1 except as noted.

Remarks These Instruments are submerged in water for a long period of time and no calibration label will be attached.

The current calibration report will be in the Instrument Log.

Initial Pre Calibration.

APPENDIX B

WBN Outfall 113 NPDES Compliance Parameters

- Current Instantaneous Upstream Temperature:

Tu_i (measured at EDS Station 30 by the first sensor below a depth of 5 feet)

- Current 1-Hour Average Upstream Temperature:

$$Tu1_i = \frac{Tu_i + Tu_{i-1} + Tu_{i-2} + Tu_{i-3} + Tu_{i-4}}{5},$$

where the subscripts $i, i-1, i-2, i-3$, and $i-4$ denote the current and previous four 15-minute (0.25 hour) values of Tu

- Current Instantaneous Downstream Temperature:

$$Td_i = \frac{Td3_i + Td5_i + Td7_i}{3},$$

where $Td3_i$, $Td5_i$, and $Td7_i$ denote the current measurements of river temperature at the downstream end of the mixing zone at water depths 3 feet, 5 feet, and 7 feet, respectively

- Current 1-Hour Average Downstream Temperature:

$$Td1_i = \frac{Td_i + Td_{i-1} + Td_{i-2} + Td_{i-3} + Td_{i-4}}{5},$$

where the subscripts $i, i-1, i-2, i-3$, and $i-4$ denote the current and previous four 15-minute (0.25 hour) values of Td

- Current Instantaneous Temperature Rise:

$$\Delta T_i = Td_i - Tu_i$$

- Current 1-Hour Average Temperature Rise:

$$\Delta T1_i = \frac{\Delta T_i + \Delta T_{i-1} + \Delta T_{i-2} + \Delta T_{i-3} + \Delta T_{i-4}}{5},$$

where the subscripts i , $i-1$, $i-2$, $i-3$, and $i-4$ denote the current and previous four 15-minute (0.25 hour) values of ΔT

- Current Temperature Rate-of-Change:

$$TROC_i = \frac{Td_i - Td_{i-4}}{1 \text{ hour}},$$

- Current 1-Hour Average Temperature Rate-of-Change:

$$TROC1_i = \frac{TROC_i + TROC_{i-1} + TROC_{i-2} + TROC_{i-3} + TROC_{i-4}}{5}$$

where the subscripts i , $i-1$, $i-2$, $i-3$, and $i-4$ denote the current and previous four 15-minute (0.25 hour) values of TROC

Enclosure 5

**Summer 2012 Compliance Survey for Watts Bar Nuclear Plant Outfall
Passive Mixing Zone**

TENNESSEE VALLEY AUTHORITY
River Operations

**SUMMER 2012 COMPLIANCE SURVEY FOR WATTS BAR
NUCLEAR PLANT OUTFALL 113 PASSIVE MIXING ZONE**

Prepared by

Daniel P. Saint
and
Paul N. Hopping

Knoxville, Tennessee
January 2013



EXECUTIVE SUMMARY

The National Pollutant Discharge Elimination System (NPDES) Permit No. TN0020168 for Watts Bar Nuclear Plant (WBN) identifies the discharge of water to the Tennessee River from the Supplemental Condenser Cooling Water (SCCW) System as Outfall 113. Furthermore, the permit identifies that when there is no flow released from Watts Bar Dam (WBH), the effluent from Outfall 113 shall be regulated based on a passive mixing zone extending in the river from bank-to-bank and 1,000 feet downstream from the outfall. Compliance with the requirements for the passive mixing zone is to be achieved by two annual instream temperature surveys—one for winter conditions and one for summer conditions. Summarized in this report are the measurements, analyses, and results for the passive mixing zone survey performed for 2012 summer conditions. The survey was conducted between 22:00 CDT on August 30 and 06:00 CDT on August 31 (eight hours) and included the collection of temperature data at twelve temporary monitoring stations deployed across the downstream end of the passive mixing zone during a period of no flow in the river. The data were analyzed to determine the three instream compliance parameters specified in the NPDES permit for the outfall: the 1-hour average temperature at the downstream end of mixing zone, T_d ; the 1-hour average temperature rise from upstream to the downstream end of the mixing zone, ΔT ; and the 1-hour average temperature rate-of-change at the downstream end of the mixing zone, TROC. The measured parameters were compared to predicted values from the thermal plume model used by TVA to help determine the safe operation of Outfall 113. The results of the comparisons, in terms of maximum values observed during the no flow event, are as follows:

| Compliance Parameter | Model | Measured | NPDES Limit |
|----------------------|-------------|-------------|-------------|
| Maximum T_d | 80.9°F | 79.3°F | 86.9°F |
| Maximum ΔT | 1.6 F° | 1.8 F° | 5.4 F° |
| Maximum TROC | 0.7 F°/hour | 0.2 F°/hour | 3.6 F°/hr |

As shown, both the model and measured values were well below the NPDES limits for all the compliance parameters. Except for the maximum ΔT , values predicted by the model were larger than those measured in the survey. The maximum value of ΔT from the model underpredicted the measured value by 0.2 F°. This difference was caused by unnatural cooling of the upstream ambient temperature from leakage of cold water through Watts Bar Dam. Based on this, as well as the fact that differences of magnitude 0.2 F° easily fall within the factor of safety currently used in performing hydrothermal forecasts, the thermal plume model is yet considered fully adequate for determining the safe operation of the SCCW system. That is, in combination with TVA procedures for predicting the impact of the Outfall 113 effluent, the model continues to provide a high level of confidence that the SCCW system is being operated in a manner that protects the limits for T_d , ΔT , and TROC specified in the NPDES permit for the passive mixing zone.

TABLE OF CONTENTS

| | <u>Page No.</u> |
|---|-----------------|
| EXECUTIVE SUMMARY | i |
| INTRODUCTION | 1 |
| INSTREAM SURVEY | 2 |
| RESULTS | 3 |
| River Conditions | 3 |
| SCCW Conditions | 4 |
| Downstream End of Passive Mixing Zone | 4 |
| NPDES Compliance Parameters | 5 |
| CONCLUSIONS | 7 |
| REFERENCES | 8 |
| APPENDIX A | 19 |
| APPENDIX B | 27 |

LIST OF FIGURES

| | |
|---|----|
| Figure 1. Watts Bar Nuclear Plant Outfall 113 (SCCW) Mixing Zones | 9 |
| Figure 2. Location of HOBO Monitoring Stations | 10 |
| Figure 3. Schematic of HOBO Water Temperature Monitoring Stations | 10 |
| Figure 4. River Conditions | 11 |
| Figure 5. SCCW Conditions | 12 |
| Figure 6. HOBO Water Temperature Measurements | 13 |
| Figure 7. Local Instantaneous Temperature Rise for HOBO Measurements | 15 |
| Figure 8. Measured and Computed Compliance Parameters for Passive Mixing Zone | 18 |

LIST OF TABLES

| | |
|--|---|
| Table 1. NPDES Temperature Limits for Outfall 113 Mixing Zones | 1 |
| Table 2. Sources of Data for Passive Mixing Zone Survey | 2 |

SUMMER 2012 COMPLIANCE SURVEY FOR WATTS BAR NUCLEAR PLANT OUTFALL 113 PASSIVE MIXING ZONE

INTRODUCTION

Outfall 113 for the Watts Bar Nuclear Plant (WBN) includes the discharge of water to the Tennessee River from the Supplemental Condenser Cooling Water (SCCW) system. Due to the dynamic behavior of the thermal effluent in the river, the National Pollutant Discharge Elimination System (NPDES) Permit No. TN0020168 for the plant specifies two mixing zones for Outfall 113—one for active operation of the river and one for passive operation of the river (TDEC, 2010). The passive mixing zone corresponds to periods when the operation of Watts Bar Dam (WBH) produces no flow in the river (i.e., hydropower and/or spillway releases). The dimensions of the passive mixing zone extend from bank-to-bank and downstream 1,000 feet from the outfall. The active mixing zone applies to all other river flow conditions. The dimensions of the active mixing zone include the right-half of the river (facing downstream) and extend downstream 2,000 feet from the outfall. The passive and the active mixing zones are shown in Figure 1.

Table 1 summarizes the NPDES instream temperature limits for Outfall 113. The limits apply to both the active and passive mixing zones. Compliance for the active mixing zone is monitored by permanent instream water temperature stations situated in the right-half of the river. Due to issues associated with placing permanent stations in the left-half of the river, which contains the navigation channel, a thermal plume model is used to determine the safe operation of Outfall 113 for the passive mixing zone. To verify the thermal plume model, the NPDES permit specifies that two instream temperature surveys shall be conducted each year—one for winter conditions and one for summer conditions. The purpose of this report is to present the results for the passive mixing zone temperature survey performed for summer 2012 conditions. The survey was conducted between 22:00 CDT on August 30 and 06:00 CDT on August 31 (total eight hours). Provided herein is a brief summary of the survey method, presentations of the measurements and analyses, and discussions of the results and conclusions.

Table 1. NPDES Temperature Limits for Outfall 113 Mixing Zones

| Compliance Parameter | Sampling Period | NPDES Limit |
|--|------------------------|--------------------|
| Maximum Temperature, Downstream End of Mixing Zone, T_d | Running 1-hr | 86.9°F |
| Maximum Temperature Rise, Upstream to Downstream, ΔT | Running 1-hr | 5.4 F° |
| Maximum Temperature Rate-of-Change, TROC | Running 1-hr | ±3.6 F°/hr |

INSTREAM SURVEY

The instream survey included the deployment of temporary water temperature stations at twelve locations across the downstream end of the passive mixing zone. Data from these and other monitoring stations were analyzed to obtain measured values for the compliance parameters listed in Table 1. These were then compared with the corresponding values estimated from the SCCW thermal plume model.

The method of conducting the instream survey is the same as that used for the first such survey, performed for winter conditions on May 6, 2005 (McCall and Hopping, 2005). Table 2 provides a summary of the sources of data for the survey. WaterView, a monitoring system for tracking hydroplant operation and performance, was used to obtain measurements for the river discharge from Watts Bar Dam. The WBN Environmental Data Station (EDS) provided measurements from existing permanent monitoring stations for the nuclear plant. These included:

- The river upstream (ambient) water temperature, measured at the EDS Station 30, which is located at the exit of the powerhouse of Watts Bar Dam.
- The river water surface elevation (WSEL) at the EDS Station 30, also known as the tailwater elevation (TWEL) at Watts Bar Dam.
- The SCCW effluent temperature, measured at the EDS Station 32, which is located at the SCCW outfall.
- The SCCW effluent discharge, measured at the EDS Station 32.
- The local air temperature, measured at the EDS meteorological tower.

Table 2. Sources of Data for Passive Mixing Zone Survey

| Data | Source | Frequency |
|---------------------------------------|---------------------------------------|-----------|
| River Discharge from Watts Bar Dam | WaterView | 1 min |
| River ambient water temperature | WBN EDS Station 30 (Tailwater at WBH) | 15 min |
| River water surface elevation | WBN EDS Station 30 (Tailwater at WBH) | 15 min |
| SCCW effluent temperature | WBN EDS Station 32 (SCCW Outfall 113) | 15 min |
| SCCW effluent discharge | WBN EDS Station 32 (SCCW Outfall 113) | 15 min |
| Air temperature | WBN EDS Met Tower | 15 min |
| Passive mixing zone water temperature | Temporary HOBO Monitors | 1 min |

The water temperature at the downstream end of the Outfall 113 passive mixing zone was measured by the aforementioned temporary water temperature stations. Using a global positioning system (GPS) device, the stations were positioned at roughly equal intervals across the river, as shown in Figure 2. The temporary stations recorded water temperatures by using HOBO temperature monitors positioned at depths of 0.5, 3, 5, and 7 feet below the water surface. Shown in Figure 3 is a schematic of the temporary stations. The stations included a string of

HOBO monitors suspended from a tire float, with weights to anchor the station and to keep the sensor string vertical in the water column. The water temperature sensors imbedded in the HOBO monitors have an accuracy of about ± 0.4 F° and resolution of about 0.04 F°, which is comparable to the accuracy and resolution of temperature sensors used elsewhere by TVA for NPDES thermal compliance. The HOBO monitors include an internal data acquisition unit that was programmed to collect measurements once per minute. All the temperature probes used in the survey, including both those contained in the HOBO monitors and the thermistors at the permanent EDS monitoring stations, were calibrated by a quality program with equipment accuracies traceable to the National Institute of Standards and Technology (NIST). The calibration procedure is summarized in APPENDIX A. The temporary monitoring stations were deployed several hours before the beginning of the survey, and retrieved several hours after the end of the survey.

RESULTS

River Conditions

Figure 4 shows the measured ambient conditions of the river during the survey. Included are the river discharge, river water surface elevation, and river temperature, all at the exit of Watts Bar Dam. The river temperature at the exit of Watts Bar Dam serves as the upstream ambient river temperature for WBN Outfall 113. To provide a period of no flow in the river, releases from Watts Bar Dam were suspended between about 22:00 CDT on August 30 and 06:00 CDT on August 31, a total of eight hours (nighttime). Leading up to the survey, as the river flow was stepping down, the WSEL at the exit of Watts Bar Dam dropped approximately 2.7 feet, from about 683.5 feet msl to about 680.8 feet msl. During the survey, the WSEL slowly increased, due to backflow from the surrounding tailwater and leakage through the hydroturbines, returning to about 681.8 feet msl after six hours of no flow in the river. Afterwards, the WSEL slowly receded, reaching about 681.3 feet msl at the end of the survey.

The ambient river temperature was about 77.9°F at the beginning of the period of no flow. The temperature held steady at 77.9°F for the first three hours of the survey, and then began to slowly decrease, reaching 77.4°F at the end of the survey. This drop in ambient river temperature is common when strong thermal stratification exists behind Watts Bar Dam. During periods of no flow, leakage occurs through the hydroturbines at the dam. Previous studies have suggested the amount of leakage to be roughly 50 cfs for each hydro unit, or a total of 250 cfs for the entire powerhouse (Harper et. al, 1998). This leakage comes from the very bottom of Watts Bar Reservoir, the coldest part of the water column in front of the dam. As the leakage occurs, it slowly fills the bottom layers of the tailrace below the powerhouse, eventually reaching the elevation of the Station 30 sensors, which are suspended downward from the water surface. Cooling of the ambient river temperature monitor in this manner falsely increases the measured

temperature rise for the SCCW system. That is, the temperature rise is elevated not by warming from the SCCW effluent, but by “artificial” cooling of the upstream monitor via a process that is beyond the operational control of the SCCW system. In forecasting values for the WBN upstream ambient river temperature, the thermal plume model for the SCCW system does not include cooling that occurs as a result of leakage through the hydroturbines at Watts Bar Dam.

SCCW Conditions

During the survey, the SCCW system at WBN was thermally loaded and operating in “summer” mode. That is, the system was operating in a manner producing the largest possible release of heat to the river. Shown in Figure 5 are the measured conditions of the SCCW system during the survey. Included are the discharge and temperature of the SCCW effluent. During the survey, the average discharge of the SCCW system to the river was about 300 cfs. The root-mean-square variation in the SCCW discharge was only about 3.1 percent of the average—thus, from the standpoint of mixing processes in the river, the discharge was essentially constant. The SCCW effluent temperature decreased throughout the survey from about 86.2°F at the beginning of the survey to about 84.8°F at the end of the survey. This trend coincides with the falling nighttime air temperature, also shown in Figure 5 (note: the temperature of the water discharging from the Unit 1 cooling tower, which provides the source for Outfall 113, varies directly with the temperature of the ambient air that is drawn through the tower). The temperature rise of the Outfall 113 effluent relative to the upstream ambient river temperature, also shown in Figure 5, decreased in a similar fashion throughout the survey, from about 8.9 F° at the beginning of the survey to about 7.4 F° at the end of the survey.

Downstream End of Passive Mixing Zone

Shown in Figure 6 are the measurements from the HOBO temperature stations at the downstream end of the passive mixing zone. The stations are labeled consecutively from WB1 to WB12, with WB1 situated near the left-hand shoreline of the river and WB12 situated near the right-hand shoreline of the river (i.e., facing downstream—see Figure 2). In Figure 7, the HOBO data has been analyzed to produce contour plots of the local “instantaneous” water temperature rise (ΔT) relative to the SCCW ambient river temperature (i.e., given in Figure 4). The horizontal (x) axis of each contour plot is the span of the river from WB1 to WB12, and the vertical (y) axis is the water depth, from 0.5 feet to 7 feet. In this manner, the plots in Figure 7 represent images of the upper 7 feet of the water column in the river, looking downstream. Note that the depth scale in the Figure 7 plots is significantly distorted so that measurements can be viewed in a meaningful manner—that is, whereas the span of the x-axis is about 1000 feet, the span of the y-axis is only about 7 feet (0.007 times smaller). Plots are provided at the top of each hour from the beginning of the survey at 22:00 CDT on August 30 to the end of the survey at 06:00 CDT on August 31. The following behaviors are emphasized from Figure 6 and Figure 7:

- At the beginning of the survey, 22:00 CDT on August 30, effluent from the SCCW resides primarily in the right-hand-side of the river. The flow in the river prevents the effluent from spreading across the river; however, the deceleration of flow from Watts Bar Dam appears to have allowed some effluent to move into the middle portion of the cross section. The maximum local instantaneous temperature rise is about 2.0 F°, occurring in the upper 3 feet of the water column.
- Without any significant flow from Watts Bar Dam, outward spreading of the SCCW effluent is unhindered, reaching the left-hand-side of the river and propagating to the downstream end of the passive mixing zone. By 01:00 CDT on August 31, the maximum local instantaneous temperature rise is about 1.6 F° and occurs in the left-hand-side of the river.
- After 01:00 CDT, the effluent continues to spread back across the river, reaching the middle of the river by 02:00 CDT. By 03:00 CDT on August 31, five hours into the survey, the SCCW effluent has returned to the right-hand-side of the river and is fully distributed across the passive mixing zone. At this point, the maximum local instantaneous temperature rise is still about 1.6 F°, occurring at several locations in the cross section, primarily in the upper 3 feet of the water column.
- In the remaining three hours of the survey, heat from the SCCW effluent continues to slowly backfill from the left-hand-side to the right-hand-side of the river. At the end of the survey, the maximum local instantaneous temperature rise is about 2.0 F°, occurring in the upper 3 feet of the water column in the left-hand-side of the river. Overall, however, at the end of the survey, there is very little temperature variation across the river—at most about 0.4 F°.

NPDES Compliance Parameters

Since heat from the SCCW effluent is distributed across the full width of the river, data from all of the HOBO stations were used to compute the NPDES compliance parameters, which is consistent with the dimensions of the passive mixing zone (i.e., the passive mixing zone spans the full width of the river). The compliance parameters examined include all those given in Table 1—the temperature at the downstream end of mixing zone, T_d ; the temperature rise from upstream to the downstream end of the mixing zone, ΔT ; and the temperature rate-of-change at the downstream end of the mixing zone, TROC. The fundamental equations used to compute the compliance parameters are provided in APPENDIX B, based on the criteria specified in the NPDES permit. The temperature at the downstream end of the mixing zone was determined from the HOBO measurements by averaging the readings from the sensors at depths 3, 5, and 7 feet for all twelve HOBO stations. The temperature rise was computed as the difference between the measured temperature at the downstream end of the mixing zone and the upstream temperature measured at Watts Bar Dam (i.e., Station 30). The temperature rate-of-change was

determined by the change in the measured temperature at the downstream end of the mixing zone from one hour to the next. The data were averaged over a period of one hour using 15-minute readings, as specified in the NPDES permit, and compared with the WBN thermal plume model. The measurements are presented in Figure 8, along with the results obtained by the thermal plume model. The following behaviors are emphasized:

- Temperature at the downstream end of the passive mixing zone, T_d : The maximum 1-hour average T_d estimated by the thermal plume model was 80.9°F, whereas the maximum measured value was about 79.3°F. Thus, the model overpredicted the maximum measured T_d by 1.6°F. Compared to the measurements, the increase in river temperature due to the no flow event was predicted to occur much more rapidly by the model. This is because the model assumes impacts due to changes in the river and/or Outfall 113 conditions are fully realized as a steady-state episode within one hour (i.e., the model time-step); whereas in reality, the actual time for the thermal plume to evolve is much longer. Both the predictions from the model and measurements from the survey were well below the NPDES limit of 86.9°F.
- Temperature rise, ΔT : The maximum 1-hour average ΔT predicted by the plume model was 1.6 F°, whereas the maximum measured value was about 1.9 F°. Thus, the model underpredicted the maximum measured temperature rise by 0.3 F°. For the reason cited above (i.e., computational time-step of one hour), the model predicted the maximum temperature rise to occur one hour into the no flow event. A close examination of the data reveals that the maximum measured value of the temperature rise occurred at end of the survey, when the impact of leakage at Watts Bar Dam reduced the upstream ambient river temperature relative to the model value (see previous discussion in section entitled “River Conditions”). The model value for the upstream ambient river temperature was 79.3°F, whereas due to leakage of cold water at Watts Bar Dam, the measured ambient temperature was unnaturally lowered to 77.4°F (i.e., 1.9 F° lower than the model value, see Figure 4). Both the predictions from the model and measurements from the survey were well below the NPDES limit of 5.4 F°.
- Temperature rate-of-change, TROC: The maximum 1-hour average TROC predicted by the plume model was 0.7 F°/hour, whereas the maximum measured value was about 0.2 F°/hour (absolute values). Thus, the model overpredicted the temperature rate-of-change by 0.5 F°/hour. Both the predictions from the model and measurements from the survey were well below the NPDES limit of ± 3.6 F°/hour.

CONCLUSIONS

The compliance survey for 2012 summer conditions was successful in measuring the NPDES instream water temperature parameters for the Outfall 113. These included the temperature, T_d , temperature rise, ΔT , and temperature rate-of-change, TROC, all at the downstream end of the passive mixing zone. The measurements were compared with values predicted by the thermal plume model that TVA currently uses to determine the safe operation of the SCCW system.

Since 2005, when the first compliance survey was performed for the Outfall 113 passive mixing zone, the model value for the maximum downstream temperature T_d , including that for the survey summarized herein, has always bounded the measured value for the maximum T_d . That is, the model value has always been greater than or equal to the measured value. Such is not the case, however, for ΔT and TROC. In this survey, the model value for the ΔT underpredicted the value for the maximum ΔT by 0.2 F°. The only other instance when the model underpredicted the actual ΔT was during the summer survey of 2011, when the model value for the maximum ΔT underpredicted the measured value by 0.1 F° (Saint and Hopping, 2011). As for temperature rate-of-change, the model value for the maximum TROC underpredicted the measured value by 0.3 F°/hour in the summer survey of 2005 (McCall and Hopping, 2006). These differences are not surprising in light of the fact that the model, like any mathematical representation of a complex physical process, contains inherent accuracy limitations. The TVA model for predicting the Outfall 113 thermal plume uses CORMIX, which has a stated accuracy of about 50% of the standard deviation of field measurements (Jirka, et al., 1996). Based on this, as well as the fact that differences as small as 0.2 F° for ΔT and 0.3 F°/hour for TROC fall within the factor of safety currently used by TVA in performing hydrothermal forecasts, the thermal plume model is yet considered fully adequate for determining the safe operation of the SCCW system. That is, in combination with TVA procedures for predicting the impact of the Outfall 113 effluent, the model continues to provide a high level of confidence that the SCCW system is being operated in a manner that protects the limits for T_d , ΔT , and TROC specified in the NPDES permit for the passive mixing zone.

REFERENCES

Harper, Walter L., and Bo Hadjerioua, Mark Reeves, Gary Hickman, and John Jenkinson, "Hydrodynamics and Water Temperature Modeling at Watts Bar SCCW Discharge Structure," TVA Resource Group, Water Management, Report No. WR98-1-85-142, November 1998.

Jirka, Gerhard H., Robert L. Doneker, and Steven W. Hinton, "User's Manual for CORMIX: A Hydrodynamic Mixing Zone Model and Decision Support System for Pollutant Discharges into Surface Waters," Office of Science and Technology, U.S. Environmental Protection Agency, Washington, DC, September 1996.

McCall, Michael J., and P.N. Hopping, "Summer 2005 Compliance Survey for Watts Bar Nuclear Plant Outfall 113 Passive Mixing Zone," TVA River Operations, Report No. WR2006-2-85-152, February 2006.

McCall, Michael J., and P.N. Hopping, "Winter 2005 Compliance Survey for Watts Bar Nuclear Plant Outfall 113 Passive Mixing Zone," TVA River Operations, Report No. WR2005-2-85-151, October 2005.

TDEC, *State of Tennessee NPDES Permit No. TN0020168*, Tennessee Department of Environment and Conservation, Issued June 2010.




Figure 1. Watts Bar Nuclear Plant Outfall 113 (SCCW) Mixing Zones




Figure 2. Location of HOBO Monitoring Stations




Figure 3. Schematic of HOBO Water Temperature Monitoring Stations




Figure 4. River Conditions




Figure 5. SCCW Conditions




Figure 6. HOBO Water Temperature Measurements




Figure 6 (Continued). HOBO Water Temperature Measurements




Figure 7. Local Instantaneous Temperature Rise for HOB0 Measurements




Figure 7 (Continued). Local Instantaneous Temperature Rise for HOB0 Measurements




Figure 7 (Continued). Local Instantaneous Temperature Rise for HOBO Measurements




Figure 8. Measured and Computed Compliance Parameters for Passive Mixing Zone

APPENDIX A

Calibration of NPDES Water Temperature Sensors



All sensors used by TVA for monitoring compliance of NPDES water temperature requirements are certified and maintained to meet the following industry and regulatory standards:

- ISO/IEC 17025—Quality assurance requirements for the competence to carry out sampling, testing, and calibrations using standard, non-standard, and laboratory-developed methods (ISO=International Organization for Standardization, IEC=International Electrotechnical Commission).
- 10CFR50 Appendix B—Quality assurance criteria for design, fabrication, construction, and testing of the structures, systems, and components of nuclear power plants (CFR=Code of Federal Regulations).
- 40CFR136—Guidelines establishing test procedures for the analysis of pollutants under the Clean Water Act.
- ANSI N45.2. 1971—Quality assurance requirements for Nuclear Power Plants (ANSI=American National Standards Institute).
- ANSI/NCSL Z540-1-1994—General requirements for calibration laboratories and equipment used for measurements and testing (NCSL=National Conference of Standards Laboratories).

The standard used to certify the thermistors for the permanent EDS stations and the temporary HOBO stations is traceable to the National Institute of Standards and Technology (NIST). The standard includes two pieces of equipment—a platinum resistance temperature detector (RTD) manufactured by Burns Engineering, Inc. and an ohmmeter manufactured by Azonix Inc. The latter is used to measure the resistance of the RTD (i.e., the resistance of platinum varies with temperature). The NTIS traceable calibration certificates for the Burns RTD and the Azonix ohmmeter used to calibrate the HOBO monitors in the field survey summarized herein are available upon request. The overall accuracy of the system for the temperature standard is about $\pm 0.05^{\circ}\text{F}$. The tolerance of the thermistors used for the WBN field survey is about $\pm 0.4^{\circ}\text{F}$, thus providing a calibration test accuracy ratio (TAR) of about 1:8. That is, the accuracy of temperature standard used for the sensor calibrations is about 8 times greater than the minimum acceptable field accuracy of temperature sensors. This is twice the recommended maximum TAR of 1:4 for sensor calibrations.

The TVA procedure to calibrate the HOBO water temperature monitors, Instruction No. 450.01-020, is provided below. Briefly, the HOBO monitors are immersed in a stirred temperature-

controlled water bath along with the standard (i.e., along with the Burns RTD probe). After the bath stabilizes, temperature readings from the HOBO monitors are compared to the temperature readings from the standard. Experience has shown that in nearly all cases, the readings from both the HOBO monitors and the standard are essentially constant, so that the 95 percent confidence interval of the readings is diminutive. Under these conditions, the accuracy of each HOBO monitor is recorded simply as the difference between the HOBO reading and that of the standard (negative difference = HOBO reading low/below standard, positive difference = HOBO reading high/above standard). The HOBO monitors are tested at three temperatures between 30°F and 100°F, covering the range of expected water temperature for natural river conditions. The three temperatures are at about the 10 percent, 50 percent, and 90 percent intervals, or 37°F, 65°F and 93°F, respectively. Any HOBO monitor with measured accuracy in excess of the maximum allowable tolerance of $\pm 0.4^\circ\text{F}$ for any one of the three temperatures fails the calibration test and is removed from the field survey inventory. The calibration certificates for HOBO monitors used in this field survey summarized herein are available upon request. All the HOBO monitors passed both the pre-survey and post-survey calibration tests. The mean square error of the HOBO monitors was 0.14°F for the pre-survey calibrations and 0.13°F for the post-survey calibrations.

| | | |
|---|---|---|
|  CENTRAL LABORATORIES SERVICES QUALITY PROGRAM INSTRUCTION | TITLE Certification of HOBO Water Temp Pro Data Acquisition Systems H ₂ O-001 | Instruction No. 450.01-020 Rev. No. 0 Page No. 1 of 7 Effective Date 5/19/03 |
| | | |
| LEVEL OF USE <input type="checkbox"/> Continuous <input checked="" type="checkbox"/> Reference <input type="checkbox"/> Information | | |
| <div style="text-align: right; font-weight: bold; font-size: 1.2em;">QA RECORD</div> <div style="display: flex; justify-content: space-between; margin-top: 20px;"> <div style="width: 45%;"> <p>_____</p> <p style="text-align: center;">Dennis T. Darby Preparer</p> </div> <div style="width: 45%;"> <p>_____</p> <p style="text-align: center;">5/19/03 Date</p> </div> </div> <div style="display: flex; justify-content: space-between; margin-top: 20px;"> <div style="width: 45%;"> <p>_____</p> <p style="text-align: center;">Paul B. Loiseau, Jr. Technical Reviewer</p> </div> <div style="width: 45%;"> <p>_____</p> <p style="text-align: center;">5/19/03 Date</p> </div> </div> <div style="display: flex; justify-content: space-between; margin-top: 20px;"> <div style="width: 45%;"> <p>_____</p> <p style="text-align: center;"> Administrative Review</p> </div> <div style="width: 45%;"> <p>_____</p> <p style="text-align: center;">6/5/03 Date</p> </div> </div> | | |
| APPROVAL | | |
| <div style="display: flex; justify-content: space-between; margin-top: 20px;"> <div style="width: 45%;"> <p>_____</p> <p style="text-align: center;">Jerry D. Hubble Department Manager</p> </div> <div style="width: 45%;"> <p>_____</p> <p style="text-align: center;">5/19/03 Date</p> </div> </div> | | |

| | |
|--|----------------------------|
| TITLE: Certification of HOBO Water Temp Pro Data Acquisition Systems H20-001 | Instruction No. 450.01-020 |
| | Rev. 0 |
| | Eff. Date 5/19/03 |
| | Page 3 of 7 |

1.0 PURPOSE

To provide uniform and effective certifications of Hobo Water Temp Pro data acquisition systems meeting the accuracy and performance requirements of TVA's water temperature-monitoring programs. This technical instruction uses the method of comparison with a laboratory standard thermometer.

2.0 SCOPE

This instruction applies to the certification of Hobo Water Temp Pro data loggers manufactured by Onset Computer Corporation of Bourne, Massachusetts. The Hobo Water Temp Pro is a data acquisition system containing a temperature sensor, data logger and battery sealed in a single submersible case. The Hobo Water Temp Pro is programmed and data retrieved by use of an infrared interface located in one end of the case. Hobo Water Temp Pros are certified upon receipt from the manufacturer at no greater than 12 month intervals during use or when requested.

3.0 SUMMARY

In this three-point certification systems are tested as actually used over the historical water temperature range of 30° to 100°F and submerged in water. The three test points are 37°, 65° and 93°F. The systems are required to perform within Onset Computer Corporation tolerances. System conformity at each temperature point is determined by comparing system temperature, logged by the Hobo Water Temp Pro and a laboratory standard thermometer.

Systems are programmed and submerged with a standard thermometer in a stirred, temperature-controlled temperature bath. The systems are read after the test by an infrared interface adapter connected to a computer running Onset Computer Corporation's Boxcar Pro software. Traceability of the certification is through the thermometer.

"As-found" certifications are performed on new systems as an acceptance test and on sensors returned from field service. "As-left" certifications are performed before delivery for field service if more than 12 months has elapsed since the last certification. "As-found" and "as-left" certifications may be combined on the same record if there is clear indication which type each system is undergoing.

Multiple HOBOs may be certified at the same time in the temperature bath.

| | |
|---|----------------------------|
| TITLE: Certification of HOBO Water Temp Pro Data Acquisition Systems H ₂ O-001 | Instruction No. 450.01-020 |
| | Rev. 0 |
| | Eff. Date 5/19/03 |
| | Page 4 of 7 |

- Accuracy of $\pm 0.2^{\circ}\text{C}$ at 25°C (0.33°F at 70°F)
- Waterproof case, submersible to 100 feet
- Capacity to store up to 21,580 temperature measurements
- Selectable sampling interval from 1 second to 9 hours
- Programmable start time/date
- Two data recording modes: Stop when full or wrap around when full.
- Two data offload modes: Halt then offload or offload while logging.
- Nonvolatile EEPROM memory that retains data even if batteries fail
- Light-emitting diode (LED) operation, indicator, which can be disabled during logging by selecting "Stealth" mode
- High-speed IR communications for offloading data; can readout full logger in less than 30 seconds while logging continues
- Battery life of 6 years with typical usage

4.0 PRACTICES/EXCEPTIONS

N/A

5.0 SAFETY

- 5.1 Standard electrical equipment safety.

6.0 STANDARDS USED

- 6.1 Laboratory reference thermometer, range 30° to 100°F or greater, 0.01°F resolution, 0.1°F accuracy or better, with current calibration sticker.

7.0 EQUIPMENT/APPARATUS

- 7.1 Temperature bath, stirred, temperature-controlled.
7.2 Computer with Onset Boxcar Pro software installed (version 4.3 or later)
7.3 IR Base station, Onset Part # BST-IR

8.0 PREREQUISITE ACTIONS

- 8.1 Turn on temperature bath and set for 37°F .
8.2 Check the IR interface to verify that it is plugged into the correct serial port on the PC. Set the correct time on the PC.
8.3 Align the IR port on the Base station with the HOBO Water Temp Pro communications window. Place the logger no further than 4 to 5 inches away from the Base station (see Figure 2) and make sure the IR windows in both devices point at each other. There is a 30° acceptance angle for the IR beam, so some misalignment is acceptable.

| | |
|--|-----------------------------------|
| TITLE: Certification of HOBO Water Temp Pro Data Acquisition Systems H₂O-001 | Instruction No. 450.01-020 |
| | Rev. 0 |
| | Eff. Date 5/19/03 |
| | Page 5 of 7 |

- 8.4 Start the Onset Box Car Software and select **Logger** then **Hobo Water Temp Pro** and **Launch**.
- 8.5 The computer will respond with a list of loggers found. The serial number in this list should match the serial number printed on the side of the logger. If these numbers do not match, click the **Refresh** button. Record this serial number on the certification form. Either wait or click the **Stop Searching** button. Using the mouse select the logger and click the **Launch** button.
- 8.6 After a few seconds the screen will display the status of the HOBO Water Temp Pro. Record the battery percentage on the certification form.
- 8.7 Verify that the Hobo is set to Fahrenheit and program it to a recording interval of 0:1:0 for a reading once a minute. Verify that the start logging immediately box is checked and that the set data logger clock with host launch is also checked.
- 8.8 Using the mouse click the **Launch Immediately** button.
- 8.9 If last HOBO is programmed click the **DONE** button, else select the **Launch Another** and repeat steps 8.5 through 8.9.
- 9.0 **TEST PROCEDURE/METHOD**
- 9.1 On the certification form record the serial number of the laboratory reference thermometer.
- 9.2 Place the HOBO Water Temp Pro in the temperature bath, making sure the end opposite the IR windows is submerged, and allow the bath to stabilize at 37°F ± 0.5°F on the thermometer. Adjust the bath set point if needed. After the bath reaches the desired temperature allow 20 minutes 'soak time' for the HOBO to reach its final temperature.
- 9.3 Record the thermometer reading on the certification form and the time. (The time will be needed to get the correct reading from the HOBO.)
- 9.4 Repeat steps 9.2 and 9.3 for bath settings of 65.0°F ± 0.5°F and 93°F ± 0.5°F.
- 9.5 Remove the HOBO from the temperature bath and align the IR port on the Base station with the HOBO Water Temp Pro communications window.
- 9.6 Restart Onset BoxCar Pro if it is not running and select **Logger** then **Hobo Water Temp Pro** and **Readout**.
- 9.7 The computer will respond with a list of loggers found. Using the mouse select the logger and click the **Readout** button. The computer will ask to download data and continue logging or the stop logging and offload data. Select the **Stop Logging and Offload data**. After a few seconds the computer will respond with a suggested file name. Select **Save** and allow the HOBO to transfer the data.
- 9.8 After a successful download click the **OK** button. The computer will then ask if the data should be displayed in Centigrade or Fahrenheit. Deselect °C and select °F and click **OK**. The computer should display a graph of the collected data. Click the view details button (this is the button just left of the question mark button.)

| | |
|--|-----------------------------------|
| TITLE: Certification of HOBO Water Temp Pro Data Acquisition Systems H₂O-001 | Instruction No. 450.01-020 |
| | Rev. 0 |
| | Eff. Date 5/19/03 |
| | Page 6 of 7 |

9.9 Scroll down the displayed list until the time recorded for the 37°F point is found. Record the corresponding temperature on the certification form. Repeat this step for 65° and 93°.

9.10 Close the view details windows and repeat steps 9.6 through 9.9 for additional HOBOS.

9.11 Fill out the rest of the certification form.

10.0 ACCEPTANCE CRITERIA

10.1 Based upon the manufacturer specifications the HOBO Water Temp Pro should be within $\pm 0.4^{\circ}\text{F}$ over the range of 32°F to 100°F. Any HOBO with an error of greater than $\pm 0.5^{\circ}\text{F}$ at any of the three measured points shall fail certification.

11.0 POST PROCEDURE ACTIVITY

11.1 Close the BoxCar Software.

12.0 RECORDS

12.1 Completed HOBO Water Temperature Pro Certification form and associated Report of Certification cover sheet is a QA record.

13.0 REFERENCE

13.1 HOBO Water Temp Pro User's Manual, version 1.0 or later

13.2 Onset BoxCar Pro4 Manual Version 1.0 or later

APPENDIX B

WBN Outfall 113 NPDES Compliance Parameters

- Current Instantaneous Upstream Temperature:

Tu_i (measured at EDS Station 30 by the first sensor below a depth of 5 feet).

- Current 1-Hour Average Upstream Temperature:

$$Tu1_i = \frac{Tu_i + Tu_{i-1} + Tu_{i-2} + Tu_{i-3} + Tu_{i-4}}{5},$$

where the subscripts i, i-1, i-2, i-3, and i-4 denote the current and previous four 15-minute (0.25 hour) values of Tu .

- Current Instantaneous Downstream Temperature:

$$Td_i = \frac{Td3_i + Td5_i + Td7_i}{3},$$

where $Td3_i$, $Td5_i$, and $Td7_i$ denote the current measurements of river temperature at the downstream end of the mixing zone at water depths 3 feet, 5 feet, and 7 feet, respectively.

- Current 1-Hour Average Downstream Temperature:

$$Td1_i = \frac{Td_i + Td_{i-1} + Td_{i-2} + Td_{i-3} + Td_{i-4}}{5},$$

where the subscripts i, i-1, i-2, i-3, and i-4 denote the current and previous four 15-minute (0.25 hour) values of Td .

- Current Instantaneous Temperature Rise:

$$\Delta_i = Td_i - Tu_i.$$

- Current 1-Hour Average Temperature Rise:

$$\Delta1_i = \frac{\Delta_i + \Delta_{i-1} + \Delta_{i-2} + \Delta_{i-3} + \Delta_{i-4}}{5},$$

where the subscripts i, i-1, i-2, i-3, and i-4 denote the current and previous four 15-minute (0.25 hour) values of ΔT .

- Current Temperature Rate-of-Change:

$$TROC_i = \frac{Td_i - Td_{i-1}}{1 \text{ hour}} \quad .$$

- Current 1-Hour Average Temperature Rate-of-Change:

$$TROC_{1i} = \frac{TROC_i + TROC_{i-1} + TROC_{i-2} + TROC_{i-3} + TROC_{i-4}}{5} \quad ,$$

where the subscripts i, i-1, i-2, i-3, and i-4 denote the current and previous four 15-minute (0.25 hour) values of TROC.

Enclosure 6

**Winter 2012 Compliance Survey for Watts Bar Nuclear Plant Outfall Passive
Mixing Zone**

TENNESSEE VALLEY AUTHORITY
River Operations

**WINTER 2012 COMPLIANCE SURVEY FOR WATTS BAR
NUCLEAR PLANT OUTFALL 113 PASSIVE MIXING ZONE**

Prepared by

Daniel P. Saint
and
Paul N. Hopping

Knoxville, Tennessee
October 2012



EXECUTIVE SUMMARY

The National Pollutant Discharge Elimination System (NPDES) Permit No. TN0020168 for Watts Bar Nuclear Plant (WBN) identifies the discharge of water to the Tennessee River from the Supplemental Condenser Cooling Water (SCCW) System as Outfall 113. Furthermore, the permit identifies that when there is no flow released from Watts Bar Dam (WBH), the effluent from Outfall 113 shall be regulated based on a passive mixing zone extending in the river from bank-to-bank and 1,000 feet downstream from the outfall. Compliance with the requirements for the passive mixing zone is to be achieved by two annual instream temperature surveys—one for winter conditions and one for summer conditions. Summarized in this report are the measurements, analyses, and results for the passive mixing zone survey performed for 2012 winter conditions. The survey was conducted between 21:00 CDT on February 21 and 05:00 CDT on February 22 (eight hours) and included the collection of temperature data at twelve temporary monitoring stations deployed across the downstream end of the passive mixing zone during a period of no flow in the river. The data were analyzed to determine the three instream compliance parameters specified in the NPDES permit for the outfall: the 1-hour average temperature at the downstream end of mixing zone, T_d ; the 1-hour average temperature rise from upstream to the downstream end of the mixing zone, ΔT ; and the 1-hour average temperature rate-of-change at the downstream end of the mixing zone, TROC. The measured parameters were compared to predicted values from the thermal plume model used by TVA to help determine the safe operation of Outfall 113. The results of the comparisons, in terms of maximum values observed during the no flow event, are as follows:

| Compliance Parameter | Model | Measured | NPDES Limit |
|----------------------|-------------|-------------|-------------|
| Maximum T_d | 51.5°F | 49.9°F | 86.9°F |
| Maximum ΔT | 4.4 F° | 3.2 F° | 5.4 F° |
| Maximum TROC | 1.2 F°/hour | 0.8 F°/hour | 3.6 F°/hr |

As shown, both the model and measured values were well below the NPDES limits for all the compliance parameters. Based on the results, the thermal plume model is considered adequate for determining the safe operation of the SCCW system. That is, in combination with TVA procedures for predicting the impact of the Outfall 113 effluent, the model continues to provide a high level of confidence that the SCCW system is being operated in a manner that does not exceed the instream limits for T_d , ΔT , and TROC as specified in the WBN NPDES permit.

TABLE OF CONTENTS

| | <u>Page No.</u> |
|---|-----------------|
| EXECUTIVE SUMMARY | i |
| INTRODUCTION | 1 |
| INSTREAM SURVEY | 2 |
| RESULTS | 3 |
| River Conditions | 3 |
| SCCW Conditions..... | 3 |
| Downstream End of Passive Mixing Zone | 4 |
| NPDES Compliance Parameters..... | 5 |
| CONCLUSIONS..... | 6 |
| REFERENCES | 8 |
| APPENDIX A..... | 19 |
| APPENDIX B | 27 |

LIST OF FIGURES

| | |
|---|----|
| Figure 1. Watts Bar Nuclear Plant Outfall 113 (SCCW) Mixing Zones | 9 |
| Figure 2. Location of HOBO Monitoring Stations | 10 |
| Figure 3. Schematic of HOBO Water Temperature Monitoring Stations | 10 |
| Figure 4. River Conditions..... | 11 |
| Figure 5. SCCW Conditions | 12 |
| Figure 6. HOBO Water Temperature Measurements | 13 |
| Figure 7. Local Instantaneous Temperature Rise for HOBO Measurements | 15 |
| Figure 8. Measured and Computed Compliance Parameters for Passive Mixing Zone | 18 |

LIST OF TABLES

| | |
|--|---|
| Table 1. NPDES Temperature Limits for Outfall 113 Mixing Zones | 1 |
| Table 2. Sources of Data for Passive Mixing Zone Survey | 2 |

WINTER 2012* COMPLIANCE SURVEY FOR WATTS BAR NUCLEAR PLANT OUTFALL 113 PASSIVE MIXING ZONE

INTRODUCTION

Outfall 113 for the Watts Bar Nuclear Plant (WBN) includes the discharge of water to the Tennessee River from the Supplemental Condenser Cooling Water (SCCW) system. Due to the dynamic behavior of the thermal effluent in the river, the National Pollutant Discharge Elimination System (NPDES) Permit No. TN0020168 for the plant specifies two mixing zones for Outfall 113—one for active operation of the river and one for passive operation of the river (TDEC, 2010). The passive mixing zone corresponds to periods when the operation of Watts Bar Dam (WBH) produces no flow in the river (i.e., hydropower and/or spillway releases). The dimensions of the passive mixing zone extend from bank-to-bank and downstream 1,000 feet from the outfall. The active mixing zone applies to all other river flow conditions. The dimensions of the active mixing zone include the right-half of the river (facing downstream) and extend downstream 2,000 feet from the outfall. The passive and the active mixing zones are shown in Figure 1.

Table 1 summarizes the NPDES instream temperature limits for Outfall 113. The limits apply to both the active and passive mixing zones. Compliance for the active mixing zone is monitored by permanent instream water temperature stations situated in the right-half of the river. Due to issues associated with placing permanent stations in the left-half of the river, which contains the navigation channel, a thermal plume model is used to determine the safe operation of Outfall 113 for the passive mixing zone. To verify the thermal plume model, the NPDES permit specifies that two instream temperature surveys shall be conducted each year—one for winter conditions and one for summer conditions. The purpose of this report is to present the results for the passive mixing zone temperature survey performed for winter 2012 conditions. The survey was conducted between 21:00 CDT on February 21 and 05:00 CDT on February 22 (total eight hours). Provided is a brief summary of the survey method, presentations of the measurements and analyses, and discussions of the results and conclusions.

Table 1. NPDES Temperature Limits for Outfall 113 Mixing Zones

| Compliance Parameter | Sampling Period | NPDES Limit |
|--|------------------------|--------------------|
| Maximum Temperature, Downstream End of Mixing Zone, T_d | Running 1-hr | 86.9°F |
| Maximum Temperature Rise, Upstream to Downstream, ΔT | Running 1-hr | 5.4 F° |
| Maximum Temperature Rate-of-Change, TROC | Running 1-hr | ±3.6 F°/hr |

* R1: Title correction from initial release (initial release contained “2011” rather than “2012”).

INSTREAM SURVEY

The instream survey included the deployment of temporary water temperature stations at twelve locations across the downstream end of the passive mixing zone. Data from these and other monitoring stations were analyzed to obtain measured values for the compliance parameters listed in Table 1. These were then compared with the corresponding values estimated from the SCCW thermal plume model.

The method of conducting the instream survey is the same as that used for the first such survey, performed for winter conditions on May 6, 2005 (McCall and Hopping, 2005). Table 2 provides a summary of the sources of data for the survey. WaterView, a monitoring system for tracking hydroplant operation and performance, was used to obtain measurements for the river discharge from Watts Bar Dam. The WBN Environmental Data Station (EDS) provided measurements from existing permanent monitoring stations for the nuclear plant. These included:

- The river upstream (ambient) water temperature, measured at the EDS Station 30, which is located at the exit of the powerhouse of Watts Bar Dam.
- The river water surface elevation (WSEL) at the EDS Station 30, also known as the tailwater elevation (TWEL) at Watts Bar Dam.
- The SCCW effluent temperature, measured at the EDS Station 32, which is located at the SCCW outfall.
- The SCCW effluent discharge, measured at the EDS Station 32.
- The local air temperature, measured at the EDS meteorological tower.

Table 2. Sources of Data for Passive Mixing Zone Survey

| Data | Source | Frequency |
|---------------------------------------|---------------------------------------|-----------|
| River Discharge from Watts Bar Dam | WaterView | 1 min |
| River ambient water temperature | WBN EDS Station 30 (Tailwater at WBH) | 15 min |
| River water surface elevation | WBN EDS Station 30 (Tailwater at WBH) | 15 min |
| SCCW effluent temperature | WBN EDS Station 32 (SCCW Outfall 113) | 15 min |
| SCCW effluent discharge | WBN EDS Station 32 (SCCW Outfall 113) | 15 min |
| Air temperature | WBN EDS Met Tower | 15 min |
| Passive mixing zone water temperature | Temporary HOBO Monitors | 1 min |

The water temperature at the downstream end of the Outfall 113 passive mixing zone was measured by the aforementioned temporary water temperature stations. Using a global positioning system (GPS) device, the stations were positioned at roughly equal intervals across the river, as shown in Figure 2. The temporary stations recorded water temperatures by using HOBO temperature monitors positioned at depths of 0.5, 3, 5, and 7 feet below the water surface. Shown in Figure 3 is a schematic of the temporary stations. The stations included a string of

HOBO monitors suspended from a tire float, with weights to anchor the station and to keep the sensor string vertical in the water column. The water temperature sensors imbedded in the HOBO monitors have an accuracy of about ± 0.4 F° and resolution of about 0.04 F°, which is comparable to the accuracy and resolution of temperature sensors used elsewhere by TVA for NPDES thermal compliance. The HOBO monitors include an internal data acquisition unit that was programmed to collect measurements once per minute. All the temperature probes used in the survey, including both those contained in the HOBO monitors and the thermistors at the permanent EDS monitoring stations, were calibrated by a quality program with equipment accuracies traceable to the National Institute of Standards and Technology (NIST). The calibration procedure is summarized in APPENDIX A. The temporary monitoring stations were deployed several hours before the beginning of the survey, and retrieved several hours after the end of the survey.

RESULTS

River Conditions

Figure 4 shows the measured ambient conditions of the river during the survey. Included are the river discharge, the river tailwater elevation, and river temperature at the exit of Watts Bar Dam. The river temperature at the exit of Watts Bar Dam serves as the upstream ambient river temperature for WBN Outfall 113. To provide a period of no flow in the river, releases from Watts Bar Dam were suspended between about 21:00 CDT on February 21 and 05:00 CDT on February 22, a total of eight hours (nighttime). Leading up to the survey, as the river flow was stepping down, the WSEL below Watts Bar Dam dropped approximately 2.8 feet, from about 679.8 feet msl to about 677.0 feet msl. For the first 5 hours of the survey, the tailwater elevation remained steady at about $677.0 \pm$ feet msl and then for the next three hours, the tailwater elevation slowly receded, reaching about 676.4 feet msl by the end of the survey.

The ambient river temperature was 46.7°F at the beginning of the period of no flow, and remained at this temperature throughout the duration of the study. This behavior is common during the winter months when the water column behind Watts Bar Dam contains little or no stratification. Under these conditions, whether the withdrawal zone from the reservoir is large (e.g., high turbine release) or small (e.g., no flow leakage), the ambient river temperature below the dam remains essentially constant.

SCCW Conditions

During the survey, the SCCW system at WBN was thermally loaded and operating in “summer” mode. That is, the system was operating in a manner producing the largest possible release of heat to the river. Shown in Figure 5 are the measured conditions of the SCCW system during the

survey. Included are the discharge and temperature of the SCCW effluent. During the survey, the average discharge of the SCCW system to the river was about 207 cfs. The root-mean-square variation in the SCCW discharge was only about 2.8 percent of the average—thus, from the standpoint of mixing processes in the river, the discharge was essentially constant. The SCCW effluent temperature decreased throughout the survey from about 70.8°F at the beginning of the survey to about 66.0°F at the end of the survey. This trend coincides with the falling nighttime air temperature, also shown in Figure 5 (note: the discharge temperature of water from the Unit 1 cooling tower, which provides the source of heat for Outfall 113, varies directly with the temperature of the ambient air that is drawn through the tower). Relative to the upstream ambient river temperature, the temperature rise of the Outfall 113 effluent released from the SCCW system, also shown in Figure 5, decreased from about 24.1 F° at the beginning of the survey to about 19.3 F° at the end of the survey.

Downstream End of Passive Mixing Zone

Shown in Figure 6 are the measurements from the HOBO temperature stations at the downstream end of the passive mixing zone. The stations are labeled consecutively from WB1 to WB12, with WB1 situated near the left-hand shoreline of the river and WB12 situated near the right-hand shoreline of the river (i.e., facing downstream—see Figure 2). In Figure 7, the HOBO data has been analyzed to produce contour plots of the local “instantaneous” water temperature rise (ΔT) relative to the SCCW ambient river temperature (i.e., given in Figure 4). The horizontal (x) axis of each contour plot is the span of the river from WB1 to WB12, and the vertical (y) axis is the water depth from 0.5 feet to 7 feet. In this manner, the plots in Figure 7 represent images of the upper 7 feet of the water column in the river, looking downstream. Note that the depth scale in the plots is very distorted so that the data can be viewed in a meaningful manner—that is, whereas the span of the x-axis is about 1000 feet, the span of the y-axis is only about 7 feet (0.007 times smaller). Plots are provided at the top of each hour from the beginning of the survey at 21:00 CDT on February 21 to the end of the survey at 05:00 CDT on February 22. The following behaviors are emphasized from Figure 6 and Figure 7:

- At the beginning of the survey, 21:00 CDT on February 21, effluent from the SCCW resides primarily on the right-hand-side of the river. This is due to the flow in the river preventing the effluent from spreading across the river. The maximum local instantaneous temperature rise at the downstream end of the passive mixing zone is about 4.8 F° and occurs in the upper 3 feet of the water column in the very right-hand-side of the river.
- Over the next two hours, the effluent from the SCCW slowly spreads across the passive mixing zone. Since there is no flow in the river, the SCCW effluent is somewhat unrestricted, reaching the left-hand-side of the river and spreading downstream alongside the

shoreline. The maximum local instantaneous temperature rise during this period is about 2.8 F° and occurs at 22:00 CDT near the middle of the river.

- By 01:00 CDT on February 22, four hours into the survey, heat from the SCCW effluent is distributed fully across the downstream end of the passive mixing zone. The maximum local instantaneous temperature rise at this point in time is about 4.0 F° and again occurs near the middle of the river.
- Throughout the remaining hours of the survey, the SCCW effluent slowly accumulates across the mixing zone. Due to buoyancy, the heat resides primarily in the upper 3 feet of the water column, with the local instantaneous temperature rise reaching, at places, around 4 F°. Between the depths of 3 feet and 7 feet, a local instantaneous temperature rise in the vicinity of 3 F° is more common.

NPDES Compliance Parameters

Since heat from the SCCW effluent is distributed across the full width of the river, data from all of the HOBO stations were used to compute the NPDES compliance parameters, which is consistent with the dimensions of the passive mixing zone (i.e., the passive mixing zone spans the full width of the river). The compliance parameters examined include all those given in Table 1—the temperature at the downstream end of mixing zone, T_d ; the temperature rise from upstream to the downstream end of the mixing zone, ΔT ; and the temperature rate-of-change at the downstream end of the mixing zone, TROC. The fundamental equations used to compute the compliance parameters are provided in APPENDIX B, based on the criteria specified in the NPDES permit. The temperature at the downstream end of the mixing zone was determined from the HOBO measurements by averaging the readings from the sensors at depths 3, 5, and 7 feet for all twelve HOBO stations. The temperature rise was computed as the difference between the measured temperature at the downstream end of the mixing zone and the upstream temperature measured at Watts Bar Dam (i.e., Station 30). The temperature rate-of-change was determined by the change in the measured temperature at the downstream end of the mixing zone from one hour to the next. The data were averaged over a period of one hour using 15-minute readings, as specified in the NPDES permit, and compared with the WBN thermal plume model. The measurements are presented in Figure 8, along with the results obtained by the thermal plume model. The following behaviors are emphasized:

- Temperature at the downstream end of the passive mixing zone, T_d : The maximum 1-hour average T_d estimated by the thermal plume model was 51.5°F, whereas the maximum measured value was about 49.9°F. Thus, the model overpredicted the maximum measured T_d by 0.6°F. Compared to the measurements, the increase in river temperature due to the no flow event was predicted to occur much more rapidly by the model. This is because the

model assumes impacts due to changes in the river and/or Outfall 113 conditions are fully realized as a steady-state episode within one hour (i.e., the model time-step); whereas in reality, the actual time for the thermal plume to evolve is much longer. Both the predictions from the model and measurements from the survey were well below the NPDES limit of 86.9°F.

- Temperature rise, ΔT : The maximum 1-hour average ΔT predicted by the plume model was 4.4 F°, whereas the maximum measured value was about 3.2 F°. Thus, the model overpredicted the maximum measured temperature rise by 1.2 F°. For the reason cited above (i.e., computational time-step of one hour), the model predicted the maximum temperature rise to occur one hour into the no flow event. Both the predictions from the model and measurements from the survey were well below the NPDES limit of 5.4 F°.
- Temperature rate-of-change, TROC: The maximum 1-hour average TROC predicted by the plume model was 1.2 F°/hour, whereas the maximum measured value was about 0.8 F°/hour (absolute values). Thus, the model overpredicted the temperature rate-of-change by 0.4 F°/hour. Both the predictions from the model and measurements from the survey were well below the NPDES limit of ± 3.6 F°/hour.

CONCLUSIONS

The compliance survey for 2012 winter conditions was successful in measuring the NPDES instream water temperature parameters for the Outfall 113. These included the temperature, T_d , temperature rise, ΔT , and temperature rate-of-change, TROC, all at the downstream end of the passive mixing zone. The measurements were compared with values predicted by the thermal plume model that TVA currently uses to determine the safe operation of the SCCW system. For the results summarized herein, the measured values for each of these parameters were bounded by the model values. That is, the model values were greater than or equal to the actual measured values, assuring compliance with the instream standards for water temperature. Since 2005, when the first compliance survey was performed for the Outfall 113 passive mixing zone, the model value for the maximum downstream temperature T_d has always bounded the measured value for the maximum T_d . The same is not true, however, for the maximum temperature rise ΔT and the maximum temperature rate-of-change TROC. In the summer survey for 2011, the model value for the maximum ΔT underpredicted the measured value for the maximum ΔT by 0.1 F° (Saint and Hopping, 2011), and in the summer survey for 2005, the model value for the maximum TROC underpredicted the measured value for the maximum TROC by 0.3 F°/hour (McCall and Hopping, 2006). These differences are not surprising in light of the fact that the model, like any mathematical representation of an actual complex physical process, contains inherent accuracy limitations.

The TVA model for predicting the Outfall 113 thermal plume uses CORMIX, which has a stated accuracy of about 50% of the standard deviation of field measurements (Jirka, et al., 1996). Based on this, as well as the fact that differences as small as 0.1 F° for ΔT and 0.3 F°/hour for TROC fall within the factor of safety currently used in performing hydrothermal forecasts, the thermal plume model is still considered adequate for determining the safe operation of the SCCW system. That is, in combination with TVA procedures for predicting the impact of the Outfall 113 effluent, the model continues to provide a high level of confidence that the SCCW system is being operated in a manner that does not exceed the instream limits for T_d , ΔT , and TROC as specified in the WBN NPDES permit for the passive mixing zone.

REFERENCES

Harper, Walter L., and Bo Hadjerioua, Mark Reeves, Gary Hickman, and John Jenkinson, "Hydrodynamics and Water Temperature Modeling at Watts Bar SCCW Discharge Structure," TVA Resource Group, Water Management, Report No. WR98-1-85-142, November 1998.

Jirka, Gerhard H., Robert L. Doneker, and Steven W. Hinton, "User's Manual for CORMIX: A Hydrodynamic Mixing Zone Model and Decision Support System for Pollutant Discharges into Surface Waters," Office of Science and Technology, U.S. Environmental Protection Agency, Washington, DC, September 1996.

McCall, Michael J., and P.N. Hopping, "Summer 2005 Compliance Survey for Watts Bar Nuclear Plant Outfall 113 Passive Mixing Zone," TVA River Operations, Report No. WR2006-2-85-152, February 2006.

McCall, Michael J., and P.N. Hopping, "Winter 2005 Compliance Survey for Watts Bar Nuclear Plant Outfall 113 Passive Mixing Zone," TVA River Operations, Report No. WR2005-2-85-151, October 2005.

Saint, Daniel P., and P.N. Hopping, "Summer 2011 Compliance Survey for Watts Bar Nuclear Plant Outfall 113 Passive Mixing Zone," TVA River Operations, March 2012.

TDEC, *State of Tennessee NPDES Permit No. TN0020168*, Tennessee Department of Environment and Conservation, Issued June 2010.




Figure 1. Watts Bar Nuclear Plant Outfall 113 (SCCW) Mixing Zones




Figure 2. Location of HOBO Monitoring Stations




Figure 3. Schematic of HOBO Water Temperature Monitoring Stations




Figure 4. River Conditions




Figure 5. SCCW Conditions




Figure 6. HOBO Water Temperature Measurements




Figure 6 (Continued). HOBO Water Temperature Measurements




Figure 7. Local Instantaneous Temperature Rise for HOB0 Measurements




Figure 7 (Continued). Local Instantaneous Temperature Rise for HOBO Measurements




Figure 7 (Continued). Local Instantaneous Temperature Rise for HOB0 Measurements




Figure 8. Measured and Computed Compliance Parameters for Passive Mixing Zone

APPENDIX A

Calibration of NPDES Water Temperature Sensors



All sensors used by TVA for monitoring compliance of NPDES water temperature requirements are certified and maintained to meet the following industry and regulatory standards:

- ISO/IEC 17025—Quality assurance requirements for the competence to carry out sampling, testing, and calibrations using standard, non-standard, and laboratory-developed methods (ISO=International Organization for Standardization, IEC=International Electrotechnical Commission).
- 10CFR50 Appendix B—Quality assurance criteria for design, fabrication, construction, and testing of the structures, systems, and components of nuclear power plants (CFR=Code of Federal Regulations).
- 40CFR136—Guidelines establishing test procedures for the analysis of pollutants under the Clean Water Act.
- ANSI N45.2. 1971—Quality assurance requirements for Nuclear Power Plants (ANSI=American National Standards Institute).
- ANSI/NCSL Z540-1-1994—General requirements for calibration laboratories and equipment used for measurements and testing (NCSL=National Conference of Standards Laboratories).

The standard used to certify the thermistors for the permanent EDS stations and the temporary HOBO stations is traceable to the National Institute of Standards and Technology (NIST). The standard includes two pieces of equipment—a platinum resistance temperature detector (RTD) manufactured by Burns Engineering, Inc. and an ohmmeter manufactured by Azonix Inc. The latter is used to measure the resistance of the RTD (i.e., the resistance of platinum varies with temperature). The NTIS traceable calibration certificates for the Burns RTD and the Azonix ohmmeter used to calibrate the HOBO monitors in the field survey summarized herein are available upon request. The overall accuracy of the system for the temperature standard is about $\pm 0.05^{\circ}\text{F}$. The tolerance of the thermistors used for the WBN field survey is about $\pm 0.4^{\circ}\text{F}$, thus providing a calibration test accuracy ratio (TAR) of about 1:8. That is, the accuracy of temperature standard used for the sensor calibrations is about 8 times greater than the minimum acceptable field accuracy of temperature sensors. This is twice the recommended maximum TAR of 1:4 for sensor calibrations.

The TVA procedure to calibrate the HOBO water temperature monitors, Instruction No. 450.01-020, is provided below. Briefly, the HOBO monitors are immersed in a stirred temperature-

controlled water bath along with the standard (i.e., along with the Burns RTD probe). After the bath stabilizes, temperature readings from the HOBO monitors are compared to the temperature readings from the standard. Experience has shown that in nearly all cases, the readings from both the HOBO monitors and the standard are essentially constant, so that the 95 percent confidence interval of the readings is diminutive. Under these conditions, the accuracy of each HOBO monitor is recorded simply as the difference between the HOBO reading and that of the standard (negative difference = HOBO reading low/below standard, positive difference = HOBO reading high/above standard). The HOBO monitors are tested at three temperatures between 30°F and 100°F, covering the range of expected water temperature for natural river conditions. The three temperatures are at about the 10 percent, 50 percent, and 90 percent intervals, or 37°F, 65°F and 93°F, respectively. Any HOBO monitor with measured accuracy in excess of the maximum allowable tolerance of $\pm 0.4^{\circ}\text{F}$ for any one of the three temperatures fails the calibration test and is removed from the field survey inventory. The calibration certificates for HOBO monitors used in this field survey summarized herein are available upon request. All the HOBO monitors passed both the pre-survey and post-survey calibration tests. The mean square error of the HOBO monitors was 0.14°F for the pre-survey calibrations and 0.13°F for the post-survey calibrations.

| | | |
|---|---|--|
|  CENTRAL LABORATORIES SERVICES QUALITY PROGRAM INSTRUCTION | TITLE Certification of HOBO Water Temp Pro Data Acquisition Systems H ₂ O-001 | Instruction No. 450.01-020 Rev. No. 0 Page No. 1 of 7 |
| | | Effective Date 5/19/03 |
| LEVEL OF USE <input type="checkbox"/> Continuous <input checked="" type="checkbox"/> Reference <input type="checkbox"/> Information | | |
| QA RECORD | | |
| _____ Dennis T. Darby Preparer | | _____ 5/19/03 Date |
| _____ Paul B. Loiseau, Jr. Technical Reviewer | | _____ 5/19/03 Date |
| _____  Administrative Review | | _____ 6/5/03 Date |
| APPROVAL | | |
| _____ Jerry D. Hubble Department Manager | | _____ 5/19/03 Date |

| | |
|--|----------------------------|
| TITLE: Certification of HOB0 Water Temp Pro Data Acquisition Systems H20-001 | Instruction No. 450.01-020 |
| | Rev. 0 |
| | Eff. Date 5/19/03 |
| | Page 3 of 7 |

1.0 PURPOSE

To provide uniform and effective certifications of Hobo Water Temp Pro data acquisition systems meeting the accuracy and performance requirements of TVA's water temperature-monitoring programs. This technical instruction uses the method of comparison with a laboratory standard thermometer.

2.0 SCOPE

This instruction applies to the certification of Hobo Water Temp Pro data loggers manufactured by Onset Computer Corporation of Bourne, Massachusetts. The Hobo Water Temp Pro is a data acquisition system containing a temperature sensor, data logger and battery sealed in a single submersible case. The Hobo Water Temp Pro is programmed and data retrieved by use of an infrared interface located in one end of the case. Hobo Water Temp Pros are certified upon receipt from the manufacturer at no greater than 12 month intervals during use or when requested.

3.0 SUMMARY

In this three-point certification systems are tested as actually used over the historical water temperature range of 30° to 100°F and submerged in water. The three test points are 37°, 65° and 93°F. The systems are required to perform within Onset Computer Corporation tolerances. System conformity at each temperature point is determined by comparing system temperature, logged by the Hobo Water Temp Pro and a laboratory standard thermometer.

Systems are programmed and submerged with a standard thermometer in a stirred, temperature-controlled temperature bath. The systems are read after the test by an infrared interface adapter connected to a computer running Onset Computer Corporation's Boxcar Pro software. Traceability of the certification is through the thermometer.

"As-found" certifications are performed on new systems as an acceptance test and on sensors returned from field service. "As-left" certifications are performed before delivery for field service if more than 12 months has elapsed since the last certification. "As-found" and "as-left" certifications may be combined on the same record if there is clear indication which type each system is undergoing.

Multiple HOB0s may be certified at the same time in the temperature bath.

| | |
|--|-----------------------------------|
| TITLE: Certification of HOBO Water Temp Pro Data Acquisition Systems H₂O-001 | Instruction No. 450.01-020 |
| | Rev. 0 |
| | Eff. Date 5/19/03 |
| | Page 4 of 7 |

- Accuracy of $\pm 0.2^{\circ}\text{C}$ at 25°C (0.33°F at 70°F)
- Waterproof case, submersible to 100 feet
- Capacity to store up to 21,580 temperature measurements
- Selectable sampling interval from 1 second to 9 hours
- Programmable start time/date
- Two data recording modes: Stop when full or wrap around when full.
- Two data offload modes: Halt then offload or offload while logging.
- Nonvolatile EEPROM memory that retains data even if batteries fail
- Light-emitting diode (LED) operation, indicator, which can be disabled during logging by selecting "Stealth" mode
- High-speed IR communications for offloading data; can readout full logger in less than 30 seconds while logging continues
- Battery life of 6 years with typical usage

4.0 PRACTICES/EXCEPTIONS

N/A

5.0 SAFETY

- 5.1 Standard electrical equipment safety.

6.0 STANDARDS USED

- 6.1 Laboratory reference thermometer, range 30° to 100°F or greater, 0.01°F resolution, 0.1°F accuracy or better, with current calibration sticker.

7.0 EQUIPMENT/APPARATUS

- 7.1 Temperature bath, stirred, temperature-controlled.
7.2 Computer with Onset Boxcar Pro software installed (version 4.3 or later)
7.3 IR Base station, Onset Part # BST-IR

8.0 PREREQUISITE ACTIONS

- 8.1 Turn on temperature bath and set for 37°F .
8.2 Check the IR interface to verify that it is plugged into the correct serial port on the PC. Set the correct time on the PC.
8.3 Align the IR port on the Base station with the HOBO Water Temp Pro communications window. Place the logger no further than 4 to 5 inches away from the Base station (see Figure 2) and make sure the IR windows in both devices point at each other. There is a 30° acceptance angle for the IR beam, so some misalignment is acceptable.

| | |
|---|----------------------------|
| TITLE: Certification of HOBO Water Temp Pro Data Acquisition Systems H ₂ O-001 | Instruction No. 450.01-020 |
| | Rev. 0 |
| | Eff. Date 5/19/03 |
| | Page 5 of 7 |

- 8.4 Start the Onset Box Car Software and select **Logger** then **Hobo Water Temp Pro** and **Launch**.
- 8.5 The computer will respond with a list of loggers found. The serial number in this list should match the serial number printed on the side of the logger. If these numbers do not match, click the **Refresh** button. Record this serial number on the certification form. Either wait or click the **Stop Searching** button. Using the mouse select the logger and click the **Launch** button.
- 8.6 After a few seconds the screen will display the status of the HOBO Water Temp Pro. Record the battery percentage on the certification form.
- 8.7 Verify that the Hobo is set to Fahrenheit and program it to a recording interval of 0:1:0 for a reading once a minute. Verify that the start logging (immediately) box is checked and that the set data logger clock with host launch is also checked.
- 8.8 Using the mouse click the **Launch Immediately** button.
- 8.9 If last HOBO is programmed click the **DONE** button, else select the **Launch Another** and repeat steps 8.5 through 8.9.
- 9.0 **TEST PROCEDURE/METHOD**
- 9.1 On the certification form record the serial number of the laboratory reference thermometer.
- 9.2 Place the HOBO Water Temp Pro in the temperature bath, making sure the end opposite the IR windows is submerged, and allow the bath to stabilize at 37°F ±0.5°F on the thermometer. Adjust the bath set point if needed. After the bath reaches the desired temperature allow 20 minutes 'soak time' for the HOBO to reach its final temperature.
- 9.3 Record the thermometer reading on the certification form and the time. (The time will be needed to get the correct reading from the HOBO.)
- 9.4 Repeat steps 9.2 and 9.3 for bath settings of 65.0°F ± 0.5°F and 93°F ± 0.5°F.
- 9.5 Remove the HOBO from the temperature bath and align the IR port on the Base station with the HOBO Water Temp Pro communications window.
- 9.6 Restart Onset BoxCar Pro if it is not running and select **Logger** then **Hobo Water Temp Pro** and **Readout**.
- 9.7 The computer will respond with a list of loggers found. Using the mouse select the logger and click the **Readout** button. The computer will ask to download data and continue logging or the stop logging and offload data. Select the **Stop Logging and Offload data**. After a few seconds the computer will respond with a suggested file name. Select **Save** and allow the HOBO to transfer the data.
- 9.8 After a successful download click the **OK** button. The computer will then ask if the data should be displayed in Centigrade or Fahrenheit. Deselect °C and select °F and click **OK**. The computer should display a graph of the collected data. Click the view details button (this is the button just left of the question mark button.)

| | |
|---|-----------------------------------|
| TITLE: Certification of HOBO Water Temp Pro Data Acquisition Systems H20-001 | Instruction No. 450.01-020 |
| | Rev. 0 |
| | Eff. Date 5/19/03 |
| | Page 6 of 7 |

9.9 Scroll down the displayed list until the time recorded for the 37°F point is found. Record the corresponding temperature on the certification form. Repeat this step for 65° and 93°.

9.10 Close the view details windows and repeat steps 9.6 through 9.9 for additional HOBOs.

9.11 Fill out the rest of the certification form.

10.0 ACCEPTANCE CRITERIA

10.1 Based upon the manufacturer specifications the HOBO Water Temp Pro should be within $\pm 0.4^{\circ}\text{F}$ over the range of 32°F to 100°F. Any HOBO with an error of greater than $\pm 0.5^{\circ}\text{F}$ at any of the three measured points shall fail certification.

11.0 POST PROCEDURE ACTIVITY

11.1 Close the BoxCar Software.

12.0 RECORDS

12.1 Completed HOBO Water Temperature Pro Certification form and associated Report of Certification cover sheet is a QA record.

13.0 REFERENCE

13.1 HOBO Water Temp Pro User's Manual, version 1.0 or later

13.2 Onset BoxCar Pro4 Manual Version 1.0 or later

APPENDIX B

WBN Outfall 113 NPDES Compliance Parameters

- Current Instantaneous Upstream Temperature:

Tu_i (measured at EDS Station 30 by the first sensor below a depth of 5 feet).

- Current 1-Hour Average Upstream Temperature:

$$Tu1_i = \frac{Tu_i + Tu_{i-1} + Tu_{i-2} + Tu_{i-3} + Tu_{i-4}}{5},$$

where the subscripts i , $i-1$, $i-2$, $i-3$, and $i-4$ denote the current and previous four 15-minute (0.25 hour) values of Tu .

- Current Instantaneous Downstream Temperature:

$$Td_i = \frac{Td3_i + Td5_i + Td7_i}{3},$$

where $Td3_i$, $Td5_i$, and $Td7_i$ denote the current measurements of river temperature at the downstream end of the mixing zone at water depths 3 feet, 5 feet, and 7 feet, respectively.

- Current 1-Hour Average Downstream Temperature:

$$Td1_i = \frac{Td_i + Td_{i-1} + Td_{i-2} + Td_{i-3} + Td_{i-4}}{5},$$

where the subscripts i , $i-1$, $i-2$, $i-3$, and $i-4$ denote the current and previous four 15-minute (0.25 hour) values of Td .

- Current Instantaneous Temperature Rise:

$$\Delta T_i = Td_i - Tu_i.$$

- Current 1-Hour Average Temperature Rise:

$$\Delta T1_i = \frac{\Delta T_i + \Delta T_{i-1} + \Delta T_{i-2} + \Delta T_{i-3} + \Delta T_{i-4}}{5},$$

where the subscripts i , $i-1$, $i-2$, $i-3$, and $i-4$ denote the current and previous four 15-minute (0.25 hour) values of ΔT .

- Current Temperature Rate-of-Change:

$$TROC_i = \frac{Td_i - Td_{i-4}}{1 \text{ hour}}.$$

- Current 1-Hour Average Temperature Rate-of-Change:

$$TROC1_i = \frac{TROC_i + TROC_{i-1} + TROC_{i-2} + TROC_{i-3} + TROC_{i-4}}{5},$$

where the subscripts i , $i-1$, $i-2$, $i-3$, and $i-4$ denote the current and previous four 15-minute (0.25 hour) values of TROC.